







PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC DESHAMANGALAM	Name of District: Thrissur
Name of Equipment: ECG MACHINE	Equipment ID & Barcode: 127119 & 0840075
Make: ECG Mac	Model: EM-301
Serial Number: EM3012018030347	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 30-05-2018	Present status of the equipment: fully damaged
Date of breakdown: 02-08-2024 Date of registration of complaint through email/ Toll free: Toll free	*AMC/ CAMC Period agreed at the time of purchase: NO AMC CAMC
Action taken: Machine checked and found that main board, printer head, amplifier board and key pad are defective. Need to replace these spares for further checking working condition of equipment.	
Recommendations for repair (required service details) – Not recommending	
Cost of spares (specify parts and cost): Main board – 6500 - Printer head – 14200 - Keyboard scan board – 3600 - Amplifier board – 9200 - TOTAL -33500/-	
# Percentage value of the cost of spares with respect to Cost of Purchase Asset Value: 112.04%	Asset Value: 29900/-
Abstract of Service Report provided by the OEM Authorized Service Provider: CYRIX (Attached or Not) – OEM quotation and cyrix service report attached	
Reasons for recommending the equipment as BER: Checked the machine and found that main board, amplifier board, printer head and key pad are defective. Equipment installed on 30-05-2018 and aged up to 6 years and 2 months. The spare cost is 112.04% of asset value. Both criteria for RBER met. So as per tender clause 5.3.14.1 we recommending the equipment for condemnation	
BIJO T JOY  Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM: <p style="text-align: center; font-size: 1.2em;">spare cost is veeey high</p>	
Recommended for BER (Yes/ No):	  Signature of JC BM (NHM)
Date: 2/9/2024	
 Date: _____ Seal	 5/9/24 Superintendent / Medical Officer (i.e) # Based on the period of life and 



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1000187

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility: <u>PHC Deshamangalam</u></p> <p>Address: <u>Deshamangalam</u> <u>Thraissur</u></p> <p>Ph: <u>9764059501</u></p>	<p>Call Registration Date : <u>2/8/2024</u></p> <p>Caller ID : <u>127119</u></p> <p>Date of Visit : <u>3/8/2024</u></p> <p>Asset No : <u>0840075</u></p> <p>EQPT Name : <u>ECU machine</u></p> <p>Manufacture: <u>ECU mac</u> Model: <u>EM-301</u></p> <p>S. No <u>EM3012018030341</u> Dept. <u>OP</u></p>
---	--

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Equipment is not working.

Action Taken : Checked the machine and found the main board, Printer Head, Amplifier board and keypad are defective. Need to replace these spares for further checking working condition of equipment.

Completed Date : 3-08-2024 Time : 12:00pm Spare Required

Spare Replaced Requested


Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Aswin A.R</u>	<u>3/8/2024</u>	<u>10:00am</u>	<u>12:00 Pm</u>

Customer Remark _____ Completed Pending

Service Engineer Name : Aswin A.R
 Signature: [Signature]
 Date: 3/8/2024
 Contact Number: 8589965265

Customer Name _____
 Signature: [Signature]
 Date: _____
 Contact Number: _____
 Designation : _____
 Hospital Seal



[Signature]
MEDICAL OFFICER INCHARGE
 FAMILY HEALTH CENTRE
 DESAMANGALAM



barcode - 0840075 caller id - 127119- phc dheshamngalam

4 messages

Bijo T Joy <bijocyrix@gmail.com>

Tue, 20 Aug, 2024 at 10:14 am

To: padmanaban.s@kiranxray.com

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Dear sir,

We would like to bring to your attention a complaint reported from the PHC DESHAMANGALAM HOSPITAL, Thrissur, regarding the ECG machine EM-301 manufactured by ECG Mac, bearing serial number EM3012018030347(Barcode: 0840075). This machine's main board, printer head, key pad and ECG module are defective. So kindly provide the quotation for these spares.

equipment - ECG machine

hospital - PHC DESHAMANGALAM

make - ECG mac

model - EM-301

SN - EM3012018030347

barcode - 0840075

caller id - 127119

quotation for - main board-printer head - key pad -ecg module board

With Regards,

BIJO T JOY

DISTRICT INCHARGE THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia

M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

Bijo T Joy <bijocyrix@gmail.com>

Tue, 20 Aug, 2024 at 10:17 am

To: Anjitha T L <purchasecyrix2@gmail.com>

##looping anjitha

With Regards,

BIJO T JOY

DISTRICT INCHARGE THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia

M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

[Quoted text hidden]

Bijo T Joy <bijocyrix@gmail.com>

Wed, 21 Aug, 2024 at 9:17 pm

To: padmanaban.s@kiranxray.com

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>, Anjitha T L <purchasecyrix2@gmail.com>

reminder

With Regards,

BIJO T JOY

DISTRICT INCHARGE THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia

M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

[Quoted text hidden]

PADMANABAN S <padmanaban.s@kiranxray.com>

Wed, 21 Aug, 2024 at 9:37 pm

To: Bijo T Joy <bijocyrix@gmail.com>

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>, Anjitha T L <purchasecyrix2@gmail.com>

Dear Bijo,

Given below the Prices for the spares requested

Item Description	Unit Price INR
Main Board	6500
Printer Module	14200
Keyboard Scan Board	3600
Amplifier Board	9200

GST 12% Extra on all the above mentioned spares

S Padmanaban

Deputy General Manager – Service (ICU Business)



M : +91 87545 80897

Triviron Healthcare D-117, TTC Industrial Area, Nerul,

Navi Mumbai 400 706, India.



www.triviron.com

Customer Care +91 98400 80008

[Quoted text hidden]

