


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	GH THALASSERY
3	Name of Equipment with Make, Model and Serial Number	CENTRIFUGE Make: ELTEK Model: TC 650D SN: JEC402
4	Equipment ID/ Barcode	1311416 #128690
5	Date of purchase/ Year of manufacture/Installation Date	In ticket master installation date is - 07.10.2011 As per the stock book Installation Date is 17.12.2015, (stock book copy attached for reference)
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 19.08.2024
9	Action taken	Checked the Equipment found Equipment is not switching ON, found that Display, Motor, RPM sensor kit Defective defective. Enquired OEM for quotation
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	Motor Display card RPM Sensor Set } 20,500/- Tax-18% Total-24,190/-



13	Asset Value	23892/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	89.8%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report & quotation attached
16	Reasons for recommending the equipment as BER	Checked the Equipment found Equipment is not switching ON, found that Display Card, Motor, RPM sensor set Defective. Enquired OEM for quotation. In software Equipment installed on 07-10-2011 As per the stock book installation date is 17.12.2015 aged up to 8 years 8 months. Spare cost is 89.8% both criteria for RBER met. As per tender clause 5.3.14.1 we recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AVINASH T 


*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

checked & identified that the machine not switching ON. Display card, motor, RPM sensor are defective. Aged up to 8 years & above. As per OEM quotation the service cost not feasible. So recommended for BER as per BER Protocol.

 Signature of JC BM (NHM)
03/09/2024



Date: 4-9-24


SUPERINTENDENT
Govt. General Hospital
Thalassery
Superintendent (Medical Officer) (c)



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 233933

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrrix.in

<p>Health Facility <u>A.H. Thalassery</u>.....</p> <p>Address <u>Thalassery, Kanner</u>..... <u>Kerala</u>.....</p> <p>Ph : <u>944 7343229</u>.....</p>	<p>Call Registration Date : <u>10-8-2024</u>.....</p> <p>Caller ID : <u>128690</u>.....</p> <p>Date of Visit : <u>12-8-2024</u>.....</p> <p>Asset No. : <u>1211416</u>.....</p> <p>EQPT Name : <u>Centrifuge</u>.....</p> <p>Manufacture <u>ELIEK</u>..... Model : <u>TL650D</u>.....</p> <p>S. No. <u>J.E.C.402</u>..... Dept. <u>Laboratory</u>.....</p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Display, Motor, RPM sensor kit are defective.....

Action Taken : checked the equipment Found Equipment is not switching on. Found that display, motor, RPM sensor kit defective. Enquired OEM for quotation......

Completed Date : ..12-8-24.. Time :11:20 Am.. Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Shankrishnan</u>	<u>12-8-24</u>	<u>10:45 Am</u>	<u>11:30 Am.</u>

Customer Remark Completed Pending

Service Engineer Name : Shankrishnan **Customer Name :** SREEKUMAR.V.K

Signature : **Signature :**

Date : 12-8-24 **Date :** 12-8-24

Contact Number : 8281188514 **Contact Number :** 9447299500

Designation : Lab Technician **Designation :** Superintendent

Hospital Seal : **Hospital Seal :**



QUOTATION REQUIRED MOTOR ,DISPLAY and RPM SENSOR FOR CENTRIFUGE, #1311416,128690 @GH THALASSERY

2 messages

avinash t <avinashtcyrix@gmail.com>
To: ADITHY SCIENTIFIC <adithyscientific@gmail.com>

Mon, 2 Sept, 2024 at 4:02 pm

Dear sir,
Please provide Quotation and stock availability of **MOTOR ,DISPLAY and RPM SENSOR** for below mentioned machine.
EQUIPMENT NAME- CENTRIFUGE
MAKE- **ELTEK**
MODEL- TC 650 D

Thanks & Regards

Avinash T
District charge, kannur
bemp kerala
mobile: +91 6364064948
E mail: avinashtcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B | Petta | Punithura | Kochi | Kerala- 682038
www.cyrixhealthcare.com

ADITHY SCIENTIFIC <adithyscientific@gmail.com>
To: avinash t <avinashtcyrix@gmail.com>

Mon, 2 Sept, 2024 at 4:27 pm

Dear Sir,

Please find attached Quotation for spare parts for Model TC 650D

Kindly place your purchase order (PO) and process the payment as soon as possible.

[Quoted text hidden]

Thanks & Regards,



SVRNSS College Jn ,T. P. Puram P. O Vazhoor , Kodungoor
Kottayam , Kerala. -686505

GSTIN :- 32CCMPR0025E1Z0

Mob :- 9747100296, 7907927607

Email :- adithyscientific@gmail.com

RAKESH M.R.

97471 00296
79079 27607



Opp. SVRNSS College, T.P. Puram P.O.
Vazhoor, Kottayam - 686505

adithyscientific@gmail.com
rakeshmecheril@gmail.com

SALES & SERVICE OF ALL LABORATORY EQUIPMENTS



BRANDS WE DEAL



ADITHY SCIENTIFIC

Opp. SVRNSS College, T. P. Puram P. O., Vazhoor , Kottayam- 686505 Phone:-9747100296, 7907927607

Email: adithyscientific@gmail.com

GSTIN: - 32CCMPR0025E1Z0

Quotation No:- 619 /24-25

Date: 02/10/2024

To
CYRIX HELTHCARE (P) LTD
30/641b , PETTA , PUNITHURA
KOCHI , KERALA -685038

Dear Sir,

With reference to the enquiry, we are pleased to offer a quotation

We hereby submitting our offer.

For further clarifications, please feel free to contact us.

Terms & Conditions:

Delivery terms : At your site
Tax : Price is inclusive of all taxes
Payment terms : 100 % against purchase order
Delivery period : Within 2-3 weeks from your purchase order
Validity : 60 days
Warranty : NA

(Consumables are not covered under warranty, e.g.:- Bulbs, Electrodes, Glass parts ...etc)

We expect your most valuable feedback and we assure the best service all time

For **Adithy Scientific**



1.	ELTEK CENTRIFUGE 1/8 HP AC/DC MOTOR FOR TC 650 D	1	20,500/-
2.	DISPLAY CARD FOR TC 650 D	1	
3.	RPM SENSOR SET FOR TC 650 D	1	
Total Without Taxes			20,500/-
Taxes @ 18%			3,690/-
Grand Total			24,190/-

For **Adithy Scientific**





SUPERINTENDENT
Govt. General Hospital
Thalassery



CENTRIFUGE

		Account	Issued	Balance
30-3-15	Received from the store NVAADCP fund - (2014-15) Model ELECTRA-AL (Prominuss Centrifuge) Issued to: Lab	1	3	4500/-
1-4-15				(Condemned as per order no C4/245/2022 dt 29/7/2022)
17-12-15	Received from the store Eitek Table Top Centrifuge (b/w/ls) Model - TC6500 S/No JEC 402 Eqpt Code 21487 Warranty - one year (KMSCL Supply)	1	4	
3-2-16	Received from the store Model T-8s Nibs 16x15ml SO. Hand Made Loco Scientific Industries & Chemicals Pvt Ltd Eqpt Code. A06A-18 (Stock page 230 stock book Vol. 1: Eqpt Vol of store) Warranty (24-year) S/No. 540/2015/406A dt: 15-10-15 Universal Agencies Ph 0487-2440319 0427038 9387021122 (KMSCL Supply)	1	5	40393/-
				(Condemned as per order no C4/245/2022 dt 29/7/2022)



C/o to Pg No: of 293 of the
Same Register (IV).
Condemned as per order no. C4/245/2022 dt 29/7/2022
SUPERINTENDENT
Govt. General Hospital
Thalassery