




PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: GH KASARGOD	Name of District: KASARGOD
Name of Equipment: NEBULIZER	Equipment ID & Barcode: #131649, 1411092
Make: READY MIST	Model: N/A
Serial Number: N/A	Warranty details: NO warranty
Date of purchase/Year of manufacture/ Installation Date: 12-12-2018	Present status of the equipment: Fully damaged
Date of breakdown: 27/08/2024	*AMC/ CAMC Period agreed
Date of registration of complaint through email/ Toll free: toll free	at the time of purchase: NO AMC/CAMC
Action taken: Checked the machine and found that Machine is not switching on, Found motor and piston are defective. Need to replace these spares for further checking.	
Recommendations for repair (required service details): No Recommendations	
Cost of spares (specify parts and cost): N/A	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: N/A	Asset Value: 1353/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Service report attached	
Reasons for recommending the equipment as BER: Checked the machine and found that Machine is not switching on, Found motor and piston are defective. Equipment installed on 12-12-18 aged up to 5 year 7 months. Quotation not submitted since spares are not available from the market. So we recommend the equipment for condemnation.	
AJITH BABU K  Name & Signature of CYRIX Authority with date 3-9-24	
Remarks of Junior Consultant (Biomedical) NHM: Checked the machine and found motor and piston are defective. Equipment installed on 12-12-18. As per cyrix spares are not available from the market so quotation not submitted.	
Recommended for BER (Yes/ No):	Ir. CUNSU TAKT Biomedical Engineering Keralam (NHM) at Kanhangad  Signature of JC BM (NHM)
Date: 10/09/2024	
 Signature of Superintendent / Medical Officer (i/c)	
Date: 11-09-2024	Seal

*Not mandatory # Based on the period of life and value as per the BER guidelines * Attach Photograph



ശ്രീ. ജനറൽ ആശുപത്രി
 കാസർഗോഡ്
 കാസർഗോഡ് ടി.ഒ. - 671 12



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1002411

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>SH Kasaragod</u>	Call Registration Date : <u>27-8-24</u>
Address..... <u>Kasaragod</u>	Caller ID : <u>131649</u>
..... <u>Kerala</u>	Date of Visit : <u>28-8-24</u>
Ph : <u>9447608668</u>	Asset No : <u>1411092</u>
	EQPT Name : <u>Nebuliser</u>
	Manufacture <u>Ready mist</u> Model : <u>NA</u>
	S. No. <u>NA</u> Dept. <u>paediatric ward</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : motor and piston damage

Action Taken : checked the machine and found that the motor and piston are damaged. need to replace those spares for further working condition of the equipment

Completed Date : 28-8-24 Time : 10:30 am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			


Cyrix Engineer <u>Ajith Babu</u>	Date <u>28-8-24</u>	Start Time <u>10:00 pm</u>	End Time <u>10:30 pm</u>
--	------------------------	-------------------------------	-----------------------------

Customer Remark Completed Pending




Service Engineer Name : Ajith Babu
Signature: [Signature]
Date: 28-8-24
Contact Number: 8848079288

Customer Name : Reena. m. D
Signature: [Signature]
Date: 28/8/24
Contact Number: 9961570724
Designation : SNO
Hospital Seal



Handwritten signature and text in Malayalam script, including a phone number: 671 121.



STOCK REGISTER OF NON PERISHABLE NAME

INSTRUMENTS & EQUIPMENTS OF ITEM

Brought forward from Stock Book of Nebulizer

Vol II Page 23 On 1-11-06

Date	Name of Manufacturer	Serial No/ Code No.	Unit Price	Opening Balance	Receipt & Source	Total	Total Not working	Date of Repair	Total Working	Remarks / Transfer out	Balance	Initial of Custodian
1-11-06				4	3674-3 2. Wd - 1	4	3		1	pd no 21-11-05 BY-08/29/06 + 77-16 MLWd on 20/9/09	4	
	C & D done as per order No. C1-11/07/2005/0m/16 Kanhenad &											
	C1-2798/2005/1 Dept G.H.C. Old 25/11/2005									CX0-3	1	
21/1/2006	Rd from the Stock (LP)			1			1		1		2	
	1. N. F. 447-V MEMO A.I.P. / N.D. 1 Let no 23-5-2006	091005022 000106557										
3-9-10	Rd from the Stock (LP)						2		0		3	
	ASHOK ENTERPRISES DELHI	07100485 LATA 210										
9-03-11	Rd from the Stock (LP)						1		1		4	
	ASHOK ENTERPRISES DELHI	07100485 LATA 210										
	ASHOK ENTERPRISES DELHI	07100485 LATA 210		5	ISSUED 3-2							
11/3/12	Tad to unserviceable equipment						2+1				1	
	Stock Register page No 90											
11/9/14	Tad from page No 93										2	
	(Rd on 29/6/12 given by Dr. Jayaraman)											
11/11/14	Tad to unserviceable book P. No 22						1				NIL	
	(Considered as order no 3/2005/122 suppl. int. no 29/6/2006)											
20/7/14	Tad to unserviceable book page 90											
24/2/16	Rd from the Stock										1	
14/5/16	Rd from the Stock (as donated item) by Dr. Jayaraman						1				2	
	1 donated items - as order no 3/2005/122											
5/2/17	Gr out to unserviceable book P. No 22						1				1	
15/11/17	in in form of order no 244										2	
12/12/17	Rd from the Stock (as donated item) by Dr. Jayaraman											
	(23 2120) memo in ist											
20/7/20	Rd from the Stock (as donated item) by Dr. Jayaraman										3	
	equipment Stock by wajayagiri pg no 39											



Carried forward to stock Book of Same book page No: 156 157

Page No 156 on 157