


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KANNUR
2	Name of Hospital	THQH PERAVOOR
3	Name of Equipment with Make, Model and Serial Number	FOETAL DOPPLER Make: BPL Model: FD01 SN: NA
4	Equipment ID/ Barcode	1320342 #125602
5	Date of purchase/ Year of manufacture/Installation Date	10.08.2020
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free 26.07.2024
9	Action taken	Checked the Equipment found Equipment is not switching ON, found that Display With main board and Probe Defective. Enquired OEM for quotation.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	DISPLAY WITH MAIN BOARD-6000/- PROBE FOR FOETAL DOPPLER-5000/- TOTAL-11,000/-

13	Asset Value	4,000/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	275%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report & quotation attached
16	Reasons for recommending the equipment as BER	Checked the Equipment found Equipment is not switching ON, found that Display and Probe Defective. Enquired OEM for quotation. Equipment installed on 10.08.2020. Aged up to 3 years and 11 months. Spare cost is 275% both criteria for RBER met. As per tender clause 5.3.14.1 we recommend the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AVINASH T 


*Not mandatory

#Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


checked & identified that equipment not switching ON, due to probe defective. Aged up to 3 years and above. its service cost economically not feasible as per quotation. So recommended for BER as per BER protocol.


03/9/2024
Signature of JC BM (NHM)

CONSULTANT
NATIONAL HEALTH MISSION
KANNUR - 670002



Date: 5/9/2024


 SUPERINTENDENT
 TALUK HOSPITAL
 PERAVOOR
 KANNUR
 670002



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 232027

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : <u>THQH PERAVOOR</u></p> <p>Address : <u>Peravoor</u> <u>KANNUR</u></p> <p>Ph : <u>9495 283064</u></p>	<p>Call Registration Date : <u>26/07/2024</u></p> <p>Caller ID : <u>125602</u></p> <p>Date of Visit : <u>27/07/2024</u></p> <p>Asset No. : <u>1320342</u></p> <p>EQPT Name : <u>Fetal doppler</u></p> <p>Manufacture : <u>BPL</u> Model : <u>FD 01</u></p> <p>S. No. <u>LLTAOC10138</u> Dept. <u>Gynec OP</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Display with Mainboard and Probe are defective

Action Taken : Checked the equipment and found equipment is not switching ON, Found that Display with mainboard and probe are defective. Enquired OEM for quotations.

Completed Date : 27/7/2024 Time : 11:30 Am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Sayanth</u>	<u>27/07/2024</u>	<u>10:30 Am</u>	<u>11:30 Am</u>

Customer Remark Completed Pending

Service Engineer Name : Sayanth

Signature :

Date : 27/07/2024

Contact Number : 8606345195

Customer Name : Admini. PL

Signature :

Date : 27/7/24

Contact Number : 9495283064

Designation : Sno

Hospital Seal :



SUPERINTENDENT
TALUK HOSPITAL
PERAVOOR
KANNUR
670673

30.08.2024
RINS/SA/24/6**TO,**
CYRIX HEALTHCARE PRIVATE LIMITED
ERNAKULAM**QUOTATION FOR DISPLAY & PROBE**

SL NO	ITEM	RATE	QTY	AMOUNT
1	PROBE FOR FOETAL DOPPLER	5000	1 no	5000.00
2	DISPLAY WITH MAIN BOARD FOR FOETAL DOPPLER	6000	1 no	6000.00
GST EXTRA 12%				

For , Rins Marketing & Services

**Authorized Signatory****OUR BANK DETAILS:**

BANK: FEDERAL BANK, A/C NO: 19975500000015, IFSC: FDRL0001997, BRANCH: TALAP



quotation required for display and probe #1320342 @thqh peravoor

4 messages

avinash t <avinashtcyrix@gmail.com>

Sat, 24 Aug, 2024 at 12:04 pm

To: rinsmarketing rinsmarketing service <rinsmarketing@gmail.com>

Dear sir,

Please provide quotation of below mentioned parts for fetal doppler(make:BPL, model:

1. probe

2.display

Thanks & Regards

Avinash T

District charge, kannur

bemp kerala

mobile: +91 6364064948

E mail: avinashtcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD

30/641B | Petta | Punithura | Kochi | Kerala- 682038

www.cyrixhealthcare.com

rinsmarketing rinsmarketing service <rinsmarketing@gmail.com>

Fri, 30 Aug, 2024 at 1:31 pm

To: avinash t <avinashtcyrix@gmail.com>

Please find the attached quotation.

Regards,

Anagha

RINS
MARKETING & SERVICES

TM 364 N, 1st Floor, City Tower

Opp. Chinmaya Balabhavan, Kannur-2

Tel.: 0497-2704624, Fax.: 2709789

Mob.: 9846008441, 9947064624

E-mail: rinsmarketing@gmail.com

[Quoted text hidden]

rinsmarketing rinsmarketing service <rinsmarketing@gmail.com>

Fri, 30 Aug, 2024 at 2:05 pm

To: avinash t <avinashtcyrix@gmail.com>

Please find the attached quotation.

Regards,

Anagha

RINS

MARKETING & SERVICES

TM 364 N, 1st Floor, City Tower
Opp. Chinmaya Balabhavan, Kannur-2
Tel.: 0497-2704624, Fax.: 2709789
Mob.: 9846008441, 9947064624
E-mail: rinsmarketing@gmail.com

[Quoted text hidden]

rinsmarketing rinsmarketing service <rinsmarketing@gmail.com>
To: avinash t <avinashscyrix@gmail.com>

Tue, 3 Sept, 2024 at 1:12 pm

Regards,

RINS

MARKETING & SERVICES

TM 364 N, 1st Floor, City Tower
Opp. Chinmaya Balabhavan, Kannur-2
Tel.: 0497-2704624, Fax.: 2709789
Mob.: 9846008441, 9947064624
E-mail: rinsmarketing@gmail.com

[Quoted text hidden]



CYBER MEDICAL EQUIPMENT
TOLL FREE NO: 1800-425-1600
SN: 1320342
SAL COM: (0804-88618225)

BPL
Model: FD-01
SN: LT10011818
Serial No: 1320342



Dr. Sahin
SUPERINTENDENT
TALUK HOSPITAL
PERAVOOR
KANNUR
670673

