

**PROFORMA  
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: GH THRISSUR	Name of District: Thrissur
Name of Equipment: B P Apparatus	Equipment ID & Barcode: 133089 & 0810963
Make: EL 520	Model: ELKO
Serial Number: 18074400594	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 13-06-2020	Present status of the equipment: Fully damaged
Date of breakdown: 03-09-2024 Date of registration of complaint through email/ Toll free)- Toll free	*AMC/CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Machine checked and found board and display are defective. Need to replace these spares for working condition of equipment.	


Recommendations for repair (required service details) - Not recommending

Cost of spares (specify parts and cost): N/A


# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: N/A	Asset Value: 2200/-
---	---------------------


Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) - Cyrix service quotation attached

Reasons for recommending the equipment as BER: checked the machine and found board and display are defective. Equipment installed on 13-06-2020 aged up to 4 years 2 months. Quotation not submitted since Spears are not available in market. So, we are recommending equipment for condemnation.

**BIJO T JOY**   
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:  
*space not available as per the rate*

  
Jr. Consultant (Biomedical)

Recommended for BER (Yes/ No):	 Signature of SC BM (NHM)
Date:	

Date:	Seal 	Signature of Superintendent / Medical Officer (i/c) <i>[Signature]</i> 10/09
-------	--	--

\*Not mandatory # Based on the period of life and value as per the BER guidelines \* Attach Photograph

**SUPERINTENDENT (Dy. DHS)  
GENERAL HOSPITAL  
THRISSUR - PIN: 680 001**



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No :1006651

# CYRIX<sup>®</sup>

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

## Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility... <u>GH</u>	Call Registration Date : <u>3-9-2024</u>
Address... <u>THRISSUR</u>	Caller ID : <u>133089</u>
<u>Kondam</u>	Date of Visit : <u>4-9-2024</u>
Ph : <u>9400884047</u>	Asset No : <u>0810963</u>
	EQPT Name : <u>BP APPARATUS</u>
	Manufacture <u>ELKO</u> Model : <u>EL920</u>
	S. No. <u>18074400594</u> Dept. <u>ward 7/8</u>

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : Not working BP apparatus.

Action Taken : Machine checked and found board and display are defective. need to replace these spares. For working condition of equipment.

Completed  Date : 6-9-2024 Time : 2:00pm Spare Required

Spare Replaced  Requested



Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Absil Tony</u>	<u>4-9-24</u>	<u>1:00pm</u>	<u>2:00pm</u>

Customer Remark  Completed  Pending

Service Engineer Name : Absil Tony  
Signature:   
Date: 4-9-2024  
Contact Number: 9745328470

Customer Name :   
Signature:   
Date:   
Contact Number:   
Designation : SUPERINTENDENT (Dy.DHS)  
Hospital Seal GENERAL HOSPITAL  
THRISSUR - PIN: 680 001



B.P. Apparatus

2 BP apparatus  
L.P. of one  
41

		Pd	Issued	Balance	By
22/9/05	Tr in from 1st 5 (no 257/2003)	1		1	H
22/9/05	Tr in from 1st Book (27-1-2001)	1		2	H
25/2/08	Rd from store no. 1475	1		3	AD
8/9/08	Condemned as per order no. 418/2007. D.H. 8/9/08 (21)		1	2	g
11/4/13	condemnation done as per order no B3-1753/2012 D.H. 15/4/13		1	1	H
16/6/14	Rd from store (page 104)	1		2	
9/4/15	Tr in from CCU Issued to wd	1	3	nil	(3)
13/4/16	Rd from store (herbal purse)	1		1	
16/4/16	Issued to wd		1	nil	
26/9/16	40 pay no: 19/Unserviceable Regd. In issue - 1 ? Condemned Not working - 2 on 4/9/18		1	3	(4)
10/2/18	Tr to Unserviceable Reg PNO 19.	2	3	1	(4)
29/8/2019	Recovery from store (Not working) 2.	1		2	7
13/6/2020	R.F.G. (Digital BP apparatus) (LP) Battery Adaptor Battery charged	3		5	8
	Issued to wd wd to wd				
	Using (1) digital cupboard not working (1) new cupboard (1) manual not working (2)				

1 new  
not working digital (1) 6  
1 (3) 9

