

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: UPHC TANUR	Name of District: MALAPPURAM
Name of Equipment: UPS	Equipment ID & Barcode: 130678 #1041804
Make: APC	Model: Back -UPS 600
Serial Number: B21732012720	Warranty details: NO Warranty
Date of purchase/Year of manufacture/ Installation Date: Ticket master installation date:03/05/2017.As per stock book installation date: 07/12/2017	Present status of the equipment: Fully damaged
Date of breakdown: 21/08/2024 Date of registration of complaint through email/ Toll free): Toll free	*AMC/CAMC Period agreed at the time of purchase: NO AMC/ CAMC
Action taken: Checked the machine & found Main pcb &12V 7Ah battery damaged. Need to replace these spare for further checking.	

Recommendations for repair (required service details): No recommendations

Cost of spares (specify parts and cost): 12 V 7AH BATTERY -944/-
 MAIN BOARD -2006/-
 TOTAL -2950/-

Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 65% Asset Value: 4500/-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Service report attached & OEM Letter attached

Reasons for recommending the equipment as BER: Checked the machine and found Power board, Main PCB &12V 7Ah Battery defective .In ticket master installation date is 03-05-2017, As per stock book the equipment was installed on 07/12/2017. Repairing cost is 65% both criteria for RBER met. As per the tender clause 5.3.14.1 we recommending the equipment for condemnation.


GOKUL KT *[Signature]* 13/09/24
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: *checked with the need to change the same. Beyond economic repair.*

Recommended for BER (Yes/ No): YES
 Date: 13/09/24
 Jr. Consultant Biomedical Engineer NHM Malappuram *[Signature]* 13/09/24
 Signature of JC BM (NHM)

Date: 17/09/2024 Seal Signature of Superintendent / Medical Officer (i.e.)
  MEDICAL OFFICER URBAN PRIMARY HEALTH CENTRE

*Not mandatory # Based on the period of life and value as per the BER guidelines * Attach Photograph





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 231384

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrilx.com | E-mail : bemp.kl@cyrilx.in

Health Facility <u>UPNC TANUR</u> Address <u>MALAPPURAM</u> Ph : <u>9526787171</u>	Call Registration Date : <u>21/08/2024</u> Caller ID : <u>130678</u> Date of Visit : <u>21/08/2024</u> Asset No. : <u>1041804</u> EQPT Name : <u>UPS</u> Manufacture <u>APC</u> Model : <u>BACK UPS 600</u> S. No. <u>B21732012720</u> Dept. <u>Laboratory</u>
--	--

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Backup problem

Action Taken : checked the machine and confirmed that the main PCB and 12V 7Ah battery damaged. Need to replace these spares for further checking.

Completed Date : 21/08/2024 Time : 11:15AM Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.	<u>Nil</u>			
2.				
3.				

Cyrilx Engineer	Date	Start Time	End Time
<u>Rohith</u>	<u>21/08/2024</u>	<u>10:30AM</u>	<u>11:15AM</u>

Customer Remark [Signature] **Completed** **Pending**
 MEDICAL OFFICER
 URBAN PRIMARY HEALTH CENTRE
 TANUR

Service Engineer Name : <u>Rohith</u> Signature : <u>[Signature]</u> Date : <u>21/08/2024</u> Contact Number : <u>9539838931</u>	Customer Name : <u>Rupeda</u> Signature : <u>[Signature]</u> Date : <u>21/8/2024</u> Contact Number : <u>9526787171</u> Designation : <u>Lab - Technician</u> Hospital Seal :
---	--



APS MARKETING COMPANY

27/233A, GROUND FLOOR, APS ARCADE
NETHAJI ROAD, PUTHIYARA, CALICUT - 673004
PH: 0495 - 2741313, Mob: 9447184933. Mail: info@apsmarketing.in

QUOTATION

Quotation No : APS-2011
Valid Upto : 13-09-2024
Customer : GOKUL KT
CYRIX HEALTH CARE PVT LTD, MALAPURAM
9809314280

#	DESCRIPTION OF ITEM	QTY	RATE	TAXABLE	GST %	TOTAL
1	7AH BATTERY 1 YEAR WARRANTY	1	800.00	800.00	18.00	944.00
2	APC 600 VA UPS MAIN BOARD W/N	1	1700.00	1700.00	18.00	2006.00

G-Total 2950.00

Amount Chargeable (in words)
Two Thousand Nine Hundred Fifty Rupees only

E & O.E

Remarks

FOR APS MARKETING COMPANY

Authorised Signatory



1041804 UPS MACHINE MAIN BOARD & BATTERY NEED PRICE QUOTATION

3 messages

Gokul K T <gokulcyrix@gmail.com>

Thu, 29 Aug, 2024 at 11:39 pm

To: apsmarketingservice@gmail.com

Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, rahuljith <rahuljithcyrix@gmail.com>

Dear Sir, Madam

Can you send the Price quotation for UPS main board & Battery Make :APC, model :Back -Ups 600, Serial number :B21732012720

Barcorde:1041804

Ticket id:130678

UPS

UPHC TANUR

GOKUL KT

CYRIX HEALTHCARE Pvt Ltd

DISTRICT INCHARGE

MALAPPURAM

9809314280

Gokul K T <gokulcyrix@gmail.com>

Wed, 4 Sep, 2024 at 8:37 am

To: apsmarketingservice@gmail.com

Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, rahuljith <rahuljithcyrix@gmail.com>

Reminder

[Quoted text hidden]

Gokul K T <gokulcyrix@gmail.com>


Fri, 20 Sep, 2024 at 12:43 pm

To: apsmarketingservice@gmail.com

Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, rahuljith <rahuljithcyrix@gmail.com>

Reminder

[Quoted text hidden]



+91 80756 42700

last seen today at 8:09 pm



BOARD & BATTERY NEED

PRICE QUOTATION Add label



Gokul K T 12:42 pm

to beeranaahraf, Ble... ▾



Dear Sir, Madam

Can you send the Price quotation for UPS main board & Battery Make :APC, model :Back -Ups 600, Serial number :B21732012720

Barcorde:1041804

Ticket id:130678

UPS

UPHC TANUR

GOKUL KT

CYRIX HEALTHCARE PVT Ltd

DISTRICT INCHARGE

MALAPPURAM

9809314280



Forwarded

APS MARKETING COMPANY						
27023A, GROUND FLOOR, APS ARCADE METHAL ROAD, PUTHYVIRA, CALICUT - 673004 PH: 9495 2747113, MO: 9447194925, Mail: info@apsmarketing.in						
QUOTATION						
Quotation No:	APS-2021	Customer:	GOKUL KT			
Date:	05-05-2024		CYRIX HEALTH CARE PVT LTD, MALAPPURAM			
Valid upto:	10-05-2024		9809314280			
#	DESCRIPTION OF ITEM	QTY	RATE	TAXABLE	GST %	TOTAL
1	TAM BATTERY + 1 year warranty	1	886.00	886.00	18.00	944.00
2	APC 600 VA UPS MAIN BOARD	1	1700.00	1700.00	18.00	2006.00

QUOTATION 123342.pdf

1 page · 141 kB · PDF

2:38 pm

Today

SEMI-AUTO ANALYSER, UPS MISPA VIVA


7/12/2017 - Received from Home.

- 1) Semi-automated clinical chemistry analyser: 1 No.
- 2) UPS mispa viva 1. Nos.



MEDICAL OFFICER
URBAN PRIMARY HEALTH CENTRE
TANUR





MEDICAL OFFICER
URBAN PRIMARY HEALTH CENTRE
TANUR

