

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	KASARGOD
2	Name of Hospital	GH KASARGOD
3	Name of Equipment with Make, Model and Serial Number	Nebulizer Make: POINT OF CARE Model: N/A SN:N/A
4	Equipment ID/ Barcode	#130467, 1412126
5	Date of purchase/ Year of manufacture/Installation Date	06-10-2022
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free- 21-08-2024
9	Action taken	Checked the machine and found that Machine is not switching on. Found motor and piston are defective. Need to replace these spares for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 1002231

# CYRIX<sup>®</sup>

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

## Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <u>General Hospital</u>	Call Registration Date: <u>21.08.2024</u>
Address: <u>Kasavaged</u>	Caller ID: <u>130467</u>
<u>Kerala</u>	Date of Visit: <u>21.08.2024</u>
Ph: <u>9846283663</u>	Asset No: <u>1412126</u>
	EQPT Name: <u>Nebuliser</u>
	Manufacture: <u>Point of care</u> Model: <u>NA</u>
	S. No: <u>NA</u> Dept: <u>FM Ward</u>

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : Fumes not coming

Action Taken : Checked the machine and found that machine is not switching on. Found motor and piston are defective. Need to replace these spares for further checking.

Completed  Date : 21.08.2024 Time : 4.00pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			


Cyrix Engineer	Date	Start Time	End Time
<u>Abhinav D</u>	<u>21.08.24</u>	<u>3.20pm</u>	<u>4.00pm</u>

Customer Remark Completed  Pending

Service Engineer Name : Abhinav D  
Signature: [Signature]  
Date: 21.08.2024  
Contact Number: 8942392989

Customer Name: [Signature]  
Signature: [Signature]  
Date: 21/8/24  
Contact Number: 9846283663  
Designation: [Signature]  
Hospital Seal: [Seal]





തിരുവനന്തപുരം  
 ജനറൽ ആശുപത്രി  
 കൗൺസിലറുടെ  
 ഓഫീസ് പി.ഒ. - 671 121





NEBULISER.

		Recd	Issue	Balance	Sign
3-11-2008	T9. in from OLD MS NO Stock Book P. NO: 153 (1)	1		1	✓
15-1-10	Rd from store (L.P)	1	1	2	✓
5/1/14	Not working (beyond repair) put in use		1	1	✓
12/1/16	Rd. from store (L.P) <sup>put in use</sup> <sub>PN: 39</sub> <sup>Grout to</sup> <sub>unserviceable</sub>	1		2	✓
	(Gr In from LP book P. No. 131)				
24/1/16	Rd from store (Donation) <sup>put in use</sup>	1		3	✓
26/1/16	not working, not repairable. Grout to unserviceable <sup>put in use</sup> PN: 39	1	2	1	✓
2/1/18	Received from the store (safe touch) Kasunya Kollara (Rs - 1562.65)	1		2	✓
6/2/19	Received from the store <sup>put in use</sup> (Rs. 3120/-)	1		3	✓
	<sup>put in use</sup>		1		✓
10/2/22	Treat to unserviceable book page no: (condemned)	1	2	1	✓
00/9/22	Rd from CFITC udayagiri	1		2	✓
6/10/22	Rd from store (Donation)	1		3	✓
	<sup>put in use</sup>		1	2	✓
	(Condemned (2) as per order no. C2 - 7637/2022 / Supt / CH / Kaseregod / Dt 20/1/22.		2	1	✓
03/3/22	Received from the store	1		2	✓
	T9. in from LPS Registar page no. 131 / FMW Ls/00109/23-24 dt. 6/3/22 scheme time. Beurer India Pvt. Ltd. SF No 288/1, Do No.: 43, Annasalai Model: IH.18, Type: IH 18/2, Art No: 101.16 230V ~ / 50Hz / 1.0A / 230VA $\omega$ min. ON 40ms dt SN 2023123/0061408			1	✓
	Not working (Beyond repair) Treat to unserviceable register page no. 39				



Handwritten signature and official stamp with Malayalam text and number 6711.