


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KASARGOD
2	Name of Hospital	TH THRIKARIPUR
3	Name of Equipment with Make, Model and Serial Number	Nebulizer Make: EASYCARE Model: EC-7011 SN:2015242010007141
4	Equipment ID/ Barcode	#130157, 1420090
5	Date of purchase/ Year of manufacture/Installation Date	Ticket master installation date is 05-12-2015. As per stock book installation date is 12-05-2015 (Stock book attached for reference).
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	Toll free- 19-08-2024
9	Action taken	Checked the machine and found that Machine is not switching on, Found motor and piston are defective. Need to replace these spares for further checking.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	No recommendations
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	1353 /-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Service report attached
16	Reasons for recommending the equipment as BER	Checked the machine and found that Machine is not switching on, Found motor and piston are defective. Ticket master installation date is 05-12-2015. As per stock book installation date is 12-05-2015 (Stock book attached for reference) aged up to 9 years 3 months. Quotation not submitted since spares are not available from the market. So we recommend the equipment for condemnation.
17	Name & Signature of CYRIX Authority	VARUN KUMAR K  17/09/24


*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Checked the machine and found motor and piston are defective. Equipment installed on 12/05/2015 (Stock book details attached). As per cyrix, ~~parts~~ spares are not available from the market so quotation not submitted. Recommending the equipment for BER.


Signature of JCBM (NHM)
Biomedical Engineer
Arogyakeralam (NH-1)
Kasaragod at Kasaragod

Date:




SUPERINTENDENT/MEDICAL OFFICER (i/c)
TALUK HOSPITAL
TALUK KARIPUR



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 198844

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : TH <u>Trikaripur</u></p> <p>Address : <u>Trikaripur</u> <u>Kasaragod (Dist)</u></p> <p>Ph : <u>9495112292</u></p>	<p>Call Registration Date : <u>19/08/2024</u></p> <p>Caller ID : <u>130157</u></p> <p>Date of Visit : <u>20/08/2024</u></p> <p>Asset No. : <u>1420090</u></p> <p>EQPT Name : <u>Nebuliser</u></p> <p>Manufacture <u>Easy care</u> Model : <u>GI-7011</u></p> <p>S. No. <u>2015242010007141</u> Dept. : <u>OP</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Machine not switching on

Action Taken : Checked the machine, found the problems with machine motor and motor piston. Need to replace motor and piston for further servicing of the machine.

Completed Date : 20/8/24 Time : 11.0M Spare Required

Spare Replaced Requested

1.	Description	Qty.	Part Number	PR Number
2.				
3.				



Cyrix Engineer	Date	Start Time	End Time
<u>Vasun kumar.k</u>	<u>20/8/24</u>	<u>10AM</u>	<u>11AM</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Vasun kumar.k</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>20/08/2024</u></p> <p>Contact Number : <u>9072522045</u></p>	<p>Customer Name : <u>S/JI.C.A.</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>20/8/2024</u></p> <p>Contact Number : <u>9495112292</u></p> <p>Designation : <u>SNO</u></p> <p>Hospital Seal :</p>
---	---





MEDICAL OFFICER
TALUK HOSPITAL
TRIKARIPUR

[Handwritten signature]
19/9/24

MEDICINE STOCK REGISTER

Name of Drug NEBULISER nebuliser

Date	Particulars of Receipt / Issue	Receipt	Issue	Balance	Rate		Cost	Batch No.	Date of Expiry	Initials
					Unit	Rs. Ps.				
10-8-2001	Bl. from stock book vol no: 1 page no: 135	2			2		165	(2)	13-7-99	
5/10/02	Received from the (APL Airjolie)	1 no	3 boxes	3			165-0			
11/7/07	Received from the store	2 no					98-0			
13/1/09	condemned as per order no. C1/5188/2008 DMD (H.) Kalyan dated 13/01/09									
2/1/16	Rd from the store Batch no. AB0100 exp. 1/2/25 Tr. out to Local medicine shop page no. 134									
6/3/12	Rd from the store									
13/6/13	Condemned as per order no. 10/2013 dt 13/6/13									
10/2/15	Rd from L.P. Stock book - vol. 11 KMSCL supply	2								
12/1/15	Rd from L.P. Stock book - vol. 11	1								
15/2/2016	Rd from L.P. Stock book - vol. 11	4								
2/10/19	Rd from the store Tr. out to Unservicable Register r.no. 513-6/9/12-2/11/10	2 no.								
17/12/21	Tr. in from Donator Register page no-15	1 no								
21/3/22	Received from store KMSCL supply	3 Nos								



MEDICAL OFFICER
TALUK HOSPITAL
TRIKARPUR

MEDICINE STOCK REGISTER

Name of Drug NEBULISER *nebuliser*

Date	Particulars of Receipt / Issue	Receipt	Issue	Balance	Rate		Batch No.	Date of Expiry	Initials
					Unit	Rs. Ps.			
10.5.00	BT for stock book vol. no. 135	2					2	13.7.99	
5/10/02	Received from the store (M.L. Arjoke)	1 no.			3	three	page no. 165-0	ward vol. 2	
11/7/07	Received from the store	2 no.			5 no.	OP	page no. 98-0	page 165	
13/1/09	Condemned as per order No. C1/5188/2008 DMO (H.) Kalyan dated 13/01/09								
2/1/16	Rd from the store								
6/3/12	Rd from the store								
17/6/13	Condemned as per order of 17/6/13 of DMO								
10/2/15	Rd from L.P. Stock book	2							
12/5/15	Rd from L.P. Stock book	1							
15/2/2016	Rd from L.P. Stock book	4							
21/1/19	Rd from the	2 no.							
17/12/21	Tr. in from District Registrar	1 no.							
21/3/22	Received from KMSCL Supply	2							

Substock

21/4/01

A

207

207 (Kans)



MEDICAL OFFICER
TALUK HOSPITAL
TRIKARPUR