

PROFORMA

**RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: THQH ALATHUR	Name of District: PALAKKAD
Name of Equipment: NEBULISER	Equipment ID & Barcode: #134147- 0920940
Make: OMARON	Model: COMP AIR PRO
Serial Number: 201701010144UF	Warranty details: NO Warranty
Date of purchase/Year of manufacture/ Installation Date: Ticket master installation date - 26/02/2019	Present status of the equipment: Fully damaged
Date of breakdown: 09/09/2024	*AMC/CAMC Period agreed at the time of purchase: NO AMC/ CAMC
Date of registration of complaint through email/ Toll free: Toll free	
Action taken: Checked the machine found not switching on. Found nebuliser motor burned , piston broken. Need to replace nebuliser motor with piston for further checking of machine.	
Recommendations for repair (required service details): No recommendations	
Cost of spares (specify parts and cost): N/A	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: N/A	Asset Value: 1353/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Service report attached	
Reasons for recommending the equipment as BER: Checked the equipment found not switching on. Found nebuliser motor burned and piston broken. Equipment installed on 26/02/2019 and aged up to 5 years 6 months. Quotation not submitted since Spare parts are not available on market. So we recommend the equipment for condemnation.	
<p align="right">AKHILESH P</p> <p align="right">Name & Signature of CYRIX Authority with date <i>18/09/24</i></p>	
Remarks of Junior Consultant (Biomedical) NHM: <i>Nebulizer installed on 26/02/2019 and aged up to 5 years. Quotation not submitted.</i>	
Recommended for BER (Yes/ No): <i>Yes</i>	Consultant (Biomedical Engg): <i>[Signature]</i> NHM (Arogyakeralam), Palakkad Signature of JC BM (NHM) <i>19/09/24</i>
Date:	
Date:	Seal Signature of Superintendent / Medical Officer (i/c) SUPERINTENDENT HOSPITAL ALATHUR

Not mandatory # Based on the period of life and value as per the BER guidelines Attach Photograph





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1003689

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>THQH ALATHUR</u> Address..... <u>ALATHUR</u> <u>PALAKKAD</u> Ph:..... <u>9847852898</u>	Call Registration Date : <u>09/09/24</u> Caller ID : <u>134147</u> Date of Visit : <u>11/09/24</u> Asset No : <u>0920940</u> EQPT Name : <u>NEBULIZER</u> Manufacture. <u>OMARON</u> Model <u>COMP-AIR-PRO</u> S. No. <u>201701010144VF</u> Dept. <u>CH.NRD</u>
----------------------------------------------------------------------------------------------------------------------------------	-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Machine not switching ON.

Action Taken : Checked the machine. Found that nebulizer motor burned and piston broken. Need to replace nebulizer motor with piston for further checking of machine.

Completed Date : 11/09/24 Time : 11:00AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
<u>DIVINIS</u>	<u>11/09/24</u>	<u>10:00AM</u>	<u>11:00AM</u>

Customer Remark Completed Pending

Service Engineer Name : DIVINIS
 Signature: [Signature]
 Date: 11/09/24
 Contact Number: 7593847155

Customer Name : THQH ALATHUR
 Signature: [Signature]
 Date: 11/9/24
 Contact Number: 9847852898
 Designation : Nursing officer AI
 Hospital Seal : [Seal]



SUPERINTENDENT
THQH HOSPITAL


Nebulizer Instrument

80
0863
0862

Date	Description	Received	Issued	Balance	Remarks
2/7/99	Received from the store	1		1	MIP
26/6/02	Received from the store	1		2	
1/3/08	Received from the store (MIP)	1		3	Reg.
2/3/08	Transfer out to Day ward		1	2	Reg.
6/2/12	Received from the store	1		3	Reg.
6/9/2012	Condensed as per order No: C/520/2012 dt 23.6.12 A. Ruppel - THCPH.		1	2	
3/13/2018	Received from the store COMPAT- Pro- Compressor NE-C-29 model - Nebulizer.		1	3	Ann
	Issued for use		3	MIL	3
26/02/19	Received from store CLP RSBY Store Stock Register Page No. 164) OHREN - [Cost ₹ 4,400/-]	1		1	Reg
26/02/19	Issued for use Kioskos Barcode: 8907579-80092139		1	NIL	4
10/03/19	Tr-out to Reg. Code G1 Page No. 36		2		2
19/1/21	Received from store (LifeCare Enterprise) Expiry - 31/07/2026 Rate - 000	1		3	@
3/12/22	Tr out to Reg. Code G1 Page No. 36		1		2
6/3/23	Del from store	2		4	n
"	Issued from store		2	2	4

2 new models
1 used
1 store - nil





Amir
SUPERINTENDENT
T.H.Q. HOSPITAL
ALATHUR