




PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: THQH MANNARKKAD	Name of District: PALAKKAD
Name of Equipment: FUMIGATOR	Equipment ID & Barcode: #132786 - 0920504
Make: FAITH INNOVATIOS	Model: STERIFOG
Serial Number: NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: Ticket master Installation date=20/10/2014 ,As per stock book equipment installed on 07/05/2015, (stock book attached for reference)	Present status of the equipment: Fully damage
Date of breakdown:02/09/2024 Date of registration of complaint through email/ Toll free): Toll Free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the fumigator found not switching on. Problem with Vaccum motor, Nozzle pin and silicon o ring and Timer motor damaged. Need to replace these spare for further checking.	
Recommendations for repair (required service details): No recommendations	
Cost of spares (specify parts and cost): Vaccum motor = 8229.32/- Nozzle pin and silicone o ring=1625.45/- Timer motor = 2920.79/- TOTAL = 12775.56/-	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 61.32%	Asset Value:20833/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Service report and Quotation attached	
Reasons for recommending the equipment as BER: Checked the fumigator found not switching on. Problem with Vaccum motor, Nozzle pin and silicon o ring and Timer motor damaged. Equipment installed on 20/10/2014 As per stock book equipment installed on 07/05/2015 and aged up to 9 years 3 months. Repairing cost is 61.32 % both criteria for RBER met. As per tender clause 5.3.14.1 we recommending for condemnation.	
<div style="text-align: right;">  AKHILESH P 18/09/24 Name & Signature of CYRIX Authority with date </div>	
Remarks of Junior Consultant (Biomedical) NHM: Machine installed on 7/05/2015 and aged up to 9 years. Quotation attached. Repairing cost is 61.32% of asset value. Both criteria of RBER met.	
Recommended for BER (Yes/ No): <i>Yes</i>	<div style="text-align: right;">  Signature of JC BM (Junior Consultant (Biomedical Engg) NHM (Arogyakeralam), Palakkad) </div>
Date:	
Date:	Signature of Superintendent / Medical Officer (i/c)

*Not mandatory #Based on the period of life and value as per the BER guidelines * Attach Photograph




താലൂക്ക് ആസ്ഥാന ആശുപത്രി
മണ്ണാർക്കാട്



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1003891

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <u>THQH... MANNARKAD</u>	Call Registration Date : <u>2/19/2024</u>
Address: <u>MANNARKAD</u>	Caller ID : <u>132786</u>
<u>PALAKKAD</u>	Date of Visit : <u>3/19/2024</u>
Ph: <u>9446063003</u>	Asset No : <u>0920504</u>
	EQPT Name : <u>Fumigator</u>
	Manufacture <u>Earth Innovati</u> Model : <u>Sterifog</u>
	S. No. <u>NIL</u> Dept. <u>OT</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Machine not working

Action Taken : checked the machine found not switching on. problem with vaccum motor, Nozzle pin and Silicon oring and timer motor damaged. Need to replace these spore for further checking

Completed Date : 3/19/2024 Time : 11:30AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>SREETHI C</u>	<u>3/19/2024</u>	<u>10:00AM</u>	<u>11:30AM</u>

Customer Remark Completed Pending

Service Engineer Name : SREETHI

Signature: [Signature]

Date: 3/19/2024

Contact Number: 907252205

Customer Name : S keebak

Signature: [Signature]

Date: 3/19/2024

Contact Number: 944-6063003

Designation : S. N. D



സുലഭൻ
താലൂക്ക് ആസ്ഥാന ആശുപത്രി
മുല്ലാർക്കാട്

REF: BEC/CHPL/ 04/09-2024
09-09-2024

The Purchase Department
M/s. Cyrix Healthcare Private Ltd
Ernakulam

Pricing for Fumigation Dispensor parts

S.No	Description	Qty	Pricing INR (exc.GST)per piece	GST	Total Price (inc.GST)
1	Vaccum Motor	1.00	6974.00	18.00	8229.32
2	Nozzle pin and Silicone O Ring	1.00	1377.50	18.00	1625.45
3	Timer motor (Spring wound timer)	1.00	2475.25	18.00	2920.79
			Total		12775.56

TERMS & CONDITIONS

Taxes: As shown above

Payment : 50% advance with PO, balance COD/bank transfer against delivery

Delivery : 6-8 weeks from date of submitted quote

With Regards

Authorized Signatory

Biomedical Engineering Company (BEC)

Bank details :

M/s. Biomedical Engineering Company

HDFC Bank, Kadavanthara Branch ,IFSC code HDFC0000295

A/c: 99999388878006



**REQUIRED SPARE QUTATION FOR FUMIGATION DISPENSER 0920504 # 132786 @ THQH
MANNARKKAD , PALAKKAD**

4 messages

Akhilesh P <akhileshcyrix@gmail.com>
To: Jobby.bec@gmail.com, prabin@bechealthcare.com
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <Zm2.klbemp@cyrix.in>

Thu, 5 Sept, 2024 at 10:04 am

Dear sir
Please share the price and stock availability of fumigation motor, nozzle and timer for fumigation dispenser @ thqh
mannarkkad palakkad District
Details are furnished below

Equipment Name : FUMIGATION DISPENSER

Make and Model : faith innovation , sterifog

Barcode : 0920504

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

Akhilesh P <akhileshcyrix@gmail.com>
To: Jobby.bec@gmail.com, prabin@bechealthcare.com
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <Zm2.klbemp@cyrix.in>

Sun, 8 Sept, 2024 at 10:48 pm

Dear sir
Kindly update the status.

Thanks & Regards

AKHILESH P
DISTRICT INCHARGE - PALAKKAD
Bemp-Kerala
Mobile : +91 7593847154
Email : akhileshcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B, Petta, Punithura
Ernakulam- 682038
www.cyrixhealthcare.com

[Quoted text hidden]

Biomedical Engineering Company <becproductquotes@gmail.com>
To: Prabin Chandraprasad <prabin@bechealthcare.com>, akhileshcyrix@gmail.com
Cc: jobby.bec@gmail.com, BEC Mail <mail@bechealthcare.com>, sales@bechealthcare.com

Mon, 9 Sept, 2024 at 3:59 pm

Please find attached

Warm regards,
Shaji
Team BEC

On Sun, 8 Sept 2024 at 23:24, Prabin Chandraprasad <prabin@behealthcare.com> wrote:

Team,
FYA
Thanks & Regards

Prabin Chandraprasad
Manager Service & Operations | Team BEC
 [+919061082277](tel:+919061082277)
 [0484-2887200](tel:0484-2887200)
 prabin@behealthcare.com

Begin forwarded message:

From: Akhilesh P <akhileshcyrix@gmail.com>
Date: 8 September 2024 at 10:50:21 PM IST
To: jobby.bec@gmail.com, prabin@behealthcare.com
Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>
Subject: Re: REQUIRED SPARE QUTATION FOR FUMIGATION DISPENSER 0920504 # 132786 @ THQH MANNARKKAD , PALAKKAD

[Quoted text hidden]

Prabin Chandraprasad <prabin@behealthcare.com>
To: Akhilesh P <akhileshcyrix@gmail.com>

Tue, 10 Sept, 2024 at 11:44 am

FYI

Thanks & Regards

Prabin Chandraprasad
Manager Service & Operations | Team BEC
 [+919061082277](tel:+919061082277)
 [0484-2887200](tel:0484-2887200)
 prabin@behealthcare.com

Begin forwarded message:

[Quoted text hidden]

111


K. F. C. No. 23/271/2000. 1,500 x 100. G.P. S. Govt. of Kerala.

K. F. C. FORM 1B

[See Chapter VII, Article 161 (b)]

Stock Account of Furniture and other office stores

Name of article..... Fumigating machine Office of.....

Date	Number and date of contingent voucher, invoice, etc.	Nature of transaction	Receipts	Issues	Balance	Remarks
7/5/15	Recd from the Store ഉടമ 10 OT 		1		1 ✓	Block panchayat 1328,2000 1328,2000

സൂപ്രണ്ട്
താലൂക്ക് ആസ്ഥാന ആശുപത്രി
മണ്ണാർകട്



WARNING



Read and understand this manual



Manufactured by
Faith Innovations
Head Office : E 107, Lalpat Nagar - I, New Delhi
Phone : +91 11 41727403, 41726534 Fax : +91 11 41727403
e-mail : faithinnovations@gmail.com




PMS DONE ON
NEXT PMS DUE
PERFORMED BY

YRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225) 0920504



Handwritten signature and text in Malayalam: "സംസ്ഥാന ആരോഗ്യ വകുപ്പ്, കണ്ണൂർ ജില്ലാ ആശുപത്രി"