


PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: TH BEDADKA	Name of District: KASARGOD
Name of Equipment: NEBULIZER	Equipment ID & Barcode: 132282- 1420319
Make: NIDEK	Model: N/A
Serial Number: N/A	Warranty details: NO Warranty
Date of purchase/ Year of manufacture/ Installation Date: 03-06-2015	Present status of the equipment: Fully damaged
Date of breakdown: 29/08/2024 Date of registration of complaint through email/ Toll free): Toll free	*AMC/CAMC Period agreed at the time of purchase: NO AMC/ CAMC
Action taken: Checked the machine and found that Machine has no fumes out, Found compressor and piston are defective. Need to replace these spares for further checking.	
Recommendations for repair (required service details): No recommendations	
Cost of spares (specify parts and cost): N/A	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: N/A	Asset Value: 1353/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Service report attached	
Reasons for recommending the equipment as BER: Checked the machine and found that Machine has no fumes out, Found compressor and piston are defective. Equipment installed on 03-06-2015 aged up to 9 year 2 months. Quotation not submitted since spares are not available from the market. So we recommend the equipment for condemnation.	
ABHINAVO 11.09.2024 Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM: Checked the machine and found machine has no fumes out, compressor and piston are defective. Equipment installed on 3/6/15. As per cyrix, spares are not available from the market so quotation not submitted. So recommending the equipment for BER.	
Recommended for BER (Yes/ No):	Jr. CONSULTANT Biomedical Engineering Arogya Keralam (NHM) Kasargod at Kanhangad
Date: 18/09/24	
Date:	Signature of Superintendent Medical Officer (i/c) 18/09/24 671541



Based on the period of life and value as per the BER guidelines. * Attach Photograph.




ബയോമെഡിക്കൽ എഞ്ചിനീയറിംഗ് വിഭാഗം പി.ഒ.-671541

NEBULISER & mask + O₂ mask paediatric

Date	Received from or issued to	Receipt	Issued	Balance	Remarks
5.11.12	Found in ward	(1)		1	
5.11.12	Issued to ward		1	Nil	
15/1/13	Nebuliser mask (Adult-1) Paediatric-1 Received from store	Adult-1 Paediatric-1		Adult-1 Paediatric-1	
15/1/13	Issued to ward nebuliser mask Adult-1 Paediatric-1		Adult-1 Paediatric-1	Nil	
6/1/13	Issued for repairing (nebuliser) Received from DMO	1	1	Nil	
7/1/11	Received from palliative OP nebuliser & mask & T. piece accessories. (adult-1) paediatric-1) mask	(1)		-2	ally
7/6/13	Transferred to unserviceable Page No: 13		1	1	
4/10/13	Adult nebulising mask & chamber (Aeromist adult Romsons)	1		1	
4/10/13	Paediatric nebulising mask & chamber (Aeromist child Romson)	1		1	
4/10/13	Issued to ward (adult mask)		1	Nil	
4/10/13	Issued to ward (paediatric mask)		1	Nil	
3/6/15	Received from store Issued to ward nebuliser	2	2	2, 3	
15.2.15	Received from store	1, 4		1	
11.2.15	Issued to ward		1	Nil	
4.4.15	Received from store (nebuliser)	2, 1		2	
4.4.15	Issued to ward		2	Nil	
26.5.15	Received from store nebu. mask Adult - 6 " mask Paediatric - 4	10		10	
26.5.15	Issued to ward & OP session.		10	Nil	
28/5/15	Transferred to unserviceable Page No: 17 (Nebuliser)	2		4, 2, 1, 1	
	Issued to ward		4	Nil	
23/5	Received from store (mask) Issued to ward (mask)	10	10	Nil	
15/1/20	Transferred to unserviceable page No: 20 (Record on 3/6/15)	1		8, 2, 3	
		Nebuliser - Balance - 2			
31/8/21	Received from store (O ₂ mask)	5		5	
31/8/21	Issued for use Pareshwari (15)			Nil	



കമ്മ്യൂണിറ്റി മെഡിസിൻ
 മെഡിക്കൽ ഓഫീസ്
 പിൻ 671541



താലൂക്ക് ആശുപത്രി ബോധധൂക്കുകൾ
 കാസർഗോഡ് ജില്ല
 പിൻ 671541
 തിരുച്ചി.....
 ബോധധൂക്കുകൾ പി. ഒ.

66
 19/9/24
 മെഡിക്കൽ ഓഫീസർ
 താലൂക്ക് ആശുപത്രി ബോധധൂക്കുകൾ
 ബോധധൂക്കുകൾ പി. ഒ.-671541