


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

456

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	Trivandrum.
2	Name of Hospital	District Model Hospital, Perootkada, Trivandrum.
3	Name of Equipment with Make, Model and Serial Number	Nebulizer. Ready Mist. Sl. no :- 187355.
4	Equipment ID & Barcode	97512 & 0113929.
5	Date of purchase / Year of manufacture /Installation Date	12/04/2016.
6	Warranty details (Yes/No)	NA
7	*AMC/ CAMC Period agreed at the time of purchase	NA
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	22/02/2024.
9	Action taken	Checked the Machine. Problem found in the motor side. Motor and pistons are damaged.
10	Present status of the equipment (Fully damaged / partially damaged)	fully Damaged.
11	Recommendations for repair (required service details)	Not recommending for repair because, the motor is stuck and piston damaged. Motor parts are rusted.
12	Cost of spares (specify parts and cost)	Not repairable.

13	Asset Value	NA
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Attached .
16	Reasons for recommending the equipment as BER	Motor is stucked and rusted. Piston is damaged. It is not repairable condition. Hence the machine is recommended for RBER.
17	Name & Signature of CYRIX Authority	Akshai Manohar, A.M 


\*Not mandatory #Based on the period of life and value as per the BER guidelines


\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Motor fully rusted and piston damaged.

**MANEESHA MOHAN M.R.**  
Junior Consultant (Biomedical)  
National Health Mission  
Thiruvananthapuram

  
28-02-2024  
Signature of JC BM (NHM)

  
DISTRICT MODEL HOSPITAL  
PEROORKADA, THIRUVANANTHAPURAM  
PIN-695 005.

Date 15/4/2024

  
Signature of  
Superintendent / Medical Officer  
District Model Hospital  
Peroorkada  
Thiruvananthapuram

Seal





**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

**CYRIX**  
HEALTHCARE PVT LTD

No. : 179653

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>DMH</u>	Call Registration Date : <u>22-02-2024</u>
Address : <u>Perinthala</u>	Caller ID : <u>97512</u>
<u>Trivandrum</u>	Date of Visit : <u>26-02-2024</u>
Ph : <u>9074678136</u>	Asset No. : <u>0113929</u>
	EQPT Name : <u>Nebulizer</u>
	Manufacture <u>Ready Mist</u> Model : <u>Ready Mist</u>
	S. No. <u>187355</u> Dept. <u>Old pay ward</u>

Service Classification : Breakdown Call  PMS  Calibration  Cust.Training

Problem Identified : Not working

Action Taken : Checked the machine. Problems found in the motor side. Serviced the machine several times. Piston and motor is damaged and the motor parts are rusted. It's not in a repairable condition.

Completed  Date : 26/02/24 Time : 10:30 AM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.	—	—	—
2. <u>NA</u>	—	—	—
3.	—	—	—

Cyrix Engineer	Date	Start Time	End Time
<u>Akshay</u>	<u>26/02/24</u>	<u>10:00 AM</u>	<u>10:30 AM</u>

Customer Remark  Completed  Pending



Service Engineer Name : Akshay  
Signature : [Signature]  
Date : 26/02/24  
Contact Number : 8921568751

Customer Name : For Susen baby  
Signature : Shreeya  
Date : 26/02/24  
Contact Number : 9074678136  
Designation : Shreeya  
Hospital Seal : SNO, 26/2/24



