

464

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GH Neyyattinkara
3	Name of Equipment with Make, Model and Serial Number	Equipment: Digital BP Apparatus Make: Ross max Model : Na Serial Number : Na
4	Equipment ID & Barcode	97989& 0113157
5	Date of purchase / Year of manufacture /Installation Date	07/11/2018
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	24/02/2024
9	Action taken	Checked and found mother board and display defective.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for service
12	Cost of spares (specify parts and cost)	NA

13	Asset Value	Rs. 1353/-
14	#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found mother board and display defective. Spares are not available in the market. Equipment completed 5+ year. So we recommended the equipment for condemnation.
17	Name & Signature of CYRIX Authority	Ajith George

\*Not mandatory **#Based on the period of life and value as per the BER guidelines**


\* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Mother board defective.

So RBER.



**MANEESHA MOHAN M.R.**  
Junior Consultant (Biomedical)  
National Health Mission  
Thiruvananthapuram

  
20.03.2023

Signature of JC BM (NHM)

Date

Seal

  
**SUPERINTENDENT**  
GENERAL HOSPITAL  
MEYYATHIPPURAM-695024  
  
Signature of  
Superintendent / Medical Officer (i/c)



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

# CYRIX

HEALTHCARE PVT LTD

No. : 178052

464

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

## Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrrix.in

Health Facility : <u>GH Meyyattinkara</u>	Call Registration Date : <u>24.2.2024</u>
Address : <u>Trivandrum</u>	Caller ID : <u>97989</u>
Ph : <u>8921163587</u>	Date of Visit : <u>24.2.2024</u>
	Asset No. : <u>0113157</u>
	EQPT Name : <u>B.P. Apparatus</u>
	Manufacture : <u>ROSSmax</u> Model : <u>NA</u>
	S. No. : <u>NA</u> Dept. : <u>OP</u>

Service Classification : Breakdown Call  PMS  Calibration  Cust.Training

Problem Identified : machine not getting on

Action Taken : checked and found mother board and display defective. need to replace the mother board and display for further checking and working condition of the equipment

Completed  Date : 24.2.2024 Time : 10.30 am Spare Required

Spare Replaced  Requested


Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrux Engineer	Date	Start Time	End Time
<u>Ajith George</u>			

Customer Remark

Completed

Pending



**SUPERINTENDENT**  
GENERAL HOSPITAL  
MEYVATTINKARA-695121

Service Engineer Name : Ajith George  
Signature : A  
Date : 24.2.2024  
Contact Number : 7593847111

Customer Name : Sambhija.V.P.  
Signature : Sadhya  
Date : 24/2/2024  
Contact Number : 9567114737  
Designation : Nursing officer nrb  
Hospital Seal :



# INDEX

	Page No		
1. Iron cot	9	39. Window bars	45
2. Wooden cot	13	40. Vaginoscopes	51
3. Labour cot	19	41. Valsalar	57
4. Stool Square top	25	42. Uterine rod	75
5. Table operation ordinary	27	43. Examinaible	1
6. Table 2 drawer	21	44. Matress	5
7. Table prescribing	31	45. Sponge holder Forceps	63
8. Aluminium Steel	35	46. 2 Cell medium mesh	119
9. Sheet	41	47. Spot light	33
10. Gowel	47	48. Forceps Alice's Pat	
11. Moping stick nil	53	49. X ray view	163
12. B.P. Apparals	59	50. Plastic culture jar	169
13. Vaccine Carrier nil	61	51. Needle Holder	75
14. Artery Forceps	37	52. Needle Holder "	77
15. Forceps Mosquito curved	62	53. Kerosine stove	40
16. Forceps Mosquito straight	49	54. E.C.G. Machine	17
17. Cheable Forceps	51	55. Syringe and Needle destroyer.	181
18. Scissors Mayo curved	63	56. Irrometer	187
19. Scissors Mayo straight	65	57. Bench 2 Back	45
20. Episiotomy Scissors Angular	67	58. Chair Armless	91
21. Scissors Shop	69	59. Arm chair	94
22. Alice Forceps Binched	71	60. Table steel (med)	105
23. Locker Iron	17	61. Aluminium Steel	
24. Saline Stand	77	62. Table steel	
25. S.S. Bowl	79	63. Small tray 2 lid	39
26. S.S. Basin	81	64. Scissors Suture Remover	75
27. Basin E.I	83	65. Com. flat steel	
28. Dressing DRUM	85	66. Scissors (Suture cut)	
29. Lock 6 lever	57	67. Froge	24
30. Lock Naval 6 lever	55	68. wheel chair	85
31. O2 Cylinder and Assurin }	45	69. Bucket (plastic)	165
32. Post merham set	115	70. Steel stool	111
33. Kidney Tray S.S	87	71. chair 'S' shaped iron	140
34. Sterilizer Electric	93	72. Table 2 drawers	105
35. Square Tray S.S	21	73. Instrument trolley	22
36. Sauce pan S.S	127	74. Plastic chair 2 corner	107
37. Hamar		75. Steel Raulc open	111
38. Wall clock		76. Revolving stool	113




NEYYATTINKARA  
SUPERINTENDENT  
GENERAL HOSPITAL  
NEYYATTINKARA - 695129



58  
Digitel B.P Apparatus

Date	Description	Reel	Time	Balance	S/P
11/1/23	Reel in store	5		5	
19/6/23	Received manifest of bone wood	5		No. 1	Spec.



*Te*  
**SUPERINTENDENT**  
 GENERAL HOSPITAL  
 NEYYATTINKARA-695124

ROSSMAX



SYS  
mmHg

DIA  
mmHg

PULSE  
/min



Complaint  
display



**CYRIX**  
HEALTH CARE PVT LTD

BIOMEDICAL EQUIPMENT MAINTENANCE  
PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669



BAR CODE - (8004-890615225)  
0113157

