



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 179260

477

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>Chc poovai</u>	Call Registration Date : <u>27/3/24</u>
Address : <u>Iravandrum</u>	Caller ID : <u>103602</u>
Ph : <u>6282656110</u>	Date of Visit : <u>27/3/24</u>
	Asset No. : <u>0131196</u>
	EQPT Name : <u>Nebuliser</u>
	Manufacture <u>Life care</u> Model : <u>Na</u>
	S. No. <u>Na</u> Dept. <u>Op</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : Checked and found piston broken and motor damaged. Need motor and piston for further checking of the equipment.

Completed Date : 27/3/24 Time : 4:30pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Salika S</u>	<u>27/3/24</u>	<u>4pm</u>	<u>4:30pm</u>

Customer Remark Completed Pending

Service Engineer Name : Salika S
Signature : [Signature]
Date : 27/3/24
Contact Number : 6282656760



Customer Name : Rama
Signature : [Signature]
Date : 27/3/24
Contact Number : 0471-2211420
Designation : Nursing officer
Hospital Seal : [Seal]


സംസ്ഥാന ആരോഗ്യ മന്ത്രിയുടെ കാര്യാലയം
ആരോഗ്യ സൗകര്യ കേന്ദ്രം, പൂനത്തൂർ

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	CHC POOVAR
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : NEBULISER Manufacturer : LIFECARE Model : NA Sl: 1486
4	Equipment ID/ Barcode	103602/ 0131196
5	Date of purchase/ Year of manufacture/Installation Date	24/06/2016
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	27/03/2024 (Toll free)
9	Action taken	Checked and found Nebulizer motor and piston damage. Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not recommending for Repair
12	Cost of spares (specify parts and cost)	Not available

13	Asset Value	Rs . 1353/-
14	Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Na
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment RBER	Checked and found nebulizer motor with piston defective. Quotation not submitted since the spares are not available in the market. The equipment was installed on 24/06/2016 and covered upto 8+years, So Recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	Kasyep PV 

*Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Motor and Piston defective.
As per the tender conditions, this can be RBER.

MANEESHA MOHAN M.R.
J MANEESH
 Junior Consultant
 Tiruvananthapuram
 Tiruvananthapuram


 10-6-2024
 Signature of JC BM (NHM)

Date:




 Signature of
 Superintendent/Medical Officer (i/c)

മെഡിക്കൽ ഓഫീസർ നൂൽ ചാർജ്ജ്
 നാമുഹിത ആരോഗ്യ കേന്ദ്രം പട്ടണൂർ

Name of drug **NEBUKIZER**



Name of drug

Date	Receipt	Issue	Balance	Date	Receipt	Issue	Balance
31/3/15	Physically found	1					
16/6/15	verified found	3					
24/6/16	2	5					
24/1/2023							1
17/8/2023		Recd from Store			Tracked	1	2
							1
							2



[Signature]

മെഡിക്കൽ സൂപ്രണ്ടർ തൃശ്ശൂർ ജില്ലാ ആരോഗ്യ കോശം, പട്ടാമ്പി



മെഡിക്കൽ സപ്ലൈസ്
സാമൂഹിക ആരോഗ്യ കേന്ദ്രം, പുറവാട