



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 223571

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <i>General Hospital</i>	Call Registration Date : <i>3/4/24</i>
Address <i>Neyyattinkara</i>	Caller ID : <i>104588</i>
<i>Travancore</i>	Date of Visit : <i>4/4/24</i>
Ph : <i>9402169162</i>	Asset No. : <i>0115441</i>
	EQPT Name : <i>Monitor (LCD)</i>
	Manufacture <i>Sony</i> Model : <i>KLV 22V320</i>
	S. No. <i>203325</i> Dept. <i>O.T.</i>

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : *Display not getting on.*

Action Take : *checked and found display defective. Need to replace the spare for further checkage and making condition of the equip. ment. Explained spare with OEM.*

Completed Date : *4/4/24* Time : *2pm* Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrlix Engineer	Date	Start Time	End Time
<i>Dish</i>	<i>4/4/24</i>	<i>1pm</i>	<i>2pm</i>

Customer Remark *Pending*



Service Engineer Name : <i>Dish</i>	Customer Name : <i>SHIMIZU</i>
Signature : <i>[Signature]</i>	Signature : <i>[Signature]</i>
Date : <i>4/4/24</i>	Contact Number : <i>414124</i>
Contact Number : <i>9593849111</i>	Designation : <i>Nursing officer</i>
	Hospital Seal : <i>8281717712</i>

SUPERINTENDENT
GENERAL HOSPITAL
NEYYATTINKARA-695121


532

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GENERAL HOSPITAL NEYYATINKARA
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : Monitor (LCD) Make : Sony Model : KLV-22BX320 Serial No : 7033255
4	Equipment ID/ Barcode	104588/0115441
5	Date of purchase/year of manufacture/installation date.	MAY 2011
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agree at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	03/04/2024
9	Action taken	Checked and found display defective. Need to replace the spare for further checking and working condition of the equipment. Enquired spare with vendor.
10	Present status of the equipment(Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	Sony Display - 14750/-

13	Asset Value	13500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	109.2%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and vendor quotation attached
16	Reasons for recommending the equipments as BER	Checked and found display defective. The unit is manufactured on may 2011 and covered upto 13+ years. The repairing cost is 109.2% . As per the tender clause 5.3.14.1 both crieteria met. So we recommend the unit for condemnation.
17	Name & Signature of CYRIX Authority	KASYEP PV 

*Not mandatory


#Based on the period of life and value as per the BER guidelines


* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Estimate ~~is~~ verified and as per tender conditions

MANESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


4.07.2024
Signature of JC BM (NHM)

Date: 


Signature of
Superintendent/Medical Officer (i/c)
SUPERINTENDENT
GENERAL HOSPITAL
NEYYATTINKARA-695121

SYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225) 0115441



SONY LCD COLOUR TV


MODEL NO. KLV-22BX320

MAY 2011

110 - 240 V ~ 50/60 Hz 55W

SER. NO 7033255

VIDEO NICKEL



MODEL NO. KLV-22BX320

SERIAL NO. 7033255

4 14 09 (264) 0 (12 1) 1173

SUPERINTENDENT
GENERAL HOSPITAL
NEYYATTINKARA





REGARDING QUOTATION - 104588 - 0115441 GH NEYYATTINKARA - MONITOR(LCD-TV)

3 messages

Abijith A <abijithacyrix@gmail.com>
To: Athulraj m <medplusbiomedicals.clt@gmail.com>
Cc: Kasyep P V <kasyepcyrix@gmail.com>, Afsal Yunuze <zm1.klbemp@cyrix.in>

Tue, 14 May 2024 at 10:53 am

Dear Sir,
kindly share a quotation for the Display panel.

Make - Sony
Model - KLV-22BX320
Serial no - 7033255

Thanks & Regards,

Abijith A
Divisional Incharge
Mob-7593847141
abijithacyrix@gmail.com
www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd
| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India
KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

Abijith A <abijithacyrix@gmail.com>
To: Athulraj m <medplusbiomedicals.clt@gmail.com>
Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>, Kasyep P V <kasyepcyrix@gmail.com>


Wed, 15 May 2024 at 9:55 pm

Reminder
[Quoted text hidden]

Athulraj m <medplusbiomedicals.clt@gmail.com>
To: Abijith A <abijithacyrix@gmail.com>
Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>, Kasyep P V <kasyepcyrix@gmail.com>

Wed, 15 May 2024 at 10:20 pm

[Quoted text hidden]



MEDPLUS BIOMEDICALS

Athulraj M
Kumaraswamy raja road
Chelannur (P.O)
Kozhikode

☎ 8848372192 ✉ medplusbiomedicals.clt@gmail.com
GSTIN 32CEYPA2140F1ZV

Quotation

To,
Cyrix Healthcare
Manager
Kochi

Quotation# Quote-14

Date: 14/05/2024

Dear Sir/Mam,

We are pleased to submit the purchase order as below.

#	DESCRIPTION	QTY	PRICE	GST	TOTAL
1	Sony Display	1 unit	₹ 12,500.00	₹ 2,250.00 18.0%	₹ 14,750.00
SUB TOTAL					₹ 12,500.00
GST					₹ 2,250.00
GRAND TOTAL					₹ 14,750.00

Your prompt attention to this order is greatly appreciated, and we look forward to a successful transaction.

Terms & Conditions:

- Quotation validity : one month
- 100 % advance
- Delivery : 7 days after confirmation
- Transportation charges included

For, MEDPLUS BIOMEDICALS

AUTHORIZED SIGNATURE