



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 206006

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>GH</u></p> <p>Address <u>Thiruvananthapuram</u></p> <p>Ph : <u>944 7465968</u></p>	<p>Call Registration Date : <u>09-04-2024</u></p> <p>Caller ID : <u>105679</u></p> <p>Date of Visit : <u>10-04-2024</u></p> <p>Asset No. : <u>0110800</u></p> <p>EQPT Name : <u>Nebuliser</u></p> <p>Manufacture <u>Life Care</u> Model : <u>CNB69009</u></p> <p>S. No. <u>LC09191160151</u> Dept. <u>ward 1</u></p>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Low fumes

Action Taken : checked the machine and found piston with motor damaged, need to replaced these spares for further checking of equipment

Completed Date : 10/04/24 Time : 12:00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>NA</u>			
3.			

Cyrux Engineer	Date	Start Time	End Time
<u>ADITHYA.S.S</u>	<u>10/4/24</u>	<u>11:30 am</u>	<u>12:00 pm</u>
<u>Stebin, Deepak</u>			
Customer Remark	Pending		



Service Engineer Name : ADITHYA.S.S
Signature : [Signature]
Date : 10-04-24
Contact Number : 9745084344

Customer Name : Celine Keena
Signature : [Signature]
Date : 10/4/24
Contact Number : 9961391392
Designation :
Hospital Seal :


[Handwritten notes and stamps]

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No.	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GENERAL HOSPITAL THIRUVANTHAPURAM
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : NEBULIZER Manufacturer : LIFECARE Model : CNB69009 Sl: LB091911160158
4	Equipment ID/ Barcode	105629 / 0110800
5	Date of purchase/ Year of manufacture/ Installation Date	19/10/2017
6	Warranty details (Yes/No)	No Warranty
7	AMC/ CMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	09/04/2024(Toll free)
9	Action taken	Checked and found Nebulizer motor and piston damage. Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not recommending for Repair
12	Cost of spares (specify parts and cost)	Not available

13	Asset Value	Rs . 1027/-
14	Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Na
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment RBER	Checked and found nebulizer motor with piston defective. The equipment was installed on 19/10/2017 and covered upto 6+years. Quotation not submitted since the spares are not available in the market., So Recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	Kasyep PV 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Motor with piston defective. So as per tender condition can be RBER

MANEESHA MOHAN M.R.
 Junior Consultant (Biomedical)
 National Health Mission
 Tiruvananthapuram

 10.07 - 2024
 Signature of JC BM (NHM)

Date:


 Signature of Superintendent/Medical Officer (Uc)



Superintendent
 General Hospital
 Tiruvananthapuram

		R	F	B	S/F
15/6/17	RF from abl RFR P-NO 183 (OAP - Not known)			2	10
19/10/17	Received from the store 19/10/17 Page no: 8B-110 Rs 1027.50. Wasimoly nil	1		3	2/10/17
28/11	Rd from store (Page no: 110. TR-8B) Rg	1		4	
5/9/19	(Rs-1902.32+Tax) Rd from stock condemnation list on no per order NOC-226/2022 CHS (2019 June) dated-26/7/22	2	3	5	3
15/11/2019					
8/4/2021					
23/3/22	Rs. 10/- from wheel etc (stock up) Vol- 01 Page no 12	1		1	
31/5/22	R.F.S (LP) 600000 B-NEB-03-188y-d-1/1/22-Exp-1/1/2030	2		6	
28/1/22	RFS (LP) Harden tech put (AM) mfg no (Order 0570)				
	mfg 1/6/22 Exp - 1/6/30	1		7	
8/2/23	Received from store (LP) [Frucbek]	1			
(HS)	Batch NO 20220826 (A) 20/01/11/2022 Exp date 1/11/22 Rs 1027.50/-	1		8	8/2
	Condemnation done as per order No: G-5549/23/GHS Dated 13/2/24				
2/4/22	Returned to store			4	dm
				6	

DICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)



FREE NO: 1800-425-7669

BAR CODE - 0110800



lifecare



Compressor Nebulizer
Model No.: CNB69009
Input : A.C-220V/50HZ
Power : 180VA
Serial No.: LG 091911160138
Mode of operation-continuous
No AP/APG IPX0



