



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. :
224127

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : THQH	Call Registration Date : 10-04-2024
Address : Vakkala	Caller ID : 105711
Thiruvananthapuram	Date of Visit : 11-04-2024
Ph : 9074748206	Asset No. : 0121137
	EQPT Name : Dental chair
	Manufacture : Confident Model : na
	S. No. : na Dept. : Dental

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : checked and found that the unit is completely Rusted, leading to the separation of the base and causing the chair to unstable

Completed **Date :** 11/04/24 **Time :** 12-Pm **Spare Required**

Spare Replaced **Requested**

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Vishnu.K	11/04/24	10:00am	12:00pm

Customer Remark Completed Pending


Service Engineer Name : Vishnu.K Signature : Date : 11/04/2024 Contact Number : 9605884413	Customer Name : V. Deepu, Superintendent Signature : Date : 11-4-2024 Contact Number : 9446846350 Designation : Assistant Dental Surgeon Hospital Seal :
---	---

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	THQH VARKALA
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : DENTAL CHAIR Make : Confident Model : NA Serial No : NA
4	Equipment ID/ Barcode	105711/0121137
5	Date of purchase/ Year of manufacture/Installation Date	1982
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	10/04/2024 (Toll free)
9	Action taken	Checked and found that the unit is completely rusted, leading to the separation of the base and causing the chair to be unstable.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	NOT AVAILABLE

13	Asset Value	65000/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Not Available
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT AND HOSPITAL SIDE LETTER ATTACHED
16	Reasons for recommending the equipment as BER	Checked and found the chair is fully rusted, leading to the separation of the base causing the chair to be unstable. The unit is installed on 1982 and outlived upto 42 years. The chair is very old and not suitable for the daily use by considering patient safety.. Hospital side letter requesting for condemnation attached. So recommended for condemnation.
17	Name & Signature of CYRIX Authority	KASYEP PV 

*Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Machine too old, fully rusted (42 years old)
So can be RBER.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram




10.7.2024

Signature of JC BM (NHM)



Date:


Signature of
Superintendent/Medical Officer (IC)
SUPERINTENDENT
TALUK HEADQUARTERS HOSPITAL
VARKALA

DENTAL CHAIR WITH OPERATOR STOOL & Compressor

Date	From whom received	Receipt	Issue	Balance	Initial
	Brought forwarded from stock register of Dental Instruments Vol. I - 1982 Page No. 47			2	(1 not working)

Condemned and auctioned as per the order no. C3/32031/08/DMOH dated 21.12.08 (1 Dent chair)

27/7/22
 from store
 Trichur Surgicals, Trissur
 KMC order no: 50/695/2018 - 2019/799 dated 08/11/2018
 Detail 8/10/2018 Dts 2017-18 TS 431
 Fully loaded Dental chair electrically open
 model of make : TS 431
 Accessories:
 Airlet Hand piece - 2 no
 Curve Angle Hand piece - 1 no
 Straight Hand piece - 1 no
 Maxwrtin cable - 1 no
 Ultrasonic scaler - 1 no
 Inbuilt Light cure - 1 no
 Dental Air Compressor 1 no
 warranty - 27/7/22 to 26/7/2024
 Rs. 3,64,710.00/-

Page no: Received from store stock Register for Instruments and Compressor - Vol. I Page no: 237



11/4/24
 SUPERINTENDENT
 TALUK HEADQUARTERS HOSPITAL
 VARKALA

Taluk Headquarters Hospital, Varkala

Temple Road, Varkala

E-mail : thqvarkala@gmail.com

Phone : 0470-2080088

Date : 18.06.2024

No.B1-2056/24/THQHV

From
Superintendent

To
Bio Medical Engineering (BME)
Thiruvananthapuram.


Sir,

Sub:-THQHV-Condemnation of Dental chair at Taluk Headquarters Hospital,
Varkala.


Ref:- Letter received from Kasyep PV District in-charge, Tvpim Cyrix Health
Care Ltd Dated, 29.05.24.

As per records (Main Stock Register -Dental Page No.36) the Dental Chair
was received in 1982 (Register for dental instruments Vol.I Book No.I Page No.47).
This is for your kind information and necessary action regarding for condemnation of
dental chair.

Yours faithfully,


19.6.24
Superintendent.

Copy to :- File /stock file




11/4/24

SUPERINTENDENT
TALUK HEADQUARTERS HOSPITAL
VARKALA



wd: condemnation of dental chair

message

ME Trivandrum <bmetrivandrum1@gmail.com>
): Kasyep P V <kasyepcyrix@gmail.com>

Wed, 19 Jun 2024 at 4

PFA

Thanks and Regards

Maneesha Mohan M R
Junior Consultant(Biomedical)
NHM, TVPM



----- Forwarded message -----

From: **TALUK HOSPITAL VARKALA** <thqvarkala@gmail.com>
Date: Wed, 19 Jun 2024 at 14:44
Subject: condemnation of dental chair
To: <bmetrivandrum1@gmail.com>

Sir, Please see the attachment.




[Save a tree. Don't print unless it's really necessary!]

Office of the Superintendent
Taluk Hospital Varkala
Thiruvananthapuram
PIN: 695141
Landline: 0470-2602549



Handwritten signature
11/4/24

SUPERINTENDENT
TALUK HEADQUARTERS HOSPITAL
VARKALA



room Shot on realme 7i



Gow
11/4/24

SUPERINTENDENT
TALUK HEADQUARTERS HOSPITAL
VARKALA