



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 206014

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>General Hospital</u></p> <p>Address <u>Thiruvananthapuram</u></p> <p>Ph : <u>7012896199</u></p>	<p>Call Registration Date : <u>13-04-2024</u></p> <p>Caller ID : <u>106175</u></p> <p>Date of Visit : <u>14-04-2024</u></p> <p>Asset No. : <u>0115893</u></p> <p>EQPT Name : <u>BP Apparatus</u></p> <p>Manufacture <u>Draminel</u> Model : <u>Dentore</u></p> <p>S. No. <u>NA</u> Dept. <u>Female Surgical Ward</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Cuff not inflating
Equipment name is Suction ~~rese~~ apparatus.

Action Taken : Checked the machine and found pump control board and solenoid valve damage - Need to replace these spares for further checking of equipment


Completed Date : 14/04/24 Time : 11:00 am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>NA</u>	-	-	-
3.			

Cyrix Engineer

ADITHYA-J.S



Date Start Time End Time


14/04/24 10:30 am 11:00 am

Customer Remark

Completed Pending

Service Engineer Name : ADITHYA-J.S
Signature : [Signature]
Date : 14/04/24
Contact Number : 9745054344

Customer Name : [Signature]
Signature : [Signature]
Date : 14/4/24
Contact Number : 7012896198
Designation : Nursing Officer Cr 2
Hospital Seal : [Signature]




513

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GENERAL HOSPITAL THIRUVANANTHAPURAM
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : BP APPARATUS Make : Diamond Model: Deluxe SI : NA
4	Equipment ID/ Barcode	106175 / 0115893
5	Date of purchase/ Year of manufacture/Installation Date	16/03/2017
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CMC
38	Date of breakdown (Date of registration of complaint through email/ Toll free)	13/04/2024
9	Action taken	Checked and found mercury spillage. Identified mercury tank,glass rod,control valve defective.Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not recommending for Repair.
12	Cost of spares (specify parts and cost)	Not available

13	Asset Value	Rs . 1323/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	NA
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report Attached
16	Reasons for recommending the equipments as BER	Checked and found mercury spillage, Identified mercury tank, control valve and glass rod defective. The apparatus was installed on 16/03/2017 and covered upto 7+ years. Quotation not attached Since mercury is not available in the market, recommending the unit for condemnation.
17	Name & Signature of CYRIX Authority	KASYEP PV 

*Not mandatory

#Based on the period of life and value as per the BER guidelines


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Mercury Spillage.

So the BP apparatus can be RBER.


MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


13-06-2024

Signature of JC BM (NHM)




Date:


Signature of
Superintendent
General Hospital
Thiruvananthapuram



74 Digitas BP Apparatus.

Date	Particulars.	R/d	Issued	Balance	Remarks
22/11/19	R/d from store (True check)	1		1	
22/2/21	R/d from store [Elko] (Exp Date 28/12/2030)	1	✓	2	By H/W
31/3/21	(L.P) Received from store Tris from local purchase Register page no. 26 (RS. 1201.46)	1		3	Loss 5/12
<u>Life time</u> <u>Exp date</u> <u>30/6/2030</u>	model B5 x 575 Lot no S/SCIBH19				
<u>Batch no</u>	BH 19 - model B5 Bar code 0115803 1) 0115816 2) 0116082 3) 0110572				
20/10/2022	Physically found ⁽¹⁾ BPL 120/80 B9 _{batch no}			4	By S/W



[Signature]
Superintendent
General Hospital
Thiruvananthapuram

