



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 205194

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Service Report

Health Facility : <u>THQ#</u>	Call Registration Date : <u>15/04/24</u>
Address : <u>THQ# Karunagappally</u>	Caller ID : <u>106311</u>
<u>KOLLAM</u>	Date of Visit : <u>17/04/24</u>
Ph : <u>8590291706</u>	Asset No. : <u>0220478</u>
	EQPT Name : <u>Micrometer</u>
	Manufacture <u>Morathon</u> Model : <u>Morathon 4</u>
	S. No. <u>1923108</u> Dept. <u>Dental</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Micrometer Transformer and hand piece complaint.

Action Taken : checked the micrometer and found that the micrometer transformer and micrometer hand piece defective. Enquired spare with OEM.

Completed Date : 17/04/24 Time : 11:30am Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time

Customer Remark Completed Pending

Superintendent
Taluk Head Quarters Hospital
Karunagappally, Kollam



Service Engineer Name : Aiswarya
Signature : [Signature]
Date : 17/04/24
Contact Number : 9645384968


Customer Name : [Signature]
Signature : [Signature]
Date : 17/4/24
Contact Number : 8521132933
Designation : [Signature]
Hospital Seal : [Signature]

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Kollam
2	Name of Hospital	THQH KARUNAGAPPALLY
3	Name of Equipment with Make, Model and Serial Number	MICROMOTOR Make : MARATHON Model : MARATHON SI no: 1923108
4	Equipment ID & Barcode	Barcode : 0220478 / Ticket id : 106311
5	Date of purchase / Year of manufacture /Installation Date	03-02-2017
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	15-04-2024
9	Action taken	Checked the machine found that the Transformer & Micro motor hand piece defective, Enquired spares with OEM
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not Recommended for Repair
12	Cost of spares (specify parts and cost)	Lab micro motor Transformer = 1770 Lab micro motor hand piece = 4720 Total = 6490 /-


13	Asset Value	Rs. 7142.86 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	90.87 %
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix Service report & OEM Quotation attached
16	Reasons for recommending the equipment as BER	Checked the machine found that the Transformer & Micro motor hand piece defective. Equipment is installed on 03-02-2017 and aged up to 7 + years. .Repairing cost of equipment is 90.87%. as per tender clause 5.3.14.1 Both criteria met So we are recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AISWARYA RAJ 

*Not mandatory #Based on the period of life and value as per the BER guidelines

* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Repairing cost is high. So recommended for Beyond economic repair after verified quotation call.


 02/06/24
 Signature of JC BM (NHM)
 ABMO
 NHM, KOLLAM

Date

Seal




 Superintendent
 Karunagappally Hospital
 Karunagappally, Kollam


STOCK REGISTER

117

Name

Name	Received from		or issued to		Rs	P	Receipts	Issues	Balance	Remarks
				LEIB MICROMETER						
3.2.17	Inv. 85 dt 16.2.16						1		1	
	Dent Smile, KNPLY Technodont Marathon 37L (7142.86)									
	Confident Dental Equipments									
	Cochin. CKM CL supply.									
	Order no. 50/400/2016-17									
	dt. 17/11/17 (Rs. 51975)									
	Main Register Page no. 57									
28.9.19	Received from state						1			3
	In. no. 1281 Dent Link, KNPLY									
	dt. 26.9.19 SONO E 3486/19									
	dt. 14.8.19. (Marathon)									
	Micro meter lab)									




 Superintendent
 Taluk Headquarters Hospital
 Karunagapally, Kollam

Estimate



MEDIDENT




25/51 KODIYERI, THALASSERIKANNUR - KERALA
 Phone no.: 9778129877 Email: medident58@gmail.com
 GSTIN: 32EMXPR1675D1ZM, State: 32-Kerala
 MD42: MD KER 13 22 004

Estimate For	Estimate Details
Cyrix healthcare pvt ltd 1ST FLOOR 30/641B CYRIX HEALTHCARE PVT LTD KOCHI DHANUSHKODI ROAD POONITHURA GSTIN : 32AAFCC2499H2ZM State: 32-Kerala	Estimate No. : MED25 Date : 28-05-2024 Place of supply: 32-Kerala

#	Item name	HSN/ SAC	Quantity	Unit	Price/ Unit	GST	Amount
1	LAB MICROMOTOR TRANSFORMER		1	Nos	₹ 1500.00	₹ 270.00 (18%)	₹ 1770.00
2	LAB MICROMOTOR		1	Nos	₹ 4000.00	₹ 720.00 (18%)	₹ 4720.00
Total			2			₹ 990.00	₹ 6490.00

Tax type	Taxable amount	Rate	Tax amount	Amounts
SGST	₹ 5500.00	9%	₹ 495.00	Sub Total
CGST	₹ 5500.00	9%	₹ 495.00	Total
				₹ 6490.00

Estimate Amount In Words
Six Thousand Four Hundred Ninety Rupees only

Bank Details	Terms and Conditions	
 <p>Name : AXIS BANK, KUTHUPARAMBA Account No. : 920020073036484 IFSC code : UTIB0003378 Account holder's name : MEDIDENT</p> <p> SCAN TO PAY</p>	<p>Terms and Conditions</p> <ol style="list-style-type: none"> 1. Customer will be billed after indicating acceptance of this quote 2. Transportation & Installation Included, Loading takes care of customer 3. Payment 50% Advance 4. Goods delivery, only receipt of 100% payment 5. Shipment within 21 working days of order confirmation <p>Thanks for doing business with us!</p>	<p>For : MEDIDENT</p>  <p>Authorized Signatory</p>