



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 223444

547

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>SPHL</u></p> <p>Address <u>Thiruvananthapuram</u> <u>Kerala</u></p> <p>Ph : <u>7510868286</u></p>	<p>Call Registration Date : <u>16/4/24</u></p> <p>Caller ID : <u>106561</u></p> <p>Date of Visit : <u>17/4/24</u></p> <p>Asset No. : <u>0111677</u></p> <p>EQPT Name : <u>Microscope</u></p> <p>Manufacture <u>Labomed</u> Model : <u>Vision 2000</u></p> <p>S. No. <u>NA</u> Dept. <u>Medica</u></p>
---	---

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not Working

Action Taken : Checked and found objective, eye piece set, Prism set, Transforms and black box defective. Need to replace these spars for further checking and working condition of the equipment

Completed Date : 17/4/24 Time : 1:00 PM Spare Required


Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Soojya Sudheep</u>	<u>17/4/24</u>	<u>12:15 PM</u>	<u>1:00 PM</u>

Customer Remark Completed Pending

<p>Service Engineer Name : <u>Soojya Sudheep</u></p> <p>Signature : <u>Soojya</u></p> <p>Date : <u>17/4/24</u></p> <p>Contact Number : <u>7907406244</u></p>	<p>Customer Name : <u>AI SWARYA-S</u></p> <p>Signature : <u>Aiswarya</u></p> <p>Date : <u>17/4/24</u></p> <p>Contact Number : <u>7510868286</u></p> <p>Designation : <u>SO</u></p> <p>Hospital Seal :</p>
--	---




**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

547

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	STATE PUBLIC HEALTH LABORATORY
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : Laboratory Microscope Make : Labomed Model : Vision 2000 Sl:081221071
4	Equipment ID/ Barcode	106561/0111677
5	Date of purchase/ Year of manufacture/Installation Date	09/03/2012
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	16/04/2024
9	Action taken	Checked and found objectives, eye piece set, Prism set, Transformer and blackbox defective. Need to replace these spares for further checking and working condition of the equipment. Enquired spares from OEM.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	Objective set - 3142/- Eye piece set - 2832/- Prism set - 3776/- Transformer and black box - 3776/- Total - 18526/-
13	Asset Value	29962/-

14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	61.8%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT AND VENDOR QUOTATION ATTACHED
16	Reasons for recommending the equipment as BER	Checked and found objectives, eye piece set, Prism set, Transformer and blackbox defective. The equipment was installed on 09/03/2012 and covered upto 12+ years. The repairing cost is 61.8%. As per the tender clause 5.3.14.1 both criteria met. So recommending the unit for condemnation
17	Name & Signature of CYRIX Authority	KASYEP PV 


*Not mandatory #Based on the period of life and value as per the BER guidelines

* **Attach Photograph**

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Objective, eye piece set, Prism set, Transformer etc defective. So as per tender conditions Microscope can be RBER.


MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


18-07-2024

Signature of JC BM (NHM)

Date:




17.7.24

Signature of
Superintendent/Medical Officer (i/c)
DIRECTOR
State Public Health &
Clinical Laboratory
Thiruvananthapuram

STOCK REGISTER OF

Name of Article

9

Date	No. and date of voucher or Invoice	From whom received or to whom Issued	Receipt	Issued	Balance after each transaction	Initials of Receiver	Remarks
		Brought forward					Kriploskan Cde
9.3.12.		Labomed Vrsion 2000		081221071			69017159 0111677
		Binocular					
		Labomed ATC 2000		090221918			69017145 0111676
		Binocular					
		(Purchased on 4-2-2010)					
		₹ 29962/-	2		18		
30.3.2016		Cartron		10071681			69017141
		CM 2001					
		Binocular					0111682
		3yr Warranty	1		19		
		₹ 29962/-					
3.4.17	1)	Labomed Vrsion 2000		161295162			70010175 0111706
		Binocular					
	2)	Labomed Vrsion 2000		161295118			70010176 0111671
		Binocular					
	3)	Labomed Vrsion 2000		161295153			70010177 0111672
		Binocular	3		22		
		(3yr Warranty)					
		Carried over					

UNIVERSAL AGENCIES



ISO 9001 : 2008

Vision for excellence

PALAKKAL ANGADI, THRISSUR - 680 001, KERALA
 0487 - 2427038, 2427031, Mob : 93870 21122, 8943341569
 E-mail : uniage@gmail.com, uniage.sales@gmail.com



Ref :

CYRIX HEALTHCARE PVT LTD ERANAKULAM KERALA		Our Quotation No: UA-Q 1023/2024	Dt 10.07.2024	
		Ref: Quotation for the LABOMED VISION 2000 MICROSCOPE service charge		
Sir,				
We hereby quote our lowest possible rates for the following items for your kind consideration and early favorable orders.				
No.	Description	Unit Rate (Rs.)	GST %	Unit Rate Inclusive of GST (Rs.)
1.	OBJECTIVE SET	6900.00	18	8,142.00
2.	EYE PIECE SET	2400.00	18	2,832.00
3.	PRISM SET	3200.00	18	3,776.00
4.	TRANSFORMER AND BLACK BOX	3200.00	18	3,776.00
TOTAL COST Rs - 18,526/-				
Terms and Conditions				
Free delivery and demonstration at your stores.				
Supplies will be made within 8-12 weeks from date of issue of purchase order.				
Rates are inclusive of GST (If there is any revision in GST, Revised GST rate will be applicable.)				
Rates will remain firm for acceptance till 31.12.2024				
GSTIN: 32AAAFU5662Q1ZQ				
Payment should be made within 30 days time on receipt & acceptance of the goods / items.				
Account Name: M/s. Universal Agencies				
Our bank details:				
STATE BANK OF INDIA				
PARAMAKKAVU TEMPLE BRANCH THRISSUR - 70166				
A/c No.00000067213868605				
IFSCODE SBIN0070166				
CONTACT DETAILS:		 Yours Faithfully, For Universal Agencies,  Managing Partner.		
Order status/Sales/Service:				
8943356134				



memmert
GERMANY



micro Tec
GERMANY

EUTECH

LABOMED



ALUF
STAR
GERMANY



BECKMAN
Coulter

Microscopes, Balances, Spectrophotometers, Microtomes & Tissue Processors



REGARDING QUOTATION - 106561 - 0111677 - STATE PUBLIC HEALTH LABORATORY - LABORATORY MICROSCOPE

messages

bijith A <abijithacyrix@gmail.com>
o: arshadtk99@gmail.com <arshadtk99@gmail.com>
c: Kasyep P V <kasyepcyrix@gmail.com>

Wed, 10 Jul 2024 at 6:4

Dear Sir,
Kindly arrange a quotation for the below spares.
Make - Labomed
Model- Vision 2000

- 1)Objective set
- 2)Eye piece set
- 3)Prism set
- 4)Transformer and black box

—
Thanks & Regards,

Abijith A
Divisional Manager | South Zone
Mobile: 7593847141
abijithacyrix@gmail.com
www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd
| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India
KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

bijith A <abijithacyrix@gmail.com>
o: Kasyep P V <kasyepcyrix@gmail.com>

Wed, 10 Jul 2024 at 10:5

— Forwarded message —

From: **mohd arshad tk arshad** <arshadtk99@gmail.com>
Date: Wed, Jul 10, 2024 at 6:49 PM
Subject: Re: REGARDING QUOTATION - 106561 - 0111677 - STATE PUBLIC HEALTH LABORATORY - LABORATORY MICROSCOPE
To: **Abijith A** <abijithacyrix@gmail.com>

On 10 Jul 2024, at 6:47 PM, **Abijith A** <abijithacyrix@gmail.com> wrote:

[Quoted text hidden]

