

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	Ernakulam
2	Name of Hospital	TALUK HOSPITAL ANGAMALY
3	Name of Equipment with Make, Model and Serial Number	Equipment: OXYGEN CONCENTRATOR Make: LONGFIAN Model: JAY-10 Sr.no : MZJIOD5648
4	Equipment ID & Barcode	107455 & 0722833
5	Date of purchase / Year of manufacture /Installation Date	25-05-2021
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CAMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	22-04-24 (Toll Free)
9	Action taken	Checked the machine and found that sieve bed, compressor unit, control board, and display are faulty. Enquired the spare.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	Sieve bed – 1 No.- 9,500 Compressor unit- 1No –10,000 Control board- 1 No – Rs. 7,000 Display- 1 No.- Rs. 4,000  Total- Rs.35,420



13	Asset Value	Rs.26,700
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	132.65%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	. Checked the machine and found that sieve bed, compressor unit, control board, and display are faulty. Equipment installed on 25-05-2021 and it is a donated equipment. As the repairing cost is 132.65% higher than asset value. So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	JIBIN THOMAS GEORGE

\*Not mandatory

#Based on the period of life and value as per the BER guidelines

\* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

*This equipment is received at the time of Covid pandemic time as donation. Manufacturing date mentioned on the unit is 25/5/21. As per the report of repair major parts were damaged and the cost submitted is beyond the Asset value. BER report submitted after 30 days.*

*Devalthy*  
24/06/24  
Signature of JC BM (NHM)

CONSULTANT  
Bio-Medical Equipment  
National Health Mission (NHM)  
Aryankulam Ernakulam

Superintendent / Medical Officer (i/c)



**Superintendent**  
Taluk Hospital, Angamaly



Signature of seal

Date

28/6/2024





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :**  
200670

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility : THQH</p> <p>Address : THQH Angamali Ernakulam, Kerala</p> <p>Ph : 9746654704</p>	<p>Call Registration Date : 22/4/2024</p> <p>Caller ID : 107455</p> <p>Date of Visit : 25/4/2024</p> <p>Asset No. : 0722533</p> <p>EQPT Name : Oxygen Concentrator</p> <p>Manufacture : LOWU FAN Model : Jay-10</p> <p>S. No. M2J10D5641 Dept. Paediatric</p>
---	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust. Training

**Problem Identified :**  
Not working

**Action Taken :** checked and O2 concentrator compressor not working identified. Compressor piston broken bearings fully damaged also cooler fan and solenoid system not working. Machine in fully damaged condition.

Completed  Date : 25/4/24 Time : 2:00pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1. Compressor	1		
2. Solenoid sy ty, Flowmeter	1		
3. cooler fan; silicon tubes 8m	1		

<b>Cyrix Engineer</b>	<b>Date</b>	<b>Start Time</b>	<b>End Time</b>
Yadhukulakshra C S	25/4/24	1:00pm	2:00pm

**Customer Remark** Completed  Pending

<p><b>Service Engineer Name :</b> Yadhukulakshra C S</p> <p><b>Signature :</b> </p> <p><b>Date :</b> 25/4/24</p> <p><b>Contact Number :</b> 8086147236</p>	<p><b>Customer Name :</b> Taluk Hospital Angamaly</p> <p><b>Signature :</b> </p> <p><b>Date :</b> 25/4/24</p> <p><b>Contact Number :</b> 9746654704</p> <p><b>Designation :</b> Secondary Palliative Nurse</p> <p><b>Hospital Seal :</b> </p>
--	---



17/9/22 Babu. T.K. 60/m

(R) Thachukudy  
Nayathode


24/9/22 Received from store [Don.  
[Incident entered in camera

30/9/22 Varghese 68/m 109

Mr. Mathias H,  
Kavayapattom.

76;





**Medical Oxygen Concentrator**

MODEL: JAY-10  
INPUT: 230V, 50Hz, 880 VA  
NET WEIGHT: 27kg

IP21

CE 0187

SN: MZ10D56198  
Lot No: 21010027  
MDF 20110525  
Exp: 5 YEARS

**KCP**

INDONESIA: KCP  
KOROR: KCP

Call FREE: 1800-2700-890

**CAUTION**

- 1. Please read user's manual before operation.
- 2. Never perform any maintenance or repair with a machine.
- 3. Do not use the machine for other uses than intended.
- 4. Do not use the machine for other purposes than intended.
- 5. Do not use the machine for other purposes than intended.

SALES AND SERVICE CENTER  
KCP

CALL FREE NO: 1800-2700-890  
TEL: 0898-072801



**Spare Estimation Request- Longfian- Jay 10- Oxygen Concentrator- Taluk Hospital- Angamaly- 0722833-107455**

6 messages

**JIBIN THOMAS GEORGE** <jibintcyrix@gmail.com>  
To: technocare9@gmail.com <technocare9@gmail.com>  
Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>, Abijith A <abijithacyrix@gmail.com>

Thu, May 23, 2024 at 7:55 AM

Dear Sir,

Gentle Reminder....

The following spares are required for repairing the oxygen concentrator Longfian Jay 10 at Taluk Hospital Angamaly having barcode No 0722833 and complaint ID 107455.

Sieve bed- 1 No.  
Compressor unit-1No  
Display-1No  
Control Board-1No

Please send me the estimate for the same at the earliest. Feel free to contact me for any clarification.

Thanks & Regards,

Jibin Thomas George  
DI-KLBEMMP ERNAKULAM  
MOB:9047958051

**JIBIN THOMAS GEORGE** <jibintcyrix@gmail.com>  
To: technocare9@gmail.com  
Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>, Abijith A <abijithacyrix@gmail.com>

Mon, May 27, 2024 at 7:42 AM

Gentle Reminder...  
[Quoted text hidden]

**JIBIN THOMAS GEORGE** <jibintcyrix@gmail.com>  
To: technocare9@gmail.com  
Cc: Afsal Yunuze <zm1.klbemp@cyrix.in>, Abijith A <abijithacyrix@gmail.com>, es.kmscl@kerala.gov.in

Wed, May 29, 2024 at 12:49 AM

ear Sir,

Gentle Reminder....

The following spares are required for repairing the oxygen concentrator Longfian Jay 10 at Taluk Hospital Angamaly having barcode No 0722833 and complaint ID 107455.

Sieve bed- 1 No.  
Compressor unit-1No  
Display-1No  
Control Board-1No

Please send me the estimate for the same at the earliest. Feel free to contact me for any clarification.

Thanks & Regards,

Jibin Thomas George  
DI-KLBEMMP ERNAKULAM



MOB:9047958051

Wed, Jun 12, 2024 at 9:01 AM

[Quoted text hidden]

**JIBIN THOMAS GEORGE** <jibintcyrix@gmail.com>  
To: technocare9@gmail.com  
Cc: zm1.klbemp@cyrix.in, abijithacyrix@gmail.com, es.kmscl@kerala.gov.in

Dear Sir....

Gentle Reminder....

[Quoted text hidden]

Wed, Jun 12, 2024 at 11:31 AM

**POOJA DHADUK** <technocare9@gmail.com>  
To: JIBIN THOMAS GEORGE <jibintcyrix@gmail.com>

Dear sir

please check below mentioned rate for spares

sieve bed pair 9500 +18%+trans.  
Compressor 10000 +18%+trans.  
Displ 4000 +18%+trans.  
Control board 7000 +18%+trans.  
[Quoted text hidden]

**BEST REGARDS ,**

**MS.POOJA DHADUK**  
**Technocare medisystems**  
C-1310/11,new Bombay market,  
Umerwada,Surat-395010

Ph no :0261-2332042 /6569308  
Mo : 9537417749  
web :www.technocarem.com

Wed, Jun 12, 2024 at 11:01 PM

**POOJA DHADUK** <technocare9@gmail.com>  
To: JIBIN THOMAS GEORGE <jibintcyrix@gmail.com>  
Cc: zm1.klbemp@cyrix.in, abijithacyrix@gmail.com, es.kmscl@kerala.gov.in

DEAR SIR ,

**PLEASE CHECK BELOW ATTACHED QUOTATION**  
[Quoted text hidden]

GSTIN : 24ACJPV8719Q1ZK

Original Copy

Sales Quotation  
**TECHNOCARE MEDISYSTEMS**

C-1310/11, NEW BOMBAY MARKET ,, UMERWADA  
SURAT, MO- 95374 17749  
email : [technocarem@gmail.com](mailto:technocarem@gmail.com)

Quotation No. : 71

Date : 12-06-2024 ( 11:00 PM )

**Quotation to :**

Jibin Thomas George  
DI-KLBEMMP ERNAKULAM  
Pin - 682011

**Shipped to :**

Jibin Thomas George  
DI-KLBEMMP ERNAKULAM  
Pin - 682011

Party Mobile No : 9047958051

Party Mobile No : 9047958051

State : Kerala (32)

State : Kerala (32)

GSTIN / UIN :

GSTIN / UIN :

S.N.	Description of Goods	HSN/SAC Code	Qty.	Unit	Price	IGST Rate	IGST Amount	Amount ( )
1.	OXYGEN BED SEIVE PAIR	90189099	1.00	Pcs.	9,500.00	12.00 %	1,140.00	10,640.00
2.	ACCESORIES Compressure	48119099	1.00	Pcs.	10,000.00	18.00 %	1,800.00	11,800.00
3.	NEW DISPLAY	48119099	1.00	Pcs.	4,000.00	18.00 %	720.00	4,720.00
4.	NEW BOARD Control Board	90189019	1.00	Pcs.	7,000.00	18.00 %	1,260.00	8,260.00

**Grand Total**                      **4.00 Pcs.**

**35,420.00**

HSN/SAC	Tax Rate	Taxable Amt.	IGST Amt.	Total Tax
48119099	18%	14,000.00	2,520.00	2,520.00
90189019	18%	7,000.00	1,260.00	1,260.00
90189099	12%	9,500.00	1,140.00	1,140.00
<b>Total</b>		<b>30,500.00</b>	<b>4,920.00</b>	<b>4,920.00</b>

**Rupees Thirty Five Thousand Four Hundred Twenty Only**

**Bank Details :** ICICI BANK , NAME - TECHNOCARE MEDISYSTEMS  
AC NO.- 058405000623 , IFSC CODE - ICIC0000584

**for TECHNOCARE MEDISYSTEMS**

**Authorised Signatory**