

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: GH THALASSERY	Name of District: KANNUR
Name of Equipment: In ticket master Equipment name is Ultrasound therapy unit, but actual equipment is IFT MACHINE	Equipment ID & Barcode: #131725- 1311317
Make: BMS	Model: DIGI STIM-PRO
Serial Number: 7990	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 18.12.2015	Present status of the equipment: Fully damaged
Date of breakdown: 27.08.2024	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Date of registration of complaint through email/Toll free): Toll free	

Action taken: Checked the Equipment Found Equipment is not switching ON. Found SMPS Board, Microcontroller Board, IFT Board, Display Control Board Defective. Enquired OEM for quotation.


Recommendations for repair (required service details): No Recommendations

Cost of spares (specify parts and cost): SMPS Board- 3,422/-
Microcontroller Board-4,248/-
IFT Board-5,664/-
Display Control Board-2,832/-
Total-16,166/-


#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 76.25%	Asset Value: 21,200/-
---	-----------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service Report & Quotation attached.

Reasons for recommending the equipment as BER: Checked the Equipment Found Equipment is not switching ON. Found SMPS Board, Microcontroller Board, IFT Board, Display Control Board Defective. Enquired OEM for quotation. Machine Installed on 18.12.2015 aged up to 11 years 9 months. Repairing Cost Is 76.25% Both criteria for RBER met. As per Tender Clause 5.3.14.1 we recommend the equipment for condemnation.

Avinash 
11/09/24
Name & Signature of CYRIX Authority with date


Remarks of Junior Consultant (Biomedical) NHM: checked & identified that machine not working due to SMPS, Microcontroller Board, IFT Board, Display board are defective. Aged up to 11 years & above. As per OEM'S quotation, machine service cost economically not feasible. So recommended for BER as per BER protocol

Recommended for BER (Yes/ No):	 11/09/24 CONSULTANT BIOMEDICAL NATIONAL HEALTH MISSION Signature of JC BM (NHM)
Date: 11/09/24.	

For Lasitha. O. (SNO)
Sudhi. P. (SNO)

Date: 11/11/24. Seal




 Signature of Superintendent / Medical Officer (i/c)
SUPERINTENDENT
Govt General Hospital
Thalassery

*Not mandatory #Based on the period of life and value as per the BER guidelines. Attach Photograph



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1014554

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <u>GH Thalassery</u>	Call Registration Date: <u>27/08/24</u>
Address: <u>Thalassery</u>	Caller ID: <u>131725</u>
<u>Kannur</u>	Date of Visit: <u>28/08/24</u>
Ph: <u>9995821486</u>	Asset No: <u>13.11317</u>
	EQPT Name: <u>IFT Machine</u>
	Manufacture: <u>BMS</u> Model: <u>DigistimPro</u>
	S. No: <u>7990</u> Dept: <u>Physiotherapy</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Machine is not switching ON

Action Taken : Checked the Equipment found Equipment is not switching ON. Found SMPS board, microcontroller board, IFT board, Display control board Defective. Enquired OEM for quotation.

Completed Date: 28/08/24 Time: 3:00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Rasikraj</u>	<u>28/08/24</u>	<u>2:00 pm</u>	<u>3:00 pm</u>

Customer Remark Completed Pending

Service Engineer Name : Rasikraj Superintendent For: Lasitha-O (SNO)
 Signature: Rasikraj Customer Name: General Hospital Thalassery
 Date: 28/8/24 Signature: [Signature]
 Contact Number: 9995521436 Date: 28/8/24
 Contact Number: 9495726609 Designation: SNO
 Hospital Seal





Phone : 91-44-24516904, 24516539
Cell : 91-0-90031 04470, 98402 12531

BHARAT MEDICAL SYSTEMS

(Manufacturer & Exporter of Physio Therapy and Medical Equipments)

M-1074 (New # 15), 7th Main Road, Thiruvalluvar Nagar, Thiruvananthapuram, Chennai - 600 041, India.

Date: 06.09.2024

Place: Chennai – 600 041.

Quotation

To:

CYRIX HEALTHCARE (P) LTD
Kannur bemp kerala

Dear Sir / Madam,

Please find below our best prices for Electro Therapy Equipment for your kind reference:

S. No	Description	Qty	Product Images	Accessories	UNIT PRICE (INR)
01.	SMPS Board	1No.			3,422.00
02.	microcontroller board	1No.			4,248.00
03.	IFT Board	1No.			5,664.00
04.	Display control board	1No.			2,832.00

TERMS & CONDITIONS:

1. Mode of Payment : 100% Advance with Firm order
2. Dispatch : Within 15 days
3. Jurisdiction : Chennai, INDIA Only
4. Warranty : 12 Months from the date of installation for the equipment's and accessories against any manufacturing defects only.
Warranty is not applicable for any physical damage ,and problems due to fell down, Rate bite, tearing , natural disaster like flood, fire etc.
5. Goods Once sold will not be taken back.
6. Credit Card Facility available
7. Any Technical Clarification contact : 9444052532
8. Packing and forwarding extra charges.
9. GST 12% Included Sr.no:1.
10. GST 18% Included Sr.no: 2,3,4,5.



AN ISO 13485 : 2016, 9001-2015 and



CERTIFIED COMPANY

OUR RANGE OF PRODUCTS:

GST No. 33AEXPJ4035M122
I.E. Code No. 0404013074
TAN No. CHEJ05558A
PAN No. AEXPJ4035M

Interferential Therapy
Ultra Sound Therapy
Electronic Traction
Shortwave Diathermy 300W & 500W
CPM Knee, Elbow, Shoulder and Wrist
TENS-Dual Channel (Pocket Size & Table Model)

Scanning LASER Therapy
LASER Therapy
Vacuum Therapy
Unweighing System
Muscle Stimulator
All Exercise Equipments

Hydro Collator
Wax bath
Combo 5 in 1
Combo 4 in 1
Combo 3 in 1
Combo 2 in 1

Visit us : www.physiotherapyequipment.in

E-mail : bmsmadras@gmail.com



Re: Quotation required for IFT Machine #1311317, @GH THALASSERY.

2 messages

avinash t <avinashtcyrix@gmail.com>

Fri, 6 Sept, 2024 at 2:54 pm

To: Bharat Medical Systems <bmsmadras@gmail.com>

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Dear sir,
#reminder,

On Wed, 4 Sept, 2024, 3:01 pm avinash t, <avinashtcyrix@gmail.com> wrote:

Dear sir,
Please provide the below mentioned spare quotation for ULTRASOUND THERAPY UNIT (make-BMS,Model-Digistim pro, sl no-7990)
1.SMPS Board
2.microcontoller board
3.IFT Board
4.Display control board

Thanks & Regards

Avinash T
District charge, kannur
bemp kerala
mobile: +91 6364064948
E mail: avinashtcyrix@gmail.com

CYRIX HEALTHCARE (P) LTD
30/641B | Petta | Punithura | Kochi | Kerala- 682038
www.cyrixhealthcare.com

Bharat Medical Systems <bmsmadras@gmail.com>

Fri, 6 Sept, 2024 at 3:16 pm

To: avinashtcyrix@gmail.com



Best regards,

Bharat Medical Systems,
M-1074, 7TH Main Road
Thiruvalluvar Nagar,
Thiruvanniyur,
Chennai – 41, India.

Office: 044-24516904, 24516539

Mobile: + 91 9003104470, +91 9840212531

Website: www.physiotherapyequipment.in

E-Mail: bmsmadras@gmail.com,

Skype Id: balasubramanian133

[Quoted text hidden]

