

#  
PROFORMA  
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR  
(RBER)

BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

|   |   |
|---|---|
| Name of Hospital: GH THALASSERY   | Name of District: KANNUR                                      |
| Name of Equipment: DIALYSIS MACHINE   | Equipment ID & Barcode: #110617- 1311107                      |
| Make: NIPRO   | Model: SURDIAL 55 PLUS  |
| Serial Number: 17CX7356   | Warranty details: NO WARRANTY                                 |
| Date of purchase/ Year of manufacture/ Installation Date: Ticket master installation date 04-12-2017, As per Stock book, Installation Date is 04.02.2017. (Stock Book copy attached for reference)  | Present status of the equipment: Fully Damaged                |
| Date of breakdown: 11.05.2024<br>Date of registration of complaint through email/ Toll free): Toll Free   | *AMC/CAMC Period agreed at the time of purchase: NO AMC/ CAMC |
| Action taken: Checked the Equipment found Equipment is not switching ON, found I/O Board, Power Board, air elimination tank assy, Rinse Port, NIBP Module Defective. Enquired OEM for Quotation.  |   |
| Recommendations for repair (required service details): No recommendations   |   |
| Cost of spares (specify parts and cost): I/O Board-1,44,256/-<br>Power Board-84,444.64/-<br>Air Elimination Tank assy- 47,488/- Rinse Port-9,842.56/-<br>NIBP Module-1,00800/-<br>Total-3,86,831.2/-  |   |
| Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 85.58%  | Asset Value: 452000/-   |
| Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service report & Quotation Attached.  |   |
| Reasons for recommending the equipment as BER: Checked the Equipment found Equipment is not switching ON, found I/O Board, Power Board, air elimination tank assy, Rinse Port, NIBP Module Defective. Enquired OEM for Quotation. In software Equipment installed on 04.12.2017. As per the stock book installation date is 04.02.2017 aged up to 7 years 7 months. Repairing cost is 85.58% Both criteria RBER met. As per tender clause 5.3.14.1 we recommend the equipment for condemnation. |   |
| Avinash T. [Signature]<br>11/09/24<br>Name & Signature of CYRIX Authority with date   |   |
| Remarks of Junior Consultant (Biomedical) NHM: checked & identified that machine not switching ON due to PB, I/O Board, air elimination tank assy, Rinse Port, NIBP module are defective. Aged up to 7 years & above. As per OEM's quotation  |   |
| Recommended for BER (Yes/ No):  | The service cost economically not feasible.                   |
| Date: 11/09/24  | [Signature]<br>11/09/2024<br>Signature of JC BM (NHM)         |
| Date: 24/9/24<br>(i/c)  |   |
| [Signature]<br>24/9/24<br>Seal<br>RESIDENT MEDICINE SUPERINTENDENT / Medical Officer<br>GENERAL HOSPITAL<br>THALASSERY  |   |

\*Not mandatory Based on the period of life and value as per BEMP guidelines\* Attach Photograph





# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

**CYRIX<sup>®</sup>**  
HEALTH CARE PVT LTD

No : 1014552

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

|   |   |
|---|---|
| Health Facility: <u>G.H. Thalassery</u> | Call Registration Date: <u>11-5-24</u>                  |
| Address: <u>Thalassery</u>              | Caller ID: <u>110617</u>                                |
| <u>Kannur</u>                           | Date of Visit: <u>13-5-24</u>                           |
| Ph: <u>9544595382</u>                   | Asset No: <u>1311107</u>                                |
|   | EQPT Name: <u>Dialysis Machine</u>                      |
|   | Manufacture: <u>Nipro</u> Model: <u>Sundial SS plus</u> |
|   | S. No. <u>17CX7356</u> Dept. <u>Dialysis</u>            |

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : Machine not switching ON.

Action Taken : checked the machine. Found machine not switching ON. Found I/O Board, Power board, Air elimination tank assy, Rinse port, Nibp module defective. Enquired OEM for Quotation.

Completed  Date : 13-5-24 Time : 2:00pm Spare Required


Spare Replaced  Requested

| Description | Qty. | Part Number | PR Number |
|-------------|------|-------------|-----------|
| 1.          |      |             |           |
| 2.          |      |             |           |
| 3.          |      |             |           |

| Cyrix Engineer  | Date           | Start Time     | End Time      |
|-----------------|----------------|----------------|---------------|
| <u>Rasilraj</u> | <u>13-5-24</u> | <u>11:00AM</u> | <u>2:00pm</u> |

Customer Remark Completed  Pending

Service Engineer Name : Rasilraj  
Signature: Rasilraj  
Date: 13-5-24  
Contact Number: 9995521436



Customer Name: Muhammed Nilul M  
Signature: [Signature]  
Date: 13/5/24  
Contact Number: 9544436433  
Designation: Dialysis Technician  
Hospital Seal

RESIDENT MEDICAL OFFICER  
GENERAL HOSPITAL  
THALASSERY



## Quotation

Quotation No.SW [ 481 ] 09-2024

Date:10-09-2024

To

Cyrix Healthcare Pvt Ltd,  
Poonithura , Ernakulam

Dear Sir,

Thank you for considering us and the opportunity to submit a quotation. We are delighted to present our quote as follows.

| SI no | Name of the Items         | Unit Price    | Qty | Amount        | GST | GST Amount  | Total         |
|-------|---------------------------|---------------|-----|---------------|-----|-------------|---------------|
| 1     | I / O BOARD               | ₹ 1,28,800.00 | 1   | ₹ 1,28,800.00 | 12% | ₹ 15,456.00 | ₹ 1,44,256.00 |
| 2     | POWER BOARD               | ₹ 75,397.00   | 1   | ₹ 75,397.00   | 12% | ₹ 9,047.64  | ₹ 84,444.64   |
| 3     | AIR ELIMINATION TANK ASSY | ₹ 42,400.00   | 1   | ₹ 42,400.00   | 12% | ₹ 5,088.00  | ₹ 47,488.00   |
| 4     | RINSE PORT                | ₹ 8,788.00    | 1   | ₹ 8,788.00    | 12% | ₹ 1,054.56  | ₹ 9,842.56    |
| 5     | NIBP MODULE               | ₹ 90,000.00   | 1   | ₹ 90,000.00   | 12% | ₹ 10,800.00 | ₹ 1,00,800.00 |
|       |                           |               |     |               |     |             | ₹ 3,86,831.2  |

### Terms & Conditions

- 100% advance payment along with your Purchase Order.
- Cheque in favor of: SmartWay Healthcare

A/c No: 0224073000005622

IFSC: SIBL0000224

Bank: South Indian Bank

Branch: Kalamassery

Thanking You

SmartWay Healthcare

GSTIN: 32AERFS9757B1ZM

Shop No: 21/497 | Ground Floor, Kaipadamugal  
Thrikkakara North | Ernakulam – 682 021

+91 8921172300 | smartwayhealthcarekm@gmail.com



---

## Dialysis Machine spare quotation required 1311107@ GH THALASSERY, KANNUR

3 messages

---

**avinash t** <avinashtcyrix@gmail.com>

Thu, 5 Sept, 2024 at 3:53 pm

To: Smartway EKM <smartwayhealthcareekm@gmail.com>

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, Blesson Jose <zm2.klbemp@cyrix.in>

Dear Sir,  
Please provide a quotation of the below mentioned parts for DIALYSIS MACHINE(make- Nipro, model-Surdial 55 plus, )

1. i/o Board
2. Power Board
3. Air elimination tank
4. BPM module
5. Rins port

Thanks & Regards

Avinash T  
District charge, kannur  
bemp kerala  
mobile: +91 6364064948  
E mail: [avinashtcyrix@gmail.com](mailto:avinashtcyrix@gmail.com)

**CYRIX HEALTHCARE (P) LTD**  
30/641B | Petta | Punithura | Kochi | Kerala- 682038  
[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)

---

**Smartway EKM** <smartwayhealthcareekm@gmail.com>

Tue, 10 Sept, 2024 at 4:42 pm

To: avinash t <avinashtcyrix@gmail.com>

Dear Sir ,

Please find the Attached Quotation for your reference

Regards

Ayas  
Smartway Healthcare  
[Quoted text hidden]

---

**Smartway EKM** <smartwayhealthcareekm@gmail.com>

Mon, 16 Sept, 2024 at 12:50 pm

To: avinash t <avinashtcyrix@gmail.com>

Dear Avinash

Please find the attached quotation for your reference

Regards

Ayas  
Smartway healthcare  
[Quoted text hidden]

# GENERAL HOSPITAL THALASSERY


## STOCK BOOK OF MEDICINES AND EQUIPMENTS

Name of Item HAEMDIALYSIS MACHINE

| Date  | Invoice No. & Date | From whom received or to whom issued | Receipt | Issue | Balance | Physical Stock Balance | Expiry Date | Cost/Unit    | Company | Remarks                                   |
|---|--------------------|--------------------------------------|---------|-------|---------|------------------------|-------------|--------------|---------|---|
| 4/2/18  | LP 439<br>17-8     | Recd from store                      | 10 ✓    |       | 10      | 10                     | 3/2/22      |              |         | 12/2/11; 17/11/22<br>Hm                   |
| 3/4/18  |                    | Issued for use                       |         | 10    | 0       | 10                     |             | 47,912.00/10 |         |   |
| 30/6/18   | LP 438<br>18-19    | Received from store                  | 2       |       | 2       | 12                     | 30/4/29     |              |         | Hm  |
|   |                    | Issued for use                       |         | 2     | 0       | 12                     |             | 10,12,480.50 |         |   |
| 15/7/24   | 15/149<br>24-25    | Recd from store<br>(Sunder 55)       | 4       |       | 16      | 16                     |             |              |         | 672 post<br>Micro<br>TOTAL<br>26,88,000/- |
| <p>(Detected Items by Lion Club only)<br/>                     executed upto 11.06.2025<br/>                     Contact Person Aarti - 924940973<br/>                     State: 240 x 33297/292/293/294<br/>                     Mfgt Supplier - Nipso MEDICAL (INDIA)<br/>                     Pvt-Ltd / Imidation &amp; emergency direct Supplier 1/e</p> |                    |                                      |         |       |         |                        |             |              |         |   |



Muhammed Nihal . M  
 27/07/24  
 Dialysis technician



GOVT. GENERAL HOSPITAL  
THALASSERY - 670101  
Nikhil M  
Dialysis technician