






**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

PROFORMA

Sl. No	Particulars	Details
1	Name of District	ERNAKULAM
2	Name of Hospital	PHC GOTHURUTHU ERNAKULAM
3	Name of Equipment with Make, Model and Serial Number	Equipment : INFRARED LAMP Make : Na Model : Na SN : Na
4	Equipment ID & Barcode	122668 & 0743572
5	Date of purchase / Year of manufacture /Installation Date	09-06-2012
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	12-07-2024 (Toll free)
9	Action taken	Checked and found bulb, holder and wiring defective. Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	Na


13	Asset Value	Rs. 495/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Na
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found wires, lamp and holder defective. Machine installed on 09-06-2012 and covered up to 12+ years and it is a old machine. Repairing cost of the equipment is beyond economical. So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	Sebin. 

\*Not mandatory #Based on the period of life and value as per the BER guidelines

\* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

As per the Direction from HQ, BER Should be Submitted in a single sheet. spare equipment not attached. bulb, holder, wiring defective reported by n/s cyrix healthcare. MET year wise and maintain as per BER guideline.

  
9/8/24  
Signature of JCBM (NHM)  
National Health Mission (NHM)  
Arogyakeralam, Ernakulam



Date 22/8/24

  
മുഖ്യ ഓഫീസർ ഇൻ ചാർജ്  
Superior Health Centre / Medical Officer (i/c)  
Arogyakeralam Health Centre, Ernakulam



LR Lamb ✓

39

Date	Items	Rd	Balance
9/6/2012	Brought forward from old register p.no: 64	1	1 ✓ Soyy S.M.S

Small Namamab - 1





Mangji  
22/8/24

മെഡിക്കൽ ഓഫീസർ ഇൻ ചാർജ്  
കുടുംബാരോഗ്യകേന്ദ്രം തോട്ടുരുത്ത്

WEIGHING MACHINE

	Brought forward from old register P.no 33			MC
13/2/17	Rd from the store	1	1	} Jony S.M.S
8/10/18	Rd from the store (Digital Scale upto 150 kg) 4105/-	2	3	
25/2/19	Rd from Flood relief Sub Stock medical camp register page No: 23 (Digital) Sanson	1	4	} Jony S.M.S
10/8/19	Rd from the store	4	8	} Jony MC
11/2/22	Rd from Store (1608.04/-)	1	9	
25/5/23	Transfer in from LMT stock (Page: 264)	1	10	} Jony S.M.S
	keil waste room - (1)			
	feed op - 1			
	pp unit - 1			
	Duty room - 1			
	Small Namamab - 6			





*Mangli*  
22/10/2024

അന്വേഷണ ഗോതുമുതൽ കൂടുതൽ സാധനങ്ങൾ ഉൾപ്പെടെയുള്ള