



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

# CYRIX

HEALTHCARE PVT LTD

**No. :** 232229

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrrix.com | E-mail : bemp.kl@cyrrix.in

<p>Health Facility <u>General Hospital</u></p> <p>Address <u>Neyyattinkara</u> <u>Thiruvananthapuram</u></p> <p>Ph : <u>9745889343</u></p>	<p>Call Registration Date : <u>22-7-2024</u></p> <p>Caller ID : <u>124633</u></p> <p>Date of Visit : <u>24-7-2024</u></p> <p>Asset No. : <u>0113322</u></p> <p>EQPT Name : <u>Nebuliser</u></p> <p>Manufacture <u>EZ Life</u> Model : <u>101N</u></p> <p>S. No. <u>NA</u> Dept. <u>ward 2</u></p>
--	---

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Not working

**Action Taken :** Checked the unit and identified nebulizer motor defective and piston broken. Need to replace these spares for further checking and working condition of the equipment.

Completed  Date : 24-7-2024 Time : 11:15 a.m Spare Required


Spare Replaced  Requested

No.	Description	Qty.	Part Number	PR Number
1.				
2.	<u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>
3.				

Cyrux Engineer	Date	Start Time	End Time
<u>Ajith George</u>	<u>24-7-2024</u>	<u>10:45 a.m</u>	<u>11:15 a.m</u>

**Customer Remark**  Pending

<p><b>Service Engineer Name :</b> <u>Ajith George</u></p> <p><b>Signature :</b> <u>A.</u></p> <p><b>Date :</b> <u>24-7-2024</u></p> <p><b>Contact Number :</b> <u>7593847111</u></p>	<p><b>Customer Name :</b> <u>Neyyattinkara</u></p> <p><b>Signature :</b> <u>[Signature]</u></p> <p><b>Date :</b> <u>24/7/24</u></p> <p><b>Contact Number :</b> <u>9745889343</u></p> <p><b>Designation :</b> <u>Nursing officer</u></p> <p><b>Hospital Seal :</b></p>
--	---



**SUPERINTENDENT**  
GENERAL HOSPITAL  
NEYYATTINKARA 695121


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GENERAL HOSPITAL NEYYATINKARA
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : NEBULISER Manufacturer : EZ Life Model : 101N Sl: NA
4	Equipment ID/ Barcode	124633/ 0113322
5	Date of purchase/ Year of manufacture/Installation Date	29/09/2021
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	22/07/2024 (Toll free)
9	Action taken	Checked the unit and identified nebulizer motor defective and piston broken. Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not recommending for Repair
12	Cost of spares (specify parts and cost)	Not available



13	Asset Value	Rs . 1353/-
14	Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Na
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment RBER	Checked and found nebulizer motor defective and piston broken. Quotation not submitted since spares are not available in the market. So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AJITH GEORGE 

\*Not mandatory


#Based on the period of life and value as per the BER guidelines

\* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Nebulizer motor and piston defective.. So  
Can RBER as per tender conditions.

MANEESHA MOHAN M.R.  
Junior Consultant (Biomedical)  
National Health Mission  
Thiruvananthapuram

  
23-08-2024  
Signature of JC BM (NHM)

Date:




  
Signature of  
Superintendent/Medical Officer (i/c)

**SUPERINTENDENT**  
GENERAL HOSPITAL  
NEYYATTINKARA-695120

Nebulizer


83

Date	Description	Qty	Issue	Balance	Remarks
29/9/2021				42	
29/9/2021	Rd from store	1		542	
	Donation page no: MSR (069)				
8/4/2022	TR from casualty sub stock register page no: 315 (001 80 0201)	1		642	
29/10/2022	Condemned				
	no. 4.823/2022/ 29/10/2022 (2)		2	6	
2/9/2023	TR out unavailible page no. 92 (1) (2) 19/11/19 (2)		2	4	
4/12/2023	Rd from store LS/11601 23-24 dated 4/12/23	1		5	
10/1/2024	Received from store LS/01296/23-24 dt. 10.1.24	1		6	
17/2/24	Rd from store	1		7	
28/2/24	TR out to unavailible Register B/L page 92. 22/2/19 " 013337 19/11/19 (2)			5	
8/5/24	Transfer in from casualty General Hospital 27/1/19	1		6	
20/5/2024	TR. out to unavailible Register page no 92 BC (19.4.18) - 9113328			4	



SUPERINTENDENT  
GENERAL HOSPITAL  
NEYYATTINKARA-695121





  
SUPERINTENDENT  
GENERAL HOSPITAL  
NEYYATTINKARA-695121