



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 232232

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrlix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>General Hospital</u> Address <u>Neyyattinkara</u> <u>Thiruvananthapuram</u> Ph : <u>9745889343</u>	Call Registration Date : <u>27-7-2024</u> Caller ID : <u>125940</u> Date of Visit : <u>29-7-2024</u> Asset No. : <u>0112914</u> EQPT Name : <u>Nebulizer</u> Manufacture <u>Life care</u> Model : <u>CNB69009</u> S. No. <u>NA</u> Dept. <u>Geriatric ward</u>
--	--

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working.

Action Taken : Checked the unit and identified nebulizer motor defective and piston broken. Need to replace the spares for further checking and working condition of the equipment.


Completed Date : 29-7-2024 Time : 3.00 p.m Spare Required

Spare Replaced Requested

No.	Description	Qty.	Part Number	PR Number
1.				
2.	<u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>
3.				

Cyrix Engineer	Date	Start Time	End Time
<u>Ajith George</u>	<u>29-7-2024</u>	<u>2.30 p.m</u>	<u>3.00 p.m</u>

Customer Remark Pending




Service Engineer Name : <u>Ajith George</u> Signature : <u>A.</u> Date : <u>29-7-2024</u> Contact Number :	Customer Name : <u>General Hospital</u> Signature : <u>[Signature]</u> Date : <u>29-7-2024</u> Contact Number : <u>9745889343</u> Designation : <u>Nursing officer</u> Hospital Seal :
---	---

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GENERAL HOSPITAL NEYYATINKARA
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : Nebuliser Manufacturer : Lifecare Model : CNB69009 SL: NA
4	Equipment ID/ Barcode	125940/0112914
5	Date of purchase/ Year of manufacture/Installation Date	26/12/2018
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	27/07/2024(Toll Free)
9	Action taken	Checked the unit and identified nebulizer motor defective and piston broken. Need to replace the spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not recommending for Repair
12	Cost of spares (specify parts and cost)	Not available

13	Asset Value	Rs . 1353/-
14	Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Na
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment RBER	Checked the unit and identified nebulizer motor defective and piston broken. The unit was installed on 26/12/2018 and covered up to 5+ years. Since the spares are not available in the market, Recommended the unit for condemnation.
17	Name & Signature of CYRIX Authority	AJITH GEORGE 

*Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Motor and Piston broken . . . this can be RBER as per tender conditions.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


23-08-2024
Signature of JC BM (NHM)

Date:




Signature of
Superintendent/Medical Officer (i/c)

SUPERINTENDENT
GENERAL HOSPITAL
NEYYATTINKARA-695121

NEBULIZER


Date	Description	QTY	Rate	Total	Remarks
	B/F from page no. 120 of SR of T & G Book B on 1.3.06				which was received from store old Book (B) page no (25)
1.3.06		3			
27.3.06	Received from store (Medicare Compact M-90) (Acrosell therapy unit)	1	Rs. 1610/-	4	I-31.1.2006 I-1.12.03 Tr. in from Casualty pag. 38 (No. 109) I-4.6.04
13.9.10	Received from store	1		5	
13.10.11	Rel from store	1		6	for
28.4.12	Te out to page No. 80			2	CP
28/9/2013	Condemned as per order NO F 1940/2013 GII NTA dtd 28/9/2013			2	RV
11/12/2013	TR: in from Trauma Care unit	1		3	
17/6/2015	TR: in from OPD unit	1		4	
14/10/16	Received from store	2		6	① Phthalenim Old ward stock
14/10.16	① given for condemnation on (7/02/2018)			1	5
20/9/16	Received from store (TRUCHER compressor) nebulizer model no: N.501 MRP. Rs. 2999/-	1		6	supplied S/N 67
30/11/17	Condemned as per order A420/17 GII dated 30/11/17				
26/12/18	nebulizer				
3/8/2019	Received from store				
15/11/2019	From do				




SUPERINTENDENT
GENERAL HOSPITAL
NEYYATTINKARA-695121

3 8 5

BAR CODE - (8004-890615225)
0112914




SUPERINTENDENT
GENERAL HOSPITAL
NEYYATTINKARA-695124

