

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: PHC EDAPATTA	Name of District: MALAPPURAM
Name of Equipment: UPS	Equipment ID & Barcode: 137570 #1043263
Make: APC	Model: BACK- UPS 600
Serial Number: B21649010226	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket master installation date is N/A, As per stock book installation date is 22/05/2017	Present status of the equipment: Fully damaged
Date of breakdown:28/09/2024	*AMC/ CAMC Period agreed
Date of registration of complaint through email/Toll free): Toll free	at the time of purchase: No AMC/CAMC
Action taken: Checked the machine & found main board, battery & transformer damaged. Need to replace these spares for further checking.	

Recommendations for repair (required service details): No Recommendation

Cost of spares (specify parts and cost): 12-230 INVERTER PCB:1298/-
7Ah BATTERY:1536/-
TRANSFORMER: 649/-
TOTAL : 3483/-

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 77% Asset Value: 4500/-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service report attached & OEM Quotation attached .

Reasons for recommending the equipment as BER: Checked the machine and found 12-230 Inverter board, 7Ah battery and Transformer defective. In ticket master installation date is Nil, As per stock book installation date is 22/05/2017 Aged up to 7 years and 5 months, Repairing cost is 77% of asset value, both criteria for RBER met. As per the tender clause 5.3.14.1 We recommending the equipment for condemnation.

GOKUL KT 11/10/2024
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
Verified with the above damage of Transformer, Battery and Inverter board. Need to replace spares and is beyond economic repair.

Recommended for BER (Yes/ No):
Date: 11/10/2024

Signature of JC BM (NHM)
Jr. Consultant Biomedical Engineer
NHM Malappuram
To Mr. Anilhe...
Kareem

Date: 14/10/2024 Seal Signature of Superintendent / Medical Officer (I/c)
*Not mandatory #Based on the period of life and value as per the BER guidelines
കുടുംബാരോഗ്യകേന്ദ്രം
എഡപ്പാട്ട് - 679326



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1023409

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility..... <u>PHC</u></p> <p>Address..... <u>Eclappatta</u></p> <p>..... <u>Malappuram</u></p> <p>Ph:..... <u>7558868130</u></p>	<p>Call Registration Date: <u>28/09/24</u></p> <p>Caller ID :... <u>137570</u></p> <p>Date of Visit: <u>30/09/24</u></p> <p>Asset No :... <u>1043263</u></p> <p>EQPT Name :..... <u>UPS</u></p> <p>Manufacture... <u>APL</u>..... Model: <u>UPS 600</u></p> <p>S. No... <u>NA</u>..... Dept... <u>Laboratory</u></p>
---	--

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Machine not working

Action Taken : Checked the machine Found 12-230 Inverter battery 7 Ah battery and transformer defective. Replace these parts and spares for further checking

Completed Date 30/09/24 Time : 12:45 PM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Ahzel Rahman</u>	<u>30/09/24</u>	<u>10:00 AM</u>	<u>12:58 PM</u>

Customer Remark Completed Pending

Service Engineer Name : Ahzel Rahman
 Signature:
 Date: 30/09/24
 Contact Number: 9567562845

Customer Name : Munshudu V
 Signature:
 Date: 30/09/24
 Contact Number: 7558868130
 Designation : LAB JUNCTION
 Hospital Seal



AION BACKUPS

YP/VIII/320 AYLARA, KERALA KOLLAM

Date: 10/10/2024

Contact: 8921950217

Email: aionbackups12@gmail.com

GSTIN: 32CGUPM9450L1ZN

APC 600VA UPS SERVICE PROPOSAL

TO,

CYRIX HEALTH CARE PVT .LTD

Barcode:1043263 Ticket ID:137570

Hospital : PHC EDAPATTA

Dear Sir/Madam,

We are pleased to submit here with our quote for the same as per below.

SL	DESCRIPTION	QTY	UNIT RATE IN INR	TOTAL	GST	TOTAL PRICE IN INR
1	12-230 INVERTER PCB	1	1100.00	1100.00	18	1298.00
2	7Ah BATTERY	1	1200.00	1200.00	28	1536.00
3	TRANSFORMER	1	550.00	550.00	18	649.00
PAYABLE AMOUNT						3483.00

Commercial terms and conditions

1. Payment 80% advance along with PO balance before dispatch, 20% after installation.
2. Delivery period; 2weeks.
3. Tax :GST@18% for spares,28%@Battery
4. Installation; Our scope.

Please arrange to release the PO at our above address

Bank name :HDFC

Account no:50200082743176

IFSC Code:HDFC0004083

We hope you will find the above offer in line with our discussions and await your valuable order please free to call or mail us on above given number if any query.

Thank you, Sincerely,
BHAGATH MOHAN



Handwritten signature and text:
Bhagath Mohan
Pin: 679326
കുടുംബാരോഗ്യകേന്ദ്രം
എഡപ്പട്ട - 679326



1043263 UPS MACHINE NEED MAIN BOARD & BATTERY SPARE QUOTATION

3 messages

Gokul K T <gokulcyrix@gmail.com>
To: aionbackups12@gmail.com
Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Mon, 7 Oct, 2024 at 12:46 pm

Dear Sir, Madam

Can you send the Price quotation for UPS main board & Battery Make :APC, model :Back -UPS 600

Barcode:1043263
Ticket id:137570
UPS
PHC EDAPATTA

GOKUL KT
CYRIX HEALTHCARE Pvt Ltd
DISTRICT INCHARGE
MALAPPURAM
9809314280

Gokul K T <gokulcyrix@gmail.com>
To: aionbackups12@gmail.com
Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Thu, 10 Oct, 2024 at 6:42 pm

Reminder
[Quoted text hidden]

Bhagath Mohan <aionbackups12@gmail.com>
To: Gokul K T <gokulcyrix@gmail.com>

Thu, 10 Oct, 2024 at 11:57 pm

Thanks & Regards

Bhagath Mohan
PH:-8921950217

[Quoted text hidden]



UPS complaint
[Signature]
To: Anil K. Jaleel
Harlem
Ph no: 3819
ബഹിരൻ ബഹിസർ
കുടുംബാരോഗ്യസേവകമ്മീഷൻ
എടപ്പു - 679326

