

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: DH KANNUR	Name of District: KANNUR
Name of Equipment: NEBULIZER	Equipment ID & Barcode: #138841-1310941
Make: MEDITECH	Model: HANDY NEB
Serial Number: NA	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 28-02-2015	Present status of the equipment: Fully damaged
Date of breakdown: 05.10.2024 Date of registration of complaint through email/ Toll free): Toll free	*AMC/CAMC Period agreed at the time of purchase: NO AMC/ CAMC
Action taken: Checked the Equipment found Equipment is not switching ON, found that Motor, piston and switch is defective. Need to replace these spares for further checking.	
Recommendations for repair (required service details): No recommendations	
Cost of spares (specify parts and cost): N/A	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: N/A	Asset Value: 1353/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Service report attached	
Reasons for recommending the equipment as BER: Checked the Equipment found Equipment is not switching ON, found that Motor, piston and switch is defective. Need to replace these spares for further checking. Equipment Installed On 28.02.2015 aged up to 9 years 7 months. Quotation not submitted since spares are not available in the market. So we recommending for condemnation.	

Avinash
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: *checked & identified that machine not working due to the motor, Piston and switch are defective. aged upto 9 years & above. The spare cost is above asset value. So recommended for BER as per BER Protocol.*

Recommended for BER (Yes/ No):
 Date: *16/10/24*

Rajesh
 JUNIOR CONSULTANT BIOMEDICAL
 NATIONAL HEALTH
 KANNUR - 670002
 Signature of JC BM (NHM)

Date: *18/10/24* Seal  Signature of Superintendent / Medical Officer (i/c) *[Signature]*
 RESIDENT MEDICAL OFFICER
 DISTRICT HOSPITAL
 KANNUR

*Not mandatory # Based on the period of life and value as per the BER guidelines * Attach Photograph



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1011496

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>D.A Kanpur</u>	Call Registration Date : .. <u>5-10-24</u>
Address..... <u>Kanpur</u>	Caller ID : .. <u>138841</u>
..... <u>Kerala</u>	Date of Visit : .. <u>7-10-24</u>
Ph:..... <u>9633829216.</u>	Asset No : .. <u>1310941</u>
	EQPT Name : .. <u>Nebuliser</u>
	Manufacture..... <u>medidex</u> Model : <u>handy.neb</u>
	S.No..... <u>na</u> Dept. <u>M.M.H.D.</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : .. water, piston, switch are defective

Action Taken : .. checked the equipment. Found. Equipment is not switching on. Found that water, piston, switch is defective. need to replace these spares for further checking.


Completed Date : .. 7-10-24 Time : .. 11:00 Am Spare Required

Spare Replaced Requested


Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
	<u>Shyamkrishnan</u>	<u>7-10-24</u>	<u>10:00 am</u>

Customer Remark



Service Engineer Name : Shyamkrishnan Customer Name : Sanitha K E
 Signature: [Signature] Signature: [Signature]
 Date: 7-10-24 Date: 11/10/2024
 Contact Number: 8281188514 Contact Number: 9633829216
 Designation : SNO
 Hospital Seal



STOCK REGISTER OF MEDICINES/EQUIPMENTS

Name of Item: Nebuliser Electrical

Date	From whom received or to whom issued	Receipt	Issue	Balance	Batch No.	Expiry Date	Price	Remarks
30/12/08	Tr in from old stock book vol. 1 page no. 173 Rd 25.10.03 B (1955) Rd 30.9.08. (Donation)			2				
	condemned as per order No. B/8991/13 Dated 3.7.2013		2	nil				
28/2/15	Rd from store (Air Compressor)	2		2			Rs 1402	
21.11.15	condemned as per order A/15900/2016 dated 12.6.17 used from		*	*			Rs 3500 each	
2/9/18	Tr. out to Unserviceable Register P. No. 24 condemned as per order no. of. reg. no. 38325 2021 order 3-8-2012		1	1				Single 8 No more left
30/3/23	Tr. out to MM IJ Instrument & equipment Registers vol. V page no. 19		1	NIL				



[Signature]
18/10/24
MEDICAL OFFICER
DISTRICT HOSPITAL
KANNUR



Sindher
 21/0 10/12
[Signature]

RESIDENT MEDICAL OFFICER
 DISTRICT HOSPITAL
 KANNUR



[Signature]
 18/10/24

RESIDENT MEDICAL OFFICER
 DISTRICT HOSPITAL
 KANNUR

