

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC MANGAD	Name of District: KOZHIKODE
Name of Equipment: CENTRIFUGE	Equipment ID & Barcode: 138371 & 1142919
Make: LABY INSTRUMENT INDUSTRY	Model: T-854/8
Serial Number: 2016-17/274	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 30/03/2017	Present status of the equipment: Fully damaged
Date of breakdown: 3/10/2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked and found Motor, tube head (swingout head 8*15 ml) & switch defective. Enquired spares with oem	
Recommendations for repair (required service details): Not recommending for repair	
Cost of spares (specify parts and cost): N/A	
#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: N/A	Asset Value: 9651/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report attached and vendor EOL Letter attached	
Reasons for recommending the equipment as BER: Checked and found Motor, tube head (swingout head 8*15 ml) & switch defective. Equipment installed on 30/03/2017 (stock book attached for reference). Equipment out lived 7 years 6 months, OEM informed that this was an discontinued model so spare parts not available. So As per tender clause 5.3.14.2 We recommending the equipment for condemnation	
JINEESH PP <i>JTP</i> 14/10/24 Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM: Checked & found that motor, tube head & switch damaged Equipment installed on 30/03/2017. OEM informed that it was discontinued model. Spares are not available. So recommended for condemnation	
Recommended for BER (Yes/ No):	Signature of JC BM (NHM) <i>Boyy P.L.</i>
Date: 14/10/24	Consultant Bio Medical Engineer National Health Mission Kozhikode Pharmacist
Date: 16/10/24	Signature of Superintendent / Medical Officer (i/c)

*Not mandatory

#Based on the period of life and value as per the BER guidelines * Attach Photograph



MEDICAL OFFICER
 FAMILY HEALTH CENTRE
 MANGAD, UNNIKULAM
 KOZHIKODE D.T-673574



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No :1013651

CYRIX[®]

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

<p>Health Facility..... <u>PHC</u></p> <p>Address..... <u>Mangad</u> <u>Kozhikode</u></p> <p>Ph:..... <u>9400576193</u></p>	<p>Call Registration Date : <u>03-10-2024</u></p> <p>Caller ID : <u>138371</u></p> <p>Date of Visit : <u>04-10-2024</u></p> <p>Asset No : <u>1142919</u></p> <p>EQPT Name : <u>Centrifuge</u></p> <p>Manufacture <u>Laby</u> Model : <u>T-8M</u></p> <p>S.No <u>2016-17/274</u> Dept. <u>Laboratory</u></p>
---	---

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Not working

Action Taken : checked and found Motor, tube head (swingout head & 15ml) and switch defective. Enquired. Spare with OEM

Completed Date : 4-10-24 Time : 3:00pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Mulharajk</u>	<u>4-10-24</u>	<u>2:00pm</u>	<u>3:00pm</u>

Customer Remark _____ Completed Pending For M-O

Service Engineer Name : Mulharajk
Signature: [Signature]
Date: 4-10-24
Contact Number: 9072522081

NATIONAL HEALTH MISSION

Customer Name : Shajina M
Signature: _____
Date: 04/10/24
Contact Number: 9400576193
Designation: Lab
Hospital: [Signature]

TECHNICAL OFFICER
FAMILY HEALTH CENTRE
MANGAD, WILKULAM
KOZHIKODE-T-673574



Required Spare Quotation - 1142919-138371- Centrifuge- PHC MANGAD,KOZHICODE

2 messages

Jineesh P P <jineeshcyrix@gmail.com>

Wed, 9 Oct, 2024 at 10:11 am

To: Mahesh Kumar Singal <laby63@gmail.com>

Cc: Anjitha T L <purchasecyrix2@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, kmidhuncyrix@gmail.com

Dear Sir

Please share price and spare availability of motor, tube head (swingout head 8×15 ml) & switch (speed control knob)

Centrifuge

Make : Laby instrument industry

Model: T- 854/8

SN: 2016-17/274

Attached photos for your reference

Thanks & Regards,

Jineesh PP

District Incharge - Kozhikode

Mob-7593847163

jineeshcyrix@gmail.com

www.cyrixhealthcare.com

Cyrix Healthcare Pvt Ltd

| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India

KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH

Mahesh Kumar Singal <laby63@gmail.com>

Wed, 9 Oct, 2024 at 11:41 am

To: Jineesh P P <jineeshcyrix@gmail.com>

Cc: Anjitha T L <purchasecyrix2@gmail.com>, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, kmidhuncyrix@gmail.com

Dear Sir,

We have discontinued this model so we do not have spare part of this machine

Thanking you

Best Regards

Mahesh Singal (C.E.O)

For **Laby Instruments Industry**

62, 1st Floor, Industrial Estate,

Ambala Cantt-133006,

MB. :- 9992584615

Email:- laby63@gmail.com, Website : www.labyindia.com,

GST No. : - 06ACTPS5454M2Z0

You can also visit on our website www.labyindia.com for accessing other products and price lists etc

MICROPROCESS BASED COLDRIMETER WITH 8FIL

30/3/17 Received from the pharmacy
 model CL 223 . invoice NO:
 054/16-17, dated 21/1/17.
 purchased from SMER (Scientific
 Ipic Medical Equipment) Pvt
 Ltd. S.L. NO: 2948. Warranty
 up to 1 yr.

in use (lab) I NO NIL

CENTRIFUGE SMALL MODEL P-854/8

30/3/17 Received from the pharmacy
 invoice NO- UA- 64/2017 I NO
 Supplied by Universal Agencies I NO

in use (lab) I NO NIL




For MEDICAL OFFICER
 FAMILY HEALTH CENTRE
 MANGAD, UNNIKULAM
 KOZHIKODE DT-673574

[Handwritten signature]
 Pharmacist



LABY™ Model-T-864/8
Centrifuge Machine
 SWING OUT HEAD 6X16ml
 Works on: 220-240 Volts 60Hz A.C.
 Manufactured by:
Laby Instrument Industry
 www.labyindia.com Sr. No. 2016-17/274



Medical Officer
 Family Health Centre
 NANGAD, UNNIKULAM
 P.O. ZHIKOLE D.T-673574
Pharmacist