


**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: THQH KOYILANDY		Name of District: KOZHIKODE
Name of Equipment: BP APPARATUS		Equipment ID & Barcode: 138731 & 1121556
Make: N/A		Model: N/A
Serial Number: N/A		Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: Ticket Master Installation Date Is 11/06/2020 As Per Stock Book Installation Date Is 29/03/2019 (Stock Book Attached For Reference)		Present status of the equipment: Fully damaged
Date of breakdown: 04/10/2024		*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Date of registration of complaint through email/Toll free): Toll free		
Action taken: Checked and found main board and motor defective. Need to replace these spares for further checking of working condition of the equipment.		
Recommendations for repair (required service details): Not recommending for repair		
Cost of spares (specify parts and cost): N/A		
#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: N/A		Asset Value: 2200/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report attached		
Reasons for recommending the equipment as BER: Checked and found main board and motor defective. Need to replace these spares for further checking of working condition of the equipment. In Ticket master installation date is 11/06/2020, As per stock book installation date is 29/03/2019 (stock book attached for reference). Equipment out lived 5 years 6 months. Quotation not submitted since spares are not available in the market so we recommend the equipment for condemnation		
		JINEESH PP <i>JPP</i> 14/10/2024 Name & Signature of CYRIX Authority with date
Remarks of Junior Consultant (Biomedical) NHM: <i>Checked & found that the main board & motor are damaged. Machine is installed on 29/03/2019. Spares are not available in the market</i>		
Recommended for BER (Yes/ No):		Consultant Bio Medical Engineer National Health Mission Kozhikode Signature of JC BM (NHM)
Date: 14/10/2024		
Date: 15/10/2024		Signature of Superintendent Medical Officer (i/c)

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph



[Signature]
15/10/24
SUPERINTENDENT
Taluk Head Quarters Hospital
Koyilandy

STOCK REGISTER OF B.P. APPARATUS (L.P) - 14 -


Name of Article.....

Date	No. and date of voucher or invoice	From whom received or to whom issued	Receipt	Issued	Balance after each transaction	Initials of Receiver	Remarks
		Brought forward			3		
25/9/18		Received from the store (L.P)	1		4		(Rs: 1500/-)
26/9/18		Put into use		1			<u>Paul</u>
1/8/15		Tr. out to unuserviceable	1	1	30		
		Page No: 9					
20/10/18		Rd from the store (Donation)	1		4		
3/4/18		Condemned as per order	1		3		
		Cy/44 72/12 dated 3/4/18			3		
29/3/19		Received from the store (Medicinal type) (L.P) LIFE CARE Exp 12/2028	1		4		Rs 2100/- <u>Paul</u>
16/8/2019		Tr. out to unuserviceable		2			<u>Paul</u>
		Register Page No: 9.			2		
20/8/2019		Trout to unuserviceable		1			<u>Paul</u>
7/2/2020		Rd from store (Donated items)	1		3		<u>Paul</u>
		included in (S1) B. case No (8011 2379)					
7/8/2022		Rd from the store 1			4		<u>Paul</u>
		(8011 2380) multoused			2		
		Rs. 1875/-					
12/11/2022		Rd from the store 1			5		<u>Paul</u>
		Put in use 5					
7/10/2023		Trnsf from ward 5, 6. Vol: 2 Stockbook: 3. Page: 58		3			<u>Paul</u>
		Carried over					
		Transferred to unuserviceable Register.		3			



SUPERINTENDENT
Taluk Head Quarters Hospital
Koyilandy

8011 2380



Handwritten signature
 15/10/24

SUPERINTENDENT
 Taluk Head Quarters Hospital
 Koyilandy

