


**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**


Name of Hospital: PHC ANDATHODE	Name of District: THRISSUR
Name of Equipment: B P Apparatus	Equipment ID & Barcode: 136488 & 0840520
Make: NA	Model: Elko
Serial Number: 2008059727	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: Ticket master installation date – 18-09-2021, As per Stock book installation date – 16-02-2021	Present status of the equipment: Fully damaged
Date of breakdown: 23-09-2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Machine checked and found display and motor are defective. Need to replace these spares for further service.	
Recommendations for repair (required service details): Not recommending	
Cost of spares (specify parts and cost): N/A	


#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: N/A	Asset Value: 2200/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix report attached	


Reasons for recommending the equipment as BER: checked the machine and found display and motor are defective. In ticket master installation date is 18-09-2021 As per stock book Machine installed on 16-02-2021 aged up to 3 years 7 months. Quotation not submitted since Spears are not available in market. So, we are recommending for condemnation.


BIJO T JOY
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
Machine is more than 3 year old.



Recommended for BER (Yes/ No): <i>Yes</i>	 Signature of JC BM (NHM)
Date:	

 Seal	 Signature of Superintendent Medical Officer (i/c)
---	--



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1007765

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>Pnc</u>	Call Registration Date : <u>23/09/24</u>
Address..... <u>Andathode</u>	Caller ID : <u>136488</u>
..... <u>Thirissur</u>	Date of Visit : <u>25/09/24</u>
Ph:..... <u>9846165538</u>	Asset No : <u>0840520</u>
	EQPT Name : <u>Bp apparatus</u>
	Manufacture..... <u>N/A</u> Model : <u>N/A</u>
	S. No..... <u>N/A</u> Dept..... <u>OP</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : equipment not getting on

Action Taken : machine checked and found display and motor are defective. Need to replace these spares for further service

Completed Date : 25/09/24 Time : 1:00pm Spare Required

Spare Replaced Requested


Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Christy</u>	<u>25/09/24</u>	<u>12:00pm</u>	<u>1:00pm</u>

Customer Remark Completed Pending

Service Engineer Name : Christy
Signature: [Signature]
Date: 25/09/24
Contact Number: 6238844347

Customer Name : Shenb. k.m
Signature: [Signature]
Date: 25/9/24
Contact Number: 9846165538
Designation: Nursing officer
Hospital Seal: [Seal]



MEDICAL OFFICER
ANDATHODE
Poonithura
Thirissur
PIN: 682038

