

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: GH THRISSUR	Name of District: THRISSUR
Name of Equipment: Ocular Ultrasound	Equipment ID & Barcode: 135510 & 0810906
Make: Sonomed	Model: Pac Scan 300 A
Serial Number: 0300-1210-7249	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: Ticket master installation date – 03-05-2013, As per Stock book installation date is 05.03.2013	Present status of the equipment: Fully damaged
Date of breakdown: 18-09-2024	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Date of registration of complaint through email/Toll free): Toll free	
Action taken: Machine checked and found the probe is defective. Need to replace probe for further checking working condition of equipment and spare enquired	

Recommendations for repair (required service details): Not recommending

Cost of spares (specify parts and cost): A probe for 300A Pac Scan – 150000/-

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 98.99%	Asset Value: 151525/-
---	-----------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): OEM Quotation and cyrix report attached

Reasons for recommending the equipment as BER: Machine checked and found probe is defective. In ticket master installation date is 03-05-2013 and As per stock book Machine installed on 05.03.2013 aged up to 11 years and 6 months. The spare cost is 98.99% with respect to asset value and both criteria met for rber. So as per tender clause 5.3.14.1 equipment recommending for condemnation.

BIJO T JOY
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:

Spare cost is high


Dr. Consultant (Biomedical)
NHM, Thrissur

Recommended for BER (Yes/ No): *Yes*

Date:

27/9/24
Signature of JC BM (NHM)

Date:



Seal

30/09/24
Signature of Superintendent / Medical Officer (i/c)

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph

THRISSUR - PIN: 680 001



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1017981

CYRIX®

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>GH</u>	Call Registration Date : <u>18-09-2024</u>
Address..... <u>THRISSUR</u>	Caller ID : <u>135510</u>
..... <u>KERALAM</u>	Date of Visit : <u>19-09-2024</u>
Ph:..... <u>9496200951</u>	Asset No : <u>0810906</u>
	EQPT Name : <u>Ocular ultrasound</u>
	Manufacture..... <u>Sonomed</u> Model : <u>Pac Scan 300A</u>
	S. No. <u>0300-1210-7249</u> Dept. <u>EYE OP</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Ocular ultrasound not working.

Action Taken : Machine checked. and found the problem is defective. Need to replace probe for further checking. working condition. of camera and spare examined.

Completed Date : 19-09-2024 Time : 11:00am Spare Required

Spare Replaced Requested


Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Ahsil</u>	<u>19-9-24</u>	<u>10:00am</u>	<u>11:00am</u>

Customer Remark Completed Pending

Service Engineer Name : Ahsil
Signature: [Signature]
Date: 19/09/2024
Contact Number: 9496200951

Customer Name :
Signature: [Signature]
Date:
Contact Number:
Designation :
Hospital Seal



SUPERINTENDENT (Dy. DHS)
GENERAL HOSPITAL
THRISSUR - PIN: 680 001



QUOTATION - GH THRISSUR - 0810906-135510

3 messages

Bijo T Joy <bijocyrix@gmail.com>

To: info@lumenmediworks.com

Mon, 23 Sept, 2024 at 11:42 am

Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear sir,

There is a complaint reported from GH THRISSUR, Thrissur, Kerala, Regarding the complaint with SONOMED, Pac Scan 300 A Ascan machine ,Please share the price and availability of the **Direct contact A-scan Probe**, as soon as possible.

Equipment - A scan

Make : SONOMED Inc

Model: PacScan 300A

Serial No: 0300-1210-7249

Hospital- GH Thrissur

District - Thrissur

Barcode - 0810906

Caller id - 135510

With Regards,

B I J O T J O Y

DISTRICT INCHARGE THRISSUR,

CYRIX HEATHCARE PRIVATE LIMITED

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia

M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

Lumen Mediworks <info@lumenmediworks.com>

Reply to: info@lumenmediworks.com

To: Bijo T Joy <bijocyrix@gmail.com>

Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, dilshad@lumenmediworks.com, Aravind Sukumar <aravind@lumenmediworks.com>

Mon, 23 Sept, 2024 at 12:27 pm

Dear Sir,

Thanks for your mail . A probe available for this model 300A.

From your mail , we understand the Unit is working well and only Probe required.

Cost of A probe for 300A Pac scan — Rs. 1,50,000/- Plus GST 12%.

50% payment Advance with PO and balance against Supply.

Delivery period : 3 weeks on receipt of Order with Advance.

Thanking you.

Best Regards.



LUMEN
MEDIWORKS (P) LTD

Ph. 0484 3551448, +91 6238826591

Email id – info@lumenmediworks.com

Website – www.lumenmediworks.com

[Quoted text hidden]

Bijo T Joy <bijocyrix@gmail.com>

Mon, 30 Sept, 2024 at 2:54 pm

To: info@lumenmediworks.com

Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>, dilshad@lumenmediworks.com, Aravind Sukumar

<aravind@lumenmediworks.com>

Kindly provide the quotation

[Quoted text hidden]

P: 126
Book (2)

STOCK REGISTER

Name

A-Scan

Date	Received from or Issued to	Inv.No. or Issue	Rate		Receipts	Issues	Balance	Remarks
			Rupees	Ps				
27-3-00	STORE	DHS.			-1-	OP		wing
23-3-11	STORE				-1-	OP		une.
	Model 300N E probe.	kmgee						
	(Installed on 5/3/2013)							
31/3/19	working Satisfactory				-1-		-2-	
3/7/2020	working Satisfactory				-1-		2	
31/3/2021	working Satisfactory				-1-		2	
31/3/2022	working Satisfactory				-1-		2	
25/2/22	working							(1) Not using.
25/4/21	One working							(2) Not using (advised Condemned)
31/1/24	One working							
31/3/24	One working							

[Handwritten signature]



