

**PROFORMA**  
**RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)**  
**BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: PHC KODIYERI	Name of District: KANNUR
Name of Equipment: WEIGHING MACHINE	Equipment ID & Barcode: #137617-1345028
Make: SAMSO	Model: NA
Serial Number: NA	Warranty details: NO Warranty
Date of purchase/ Year of manufacture/ Installation Date: 10-03-2019	Present status of the equipment: Fully damaged
Date of breakdown: 28-09-2024 Date of registration of complaint through email/ Toll free: Toll free	*AMC/CAMC Period agreed at the time of purchase: NO AMC/ CAMC

Action taken: Checked the machine and found that Machine is not switching on, Found main board, Load cells damaged. Need to replace these spares for further checking.

Recommendations for repair (required service details): No recommendations

Cost of spares (specify parts and cost): N/A

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: N/A	Asset Value: 1800/-
---	---------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Service report attached


Reasons for recommending the equipment as BER: Checked the machine and found that machine is not switching ON. Found that main board and load cells damaged. Equipment installed on 10-03-2019. Aged up to 5 years 6 months. Quotation not submitted since spares are not available in the market. So we recommending for condemnation.

Avinash  09/10/24  
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: *checked & identified that the machine not switching on, aged upto 5 years. and spare also not available, so recommended for BER as per BER protocol.*


Recommended for BER (Yes/ No):

Date: 09/10/24

 09/10/2024  
 Signature of JC BM (NHM)

Date: \_\_\_\_\_ Seal \_\_\_\_\_ Signature of Superintendent / Medical Officer (i/c)

\*Not mandatory



# Based on the period of life and value as per the BER guidelines

\* Attach Photograph

**മെഡിക്കൽ ഓഫീസർ**  
**എ.മി.കെ.എ. കോളേജ് - കോടിയേര**  
**കെ. മുഴിക്കര, കണ്ണൂർ - 670 103**  
**ഫോൺ: 0490 2356107**



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 1011533

**CYRIX**<sup>®</sup>  
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility.....PHC...KODIYERI.....	Call Registration Date : 28/09/2024.....
Address.....KODIYERI.....	Caller ID : 137617.....
.....KANNUR.....	Date of Visit : 30/09/2024.....
Ph : 9947506283.....	Asset No : 1345028.....
	EQPT Name : Weighing Machine.....
	Manufacture.....Samso..... Model : NA.....
	S.No.....NA..... Dept.....OP.....

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : Equipment is not switching ON.....

Action Taken : checked the equipment and found equipment is not switching on. Found that main board and load cells damaged. Enquired for quotation.....

Completed  Date : 30/09/2024..... Time : 1:00pm..... Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
Prathyush T. Bhaskar	30/09/2024	12:20pm	1:00pm


Customer Remark Completed

*for* *Prathyush*

Pending  
മുൻകരുതൽ നടപടികൾ  
എടുക്കപ്പെട്ടിട്ടുണ്ട്. കേന്ദ്രം-കോഴിക്കോട്  
ഒ. മുതിരമം. കണ്ണൂർ - 074  
ഫോൺ നമ്പർ : 9947506283

Service Engineer Name : Prathyush T. Bhaskar  
Signature: *Prathyush*  
Date: 30/09/2024  
Contact Number: 9947025643

Customer Name : Sindhya Ram  
Signature: *Sindhya Ram*  
Date: 30/9/2024  
Contact Number: 9947506283  
Designation : Nurse Officer gr I  
Hospital Seal




# STOCK REGISTER OF MEDICINES

151

Name of Items:

B.P. Apparatus (K.C.D)

Date	No. & Date of Invoice/Voucher	From Whom received/ to whom issued	Received	Issued	Balance	Remarks
10/3/19	Pl store (Ardran supply)		2		2	
11/10/23	1 (LED	BP apparatus. Trout for purpose.		MLPS for field } MLSP Shahnig. }		
10/3/19	Received from store (K.C.D)	Weighing machine (Ardran supply)	2		2	




*[Handwritten signature]*

ഓഡിറ്റർ (ആരോഗ്യം)  
 ഗവൺമെന്റ് കോളേജ് - കോട്ടയം  
 കെ. മുഹമ്മദ്, കല്ലൂർ - 670 103  
 ഫോൺ 0490 2356113

SAMSO

**CYRIX**  
LINEAR REFILL  
CN: phkily  
PRO: Digital weighing Mach  
ID: 1345028  
CALIB: 10-01-24  
CAL DUE: 09-01-25  
ISO 9001: 2008 Certified Co. Ph: 8764155072



**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)  
TOLL FREE NO: 1800 - 425-7669  
BAR CODE - (8004-890615225) 1345028



By *[Signature]*

കേരള സർക്കാർ  
ആരോഗ്യ വകുപ്പ്  
8, ഗവൺമെന്റ് റോഡ്  
കൊച്ചി 682 002