

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: TH PANATHADY	Name of District: KASARGOD
Name of Equipment: ELECTROLYTE ANALYZER	Equipment ID & Barcode: 133616- 1420464
Make: MEDICA	Model: EASY LYTE
Serial Number: 47353CNAK	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket mater installation date is 09/02/2017, As per stock register installation date is 19/3/2014 (Stock book attached for reference)	Present status of the equipment: Fully damaged
Date of breakdown: 05/09/24 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC

Action taken: Checked the machine, found CPU board, Probe motor assembly and electrodes are defective. Enquired OEM for quotation.

Recommendations for repair (required service details): No Recommendation

Cost of spares (specify parts and cost): CPU BOARD NA/K- 90505/-
POTTASIMUM ELECTRODE- 14375/-
SODIUM ELECTRODE- 16100/-
REFERENCE ELECTRODE- 16215/-
PROBE MOTOR ASSEMBLY WITH LEAD SCREW- 38985/-
TOTAL: 176180+ (GST 31712) =207892/-

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 277.1% Asset Value: 75000/-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): CYRIX Service report and OEM quotation attached


Reasons for recommending the equipment as BER: Checked the machine, found machine CPU board, Potassium electrode, Sodium electrode, Reference electrode and Probe motor assembly are defective. In ticket master installation date is 09-02-2017, As per stock installation date is 19/03/2014 and aged up to 10 years and 8 months. Repairing cost of equipment is more than 100% with respect to asset value. Both criteria for RBER met. Considering the tender clause 5.3.14.1, We recommending the equipment for condemnation.

ABHINAV O 24.09.2024
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: Checked the machine and found machine CPU board, Pottassium electrode, sodium electrode, reference electrode & Probe motor assembly are defective. Equipment installed on 19/03/2014 (stock book details attached). Repairing cost of the equipment is more than 100%. With respect to asset value OEM quotation and cyrix service report attached.

Recommended for BER (Yes/ No) _____
Signature of J.C. CONSULTANT Biomedical Engineering (NHM) Kasaragod at Kasaragod

Date: 27/09/2024



Date: 28.09.2024

Signature of Superintendent / Medical Officer (i/c)

*Not mandatory

#Based on the period of life and value as per the BER guidelines.

Attach Photograph

NHM KASARGOD



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1002266

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi-682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility... <i>Taluk Hospital.</i>	Call Registration Date : .. <i>05.09.2024</i>
Address..... <i>Panathady</i>	Caller ID : .. <i>133616</i>
..... <i>Kasaragod.</i>	Date of Visit : .. <i>06.09.2024</i>
Ph : .. <i>894-3119688</i>	Asset No : .. <i>1420464</i>
	EQPT Name : <i>Electrolyte Analyzer</i>
	Manufacture <i>Medica</i> Model : <i>Easy Lyte</i>
	S. No... <i>47353LNAK</i>Dept. <i>Laboratory</i>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : ..*Not Initialising.*.....

Action Taken : ..*Checked the machine and found problem with machine CPU Board, Probe motor assembly with lead screw, Potassium electrode, Sodium electrode and reference electrode. Need to replace these spares for further servicing of the equipment.*.....

Completed Date : ..*06.09.24*..... Time : ..*3.30pm*..... Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
	<i>06.09.24</i>	<i>1.30pm</i>	<i>3.30 pm</i>

Customer Remark Completed Pending

Service Engineer Name : *Abhinav O*
Signature: *[Signature]*
Date: *06.09.2024*
Contact Number: *8948392989*



Customer Name : *DIVU. E.C*
Signature: *[Signature]*
Date: *06.09.24*
Contact Number: *9996970078*
Designation : *Lab Technician*
Hospital Seal: *[Seal]*

QUOTATION


Bill To: CYRIX HEALTHCARE PVT LTD
First Floor,petta Junction 30/641b, Poonithura,Ernakulam, Kerala,
Pincode 682038 , , 673508

Ref. No : 2425QUOTE001933
Date : Sep 19, 2024 2:34 PM
Validity : 30 days
TBM GSTIN : 27AAACTP2038C1ZT

Ship to, THQH Panathady, Kasargod

SUB : QUOTATION FOR SUPPLY/REPLACEMENT OF SPARES IN INSTRUMENT MODEL: EASYLITE (NA/K) ANALYSER MODEL - NA/K (2004) SR.NO.: 47353CNAK

With reference to the above, following Spares / Consumables are required to be replaced / supplied. Price/s of the same is / are given below, for approval:

SR. NO	Part Code	HSN Code	Description of Spares/Consumables	Qty	Unit Price	Discount%	GST%	Amount	
1	160303	90278990	CPU BOARD NA / K	1	90505		18	90505	
2	160267	90278990	POTTASIMUM ELECTRODE CAT 2101	1	14375		18	14375	
3	160269	90278990	SODIUM ELECTRODE CAT 2102	1	16100		18	16100	
4	160281	90278990	REFERENCE ELCTRODE CAT NO.2103	1	16215		18	16215	
5	160294	90278990	PROBE MOTOR ASSEMBLY WITH LEAD SCREW CAT 001392-001	1	38985		18	38985	
Merchant Name : TRANSASIA BIO-MEDICALS LTD								Total (Rs)	176180
<p>SCAN AND PAY WITH UPI QR</p> <p>Merchant Name : TRANSASIA BIO MEDICALS LTD</p>  <p>Payee VPA : transasiabiomedicalsltd@citibank</p>								Discount (Rs.) (-)	
								GST (Rs.)	31712.4
								Grand Total (Rs.)	207892.4
Rupees in words: Two lakhs Seven thousand Eight hundred Ninety Two rupees and Four paise only									

TERMS AND CONDITIONS :

- Payment :** Full payment should be made in advance through NEFT / RTGS/ Demand Draft in favour of "**Transasia Bio-Medicals Ltd**" to **Bank A/C no.:** 00600330002538 **IFSC code:** HDFC0000060, **UPI ID :** transasiabiomedicalsltd@citibank. You can also pay us through **UPI apps like BHIM/ Google Pay/ Paytm/PhonePe** etc. by scanning the above mentioned code. Cash Payment is strictly prohibited. No employee of the TBM is authorized to collect cash and or issue receipt against collection of Cash.
- GST :** As applicable
- Other Cost / Local duties :** Customer's responsibility
- Delivery :** Within 1-2 weeks from the receipt of your work order with payment confirmation

Thanking You
Your faithfully,
For **TRANSASIA BIO-MEDICALS LTD**

RECEIVED & ACCEPTED
CUSTOMER SIGNATURE & STAMP:
CUSTOMER NAME: _____
CONTACT NO: _____

TRANSASIA BIO-MEDICALS LTD.

Head Office: Transasia House. 8. Chandivali Studio Road, Andheri (E), Mumbai - 400 072. Tel.: (022) 4030 9000 Fax: (022) 2857 3030 Mumbai : Toll Free No. 1800 103 8226, Whatsapp No.: 8451048434, SMS No.: 9212433444



Jayesh Krishnan <jayeshcyrix@gmail.com>

PRICE ENQUIRY 1420464

6 messages

Jayesh Krishnan <jayeshcyrix@gmail.com>

Tue, Sep 10, 2024 at 10:33 AM

To: SANDEEP DAS M <s.das@transasia.co.in>

Cc: s.john@transasia.co.in, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear sir,

Kindly share price and availability of Main board, sodium electrode, Potassium electrode, reference electrode and probe assembly for electrolyte analyzer.

Make: Medica

model: Easy Lyte

S/N: 47353CNAK

--

Thanks & Regards

JAYESH KRISHNAN,
Sr. Biomedical Engineer,
District In-charge
Kasargod.
PH:- 7593847156

jayeshcyrix@gmail.comwww.cyrixhealthcare.com**Cyrix Healthcare Pvt Ltd****| 30/641 B | Pettah Jn | Poonithura | Cochin -38 | Kerala | India****KERALA || KARNATAKA || TAMIL NADU || UTTAR PRADESH**

Jayesh Krishnan <jayeshcyrix@gmail.com>

Wed, Sep 11, 2024 at 12:49 PM

To: SANDEEP DAS M <s.das@transasia.co.in>

Cc: s.john@transasia.co.in, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Sir,

Kindly Update.

[Quoted text hidden]

Jayesh Krishnan <jayeshcyrix@gmail.com>

Thu, Sep 12, 2024 at 7:13 PM

To: SANDEEP DAS M <s.das@transasia.co.in>

Cc: s.john@transasia.co.in, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Sir,

Kindly update urgently.

[Quoted text hidden]

Jayesh Krishnan <jayeshcyrix@gmail.com>

Mon, Sep 16, 2024 at 8:49 PM

To: SANDEEP DAS M <s.das@transasia.co.in>

Cc: s.john@transasia.co.in, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Sir,

Kindly share the quote at the earliest.

[Quoted text hidden]

Jayesh Krishnan <jayeshcyrix@gmail.com>

Wed, Sep 18, 2024 at 9:19 AM

To: SANDEEP DAS M <s.das@transasia.co.in>

Cc: s.john@transasia.co.in, Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Sir,
Kindly update.
[Quoted text hidden]

SANDEEP DAS M <s.das@transasia.co.in>
To: Jayesh Krishnan <jayeshcyrix@gmail.com>
Cc: Blesson Jose <zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Thu, Sep 19, 2024 at 2:37 PM

Dear jayesh ,

Please find the quote of Easylyte spares

Get [Outlook for Android](#)

From: Jayesh Krishnan <jayeshcyrix@gmail.com>
Sent: Tuesday, September 10, 2024 10:33:53 AM
To: SANDEEP DAS M <s.das@transasia.co.in>
Cc: s.john@transasia.co.in <s.john@transasia.co.in>; Blesson Jose <zm2.klbemp@cyrix.in>; DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>
Subject: PRICE ENQUIRY 1420464

[Quoted text hidden]

 **Quote-SR.No.pdf**
1460K


099

SODIUM POTASSIUM ELECTROLYTE ANALYSER

Date	Issue Note No	MFG	EXP	Batch No	Receipt	Issue	Balance	Remarks
2/14	L.No: SMSEKSD/P88/198 Kerala Corporation Kochin 68 2020, Ab 0484-220565/ Supplied from Sidra Marketing Centre Kollam Vidyamangal. Ab: 04994 25593/ R.O. No: A/354/2013 18.2.14 G.M.O CHE Kollam.				1 Nos	RB. 174354.00		✓
3/14	I issued above item to lab (Lab Page: 202)	Received			One item Model: Easy type Firmys mfg: Easy type Lab: tech. mfg: Easy type installed on: 19.03.14	Nil SL No: HT353CA Type: TO-BAL 250.V mfg: Mech Cal.		✓

മെഡിക്കൽ ഓഫീസർ
 മെട്രിക് യൂണിറ്റ് സർവ്വീസ് ഓഫീസർ
 എം.എ.പി. കമ്മീഷൻ
 എ.എ.പി. കമ്മീഷൻ
 28/9/24





കൗൺസിലർ പി. ഒ. രാജപുരം
 കൗൺസിലറുടെ ഓഫീസ്
 പി. ഒ. രാജപുരം കൗൺസിലറുടെ ഓഫീസ്
 22/09/2024