

PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
 BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

| | |
|--|---|
| Name of Hospital: PHC PUZHATHY | Name of District: KANNUR |
| Name of Equipment: BP APPARATUS | Equipment ID & Barcode: 138287 #1343730 |
| Make: INFI CHECK | Model: INFI |
| Serial Number: NA | Warranty details: NO Warranty |
| Date of purchase/ Year of manufacture/ Installation Date: In ticket master installation date is 07-08-2020, As per stock book Installation date is 16.09.2019 (Stock book copy attached for reference) | Present status of the equipment: Fully damaged |
| Date of breakdown: 02.10.2024 Date of registration of complaint through email/ Toll free): Toll free | *AMC/CAMC Period agreed at the time of purchase: NO AMC/ CAMC |
| Action taken: Checked the Equipment found Equipment is not switching ON, found that Main Board, Pressure sensor and Display Defective. Need to replace these spare for further checking. | |


Recommendations for repair (required service details): No recommendations

Cost of spares (specify parts and cost): N/A

| | |
|---|---------------------|
| # Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: N/A | Asset Value: 2200/- |
|---|---------------------|

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Service report attached


Reasons for recommending the equipment as BER: Checked the Equipment found Equipment is not switching ON, found that Main Board, Pressure sensor and Display Defective. Need to replace these spare for further checking. In Ticket master Equipment installed on 07.08.2020. As per the stock book installation date is 16.09.2019 aged up to 5 Years. Quotation not submitted since spares are not available in the market. So we recommending for condemnation.

Avinash  09/10/24
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM: checked and identified that the machine not switching on aged upto 5 years and spare also not available, so recommended for BER as per BER protocol.

| | |
|--------------------------------|--|
| Recommended for BER (Yes/ No): |  09/10/24 Signature of JC BM (NHM) |
| Date: | Fin  Signature of Superintendent / Medical Officer (i/c) |
| Date: 10/10/24 | Seal  |

*Not mandatory # Based on the period of life and value as per the BER guidelines


 (പ്രാഥമിക ആരോഗ്യ) കേന്ദ്രം (PHC) കണ്ണൂർ
 കൊറ്റാളി പി.ഒ, കണ്ണൂർ - 670005



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1004098

CYRIX®

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

| | |
|--|---|
| Health Facility..... <u>PHC Puzhathy</u> | Call Registration Date : .. <u>02/10/24</u> |
| Address..... <u>Kannur</u> | Caller ID : .. <u>138287</u> |
| <u>Kottali</u> | Date of Visit : .. <u>03/10/24</u> |
| Ph : .. <u>9446493955</u> | Asset No : .. <u>13.43730</u> |
| | EQPT Name : .. <u>BP Apparatus</u> |
| | Manufacture .. <u>Inficheck</u> Model : <u>Inti</u> |
| | S. No. .. <u>NA</u> .. Dept. .. <u>OP</u> |

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : .. Machine is not switching ON

Action Taken : .. checked the equipment found equipment is not switching ON, found that main board, Pressure sensor and Display defective. need to replace these spares for further checking.

Completed Date : .. 03/10/24 .. Time : .. 11:00 AM .. Spare Required

Spare Replaced Requested

| Description | Qty. | Part Number | PR Number |
|-------------|------|-------------|-----------|
| 1. | | | |
| 2. | | | |
| 3. | | | |

| Cyrix Engineer | Date | Start Time | End Time |
|----------------|-----------------|-----------------|-----------------|
| <u>Avinash</u> | <u>03/10/24</u> | <u>10:00 AM</u> | <u>11:00 AM</u> |

Customer Remark

Completed

Pending

| | |
|--|--|
| Service Engineer Name : <u>Avinash</u> | Customer Name : <u>Divya</u> |
| Signature: <u>[Signature]</u> | Signature: <u>[Signature]</u> |
| Date: <u>03/10/24</u> | Date: <u>3/10/24</u> |
| Contact Number: <u>6364064948</u> | Contact Number: <u>9446493955</u> |
| | Designation : <u>Pharmacy</u> |
| | Hospital Seal: <u>മെഡിക്കൽ ഓഫീസർ</u> <u>പ്രാഥമിക ആരോഗ്യ കേന്ദ്രം പുഴുതി</u> <u>കോട്ടി പ. കണ്ണൂർ - 670005</u> |

STOCK REGISTER OF

Name of Article..... B.P. APPARATUS

| Date | No. and date of voucher or invoice | From whom received or to whom issued | Receipt | Issued | Balance after each transaction | Initials of Receiver | Remarks |
|----------|------------------------------------|--|---------|--------|--------------------------------|----------------------|--|
| | | Brought forward | | | | | |
| 16/9/19 | | Rd from pharmacy | 2 | | 2 | | pro d d |
| 16/9/19 | | Issued to op | | 2 | | | pharmacy equipment to set B |
| 16/11/20 | | Rd from pharmacy | 1 | | 3 | | " |
| 17/11/20 | | Issued to op | | 1 | | | " |
| 9/12/22 | | Rd from pharmacy | 1 | | 4 | | " |
| 9/12/22 | | Issued to op | | 1 | | | " |
| 15/1/24 | | Rd from pharmacy BP monitor Hg.p. 1320 OMRON (anurim) (1200) detector kit 24. Issued to op | 1 | | 5 | | " |
| 15/1/24 | | Rd from pharmacy LEC/LCO syringe Kerele surgical equip (mut) Issued to Medic Chem. Issued to Main center | 2 | | 7 | | page No 89 equipment register vol 1 10/12/23 |
| 11/3/24 | | Rd from pharmacy (surgical surgical) (CIRCA) 120150 | 1 | | 8 | | op equip (121) |
| 14/3/23 | | Issued to op | | 1 | | | (2500/-) |

Carried over



[Signature]
Pharmacist


CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800 - 425-7669

BAR CODE - (8004-890615225)
1343730



KMSC



Praveen
Chy



SYS
mmHg

DIA
mmHg

PUL
/MIN



WHO

MEM

START
STOP

SET