



PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: THQH KOYILANDY	Name of District: KOZHIKODE
Name of Equipment: X RAY MACHINE	Equipment ID & Barcode: 136281 & 1121140
Make: ALLENGERS MEDICAL SYSTEMS LIMITED	Model: ALLENGERS 325 FC/HT
Serial Number: 2K101130749 -X	Warranty details: NO Warranty
Date of purchase/Year of manufacture/ Installation Date: 21/01/2011	Present status of the equipment: Fully damaged
Date of breakdown: 22/09/2024	*AMC/ CAMC Period agreed
Date of registration of complaint through email/ Toll free): Toll free	at the time of purchase: NO AMC/CAMC
Action taken: Checked and found Xray not generated, found Cable H.V 6 MTR, Card Mas 325 MA , Card KVP FOR 325 MA, Tank Double 325\525 MA Complete (Assembled) and X-RAY Tube defective. Enquired spares with oem	
Recommendations for repair (required service details): Not recommending for repair	
Cost of spares (specify parts and cost): <ol style="list-style-type: none"> 1) Cable H.V 6 MTR -1 nos = 22400 /- 2) Card Mas 325 MA – 1 nos = 8920 /- 3) Card KVP FOR 325 MA – 1nos = 8960 /- 4) Tank Double 325\525 MA Complete (Assembled) – 1nos = 179200 /- 5) X-RAY Tube – 1nos = 179200 /- GRAND TOTAL = 398640 /-	
# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 108.31%	Asset Value: 368040/-
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix service report attached and vendor quotation attached	
Reasons for recommending the equipment as BER: Checked and found Xray not generated, found Cable H.V 6 MTR , Card Mas 325 MA ,Card KVP FOR 325 MA, Tank Double 325\525 MA Complete (Assembled) and X-RAY Tube defective. Equipment installed on 21/01/2011 (stock book attached for reference). Equipment out lived 13 years 8 months. Spare cost is 108.31% with respect to asset value, both criteria for RBER met, As per tender clause 5.3.14.1 we recommending the equipment for condemnation.	
JINEESH PP 9/11/2024 Name & Signature of CYRIX Authority with date	
Remarks of Junior Consultant (Biomedical) NHM:	
Checked & found that x-Ray machine not working. Cable HV6 make Card Mas 325M Card KVP for 325MA, Tank Double 325/525 MA, and X-Ray tube damaged. Equipment installed on 21/01/2011 So recommended for condemnation	
Recommended for BER (Yes/ No): Yes	Consultant Bio Medical Engineer
Date: 10/10/24	National Health Mission P.R. Kozhikode Signature of JC BM (NHM)
	
Date: 14/10/24	 Signature of Superintendent / Medical Officer (i/c)

*Not mandatory

Based on the period of life and value as per the BER guideline

SUPERINTENDENT
 Taluk Head Quarters Hospital
 Koyilandy
 Attach Photograph



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No: WO-37/2021-2022/698

No : 1013720

CYRIX[®]
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>THQH</u>	Call Registration Date : <u>22-09-2024</u>
Address..... <u>Koyilandy</u>	Caller ID : <u>136281</u>
..... <u>Badussery</u>	Date of Visit : <u>23-09-2024</u>
Ph : <u>9495669140</u>	Asset No : <u>1121140</u>
	EQPT Name : <u>X-Ray Maehiao</u>
	Manufacture <u>Allengers</u> Model : <u>325EC/HT</u>
	S. No <u>2K101130749</u> Dept. <u>X-ray</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Radiation (x-ray) not generated

Action Taken : Checked and found that cable HV6 meter Card mA's 325, Card KVP 325, Tank Drable 325mA and x-ray tube defective. Enquired spares with OEM

Completed Date : 23/9/24 Time : 2:10pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Midhan rajk</u>	<u>23/9/24</u>	<u>11:00am</u>	<u>2:10pm</u>

Customer Remark Completed Pending

Service Engineer Name : Midhan rajk Customer Name : Dina karan
 Signature: [Signature] Signature: [Signature]
 Date: 23/9/24 Date: 23.09.24
 Contact Number: 9072522081 Contact Number: 9495669140
 Designation : Radiographer



SUPERINTENDENT
rajuk Head Quarters Hospital
Koyilandy

To,
Cyrix Healthcare (P) Ltd
30/641B, Vandipetta Rd, near SBI Bank,,
Poonithura, Maradu, Ernakulam,
Kerala-682038

Date: 07/10/2024

Sub: Estimate for Spares of Allengers-325 (THQH-Kovilandy)

Dear Sir,

As per the above mentioned subject please find the below estimate.
Kindly go through the same and we are expecting your work order at the earliest.

SI No.	Item description	QTY	Price	Taxes	Amount in Rs.
1	CABLE H.V 6 MTR W/O SILICON WASHER (11008)	1	20000	12%	22400.00
3	CARD mAS 325/525/625 MA	1	8000	12%	8920.00
4	CARD KVP FOR 325/525/625 MA (SA2958-JW-2)	1	8000	12%	8960.00
5	TANK DOUBLE 325\525 MA COMPLETE [ASSEMBLED] (5554)	1	160000	12%	179200.00
6	X-RAY TUBE E-7239-X (2169)	1	160000	12%	179200.00
Grand Total					398640.00
(Rupees three lakh ninety eight thousand six hundred forty only)					

*Above information is not an invoice and only an estimate of services described above.If prices and tax change or additional parts, and labor are required, we will inform you prior to proceeding with the work. Kindly confirm your billing address (as mentioned above) and intimate us if any changes required. This is not included our service charges.

Kindly give an approval for the service in written (Work order) to following address Email id, so that we can perform the Job accordingly.

Payment: 100% in advance.

Our GSTIN number: 03AABCA2332H1ZE

For Allengers Medical Systems Ltd



Authorized Signatory
RENJITH RAMESH C Executive, Office
Operations Cochin Branch
Allengers Medical Systems Ltd
5-B, 5th Floor, Hedge House (Pearl Opulence),
Edapally-Palarivattom Road, Ernakulam - 682 025
Tel (484) 4035626 mobile +91 7087029374

Fwd: Quotation- THQH Koylandy

From: Aneesh P B mklse@cyrix.in
To: Blesson Jose zm2.klbemp@cyrix.in
Sent: Monday, 7 October at 10:49 am

Dear Blesson,

PFA quote from Allengers.

Kind Regards,

Aneesh P B

Deputy Manager - Service

Ph: 9744827770

**Cyrix Healthcare Pvt Ltd | 30/641 B |
Pettah Jn | Poonithura | Cochin -38 |
Kerala | India**

Ph: 9744555073 | M 9072522093 |

W www.cyrixhealthcare.com

KERALA || KARNATAKA || UTHAR PRADESH ||

TAMIL NADU || KUWAIT || KSA ||UAE

Begin forwarded message:

From: RENJITH

[<ker.service@allengers.net>](mailto:ker.service@allengers.net)

Date: 7 October 2024 at 10:45:41 AM

GMT+5:30

To: Aneesh P B [<mklse@cyrix.in>](mailto:mklse@cyrix.in)

Subject: RE: Quotation- THQH Koylandy

Dear Aneesh,

Please find the attached estimate.

Thanks and regards

RENJITH RAMESH C Executive, Office Operations | Cochin

Branch

Allengers Medical Systems Ltd

5-B, 5th Floor, Hedge House (Pearl Opulence), Edapally-

Palarivattom Road, Ernakulam - 682 025

8800

www.allengers.com |

ker.service@allengers.net

Confidentiality Note: This e-mail and any attachments are confidential and may be protected by legal privilege. If you are not the intended recipient, be aware that any disclosure, copying, distribution or use of this e-mail or any attachment is prohibited. If you have received this e-mail in error, Please notify us immediately by returning it to the sender and delete this copy from your system. Thank you for your cooperation.

P.Please consider the environment before printing this e-mail!

From: Aneesh P B [<mailto:mklse@cyrix.in>]

Sent: Friday, September 27, 2024 2:18 PM

To: ker.service@allengers.net

Subject: Re: Quotation- THQH Koylandy

Pls share estimate asap.

Kind Regards,

Aneesh P B

Deputy Manager - Service

Ph: 9744827770

**Cyrix Healthcare Pvt Ltd | 30/641 B |
Pettah Jn | Poonithura | Cochin -38 |
Kerala | India**

Ph: 9744555073 | **M** [9072522093](tel:9072522093) |

W www.cyrixhealthcare.com

KERALA || KARNATAKA || UTHAR

PRADESH || TAMIL NADU ||

KUWAIT || KSA || UAE

On 24 Sep 2024, at 1:17 PM, Aneesh P B
<mklse@cyrix.in> wrote:

Dear Team,

Please share quote for below spares
to THQH Koylandy.

Equipment: Allengers 325

SN: 2K101130749-X

Parts needed:

1. HT Transformer
2. X Ray Tube
3. HV Cable -1 pair
- 4: MAS Card.

Kind Regards,

Aneesh P B

Deputy Manager - Service

Ph: 9744827770

Cyrix Healthcare Pvt Ltd | 30/641

B | Pettah Jn | Poonithura | Cochin

-38 | Kerala | India

Ph: 9744555073 | M 9072522093 |

W www.cyrixhealthcare.com

KERALA || KARNATAKA || UTHAR

PRADESH || TAMIL NADU ||

KUWAIT || KSA || UAE



Allengers

OPERATING CONDITIONS


LINE VOLTAGE	380 - 440V
FREQUENCY	50 / 60 Hz
MAX. LINE RESISTANCE	0.2 Ω
OPERATING TEMP	10 - 40°
OPERATING HUMIDITY	80%

MODEL	Allengers 325 FC/HT
SR. NO.	2K101130749-X
AERB NO.	12/400
POWER	440V, ± 10 % 50/60 HZ.

Manufactured by: -
 ALLENGERS MEDICAL SYSTEMS LIMITED
 WORKS: 101 INDIA STREET, PHASE - 01 BADOI (DISTT. SOLAN) H.P.
 INDIA. TEL: 0191-240-215 FAX: 0191-240-215
 CHANDIGARH (INDIA)
 TEL: 0172-270-1000 FAX: 0172-270-1000
 WWW.ALENGERS.COM

SUPER MEDICAL
 Head Office
 Koyilandy

KOYILANDY



Taluk Head & ...
 Koyilandy

FOR SERVICE COMPLAINTS PLEASE CONTACT

SOUTH ZONE

ANCH			
CHENNAI	: 044	- 32966498	MOB.: 093800 18566
MANGALORE	: 080	- 3293480	MOB.: 093425 18152
MUMBAI	: 040	- 32965077	MOB.: 093467 99791
MANGALORE	:		MOB.: 098452 85808
CHENNAI	:		MOB.: 093413 62458
MUMBAI	:		MOB.: 098460 83167
MUMBAI	: 0495	- 2766212	
HEAD OFFICE (SOS)	0172-3258435		MOB.: 093180 19218
CUSTOMER CARE	0172-3258435		

CYRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)
TOLL FREE NO: 1800 - 425-7669
 S.C.O. 21 PH BAR CODE - (8004-890615225) 1121140
 E-mail: ...

X-RAY, C-ARM, FLUORO, MAMMOGRAPHY, MOBILE CATH-LAB & DSA SYSTEMS
 TALUK HEAD QUARTERS HOSPITAL
 Koyilandy

X-ray Machine :

From substock register
vol-I page No: 55

18.8.72 150 mA machine.
Condemned as per order No 2245/2004
dt. 9/11/2005

21.01.11 300 mA X-ray machine
C table.
(Allengers 325/FC/Horizontal)
(KMCL-G page No: 28)

1. Issued
for use!

AERB Reg: From 23/07/2017
to 23/07/2022

AERB Reg: From - 12-10-2022
To : 12-10-2027


SUPERINTENDENT
Taluk Head Quarters Hospital
Kovilandy

