

**PROFORMA**  
**RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)**  
**BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: PHC ATHANIKKAL	Name of District: MALAPPURAM
Name of Equipment: UPS	Equipment ID & Barcode: 136526 #1040260
Make: APC	Model: Back -UPS 600
Serial Number: B21718009363	Warranty details: NO Warranty
Date of purchase/Year of manufacture/ Installation Date: 08-11-2017	Present status of the equipment: Fully damaged
Date of breakdown: 23-09-2024 Date of registration of complaint through email/ Toll free): Toll free	*AMC/CAMC Period agreed at the time of purchase: NO AMC/ CAMC

Action taken: Checked the machine & found Main board , Transformer & 12A 7Ah battery damaged. Need to replace these Spare for further checking

Recommendations for repair (required service details): No recommendations

Cost of spares (specify parts and cost):  
 12 V 7AH BATTERY:1536/-  
 12-230 inverter PCB :1298/-  
 Transformer:649 /-  
**TOTAL:3483/-**

# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 77%      Asset Value: 4500/-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Service report attached & OEM Quotation attached

Reasons for recommending the equipment as BER: Checked the machine and found 12-230 inverter PCB ,Transformer & 12A 7Ah Battery defective .The equipment was installed on 08/11/2017 aged up to 6 years and 10 months .Repairing cost is 77% of asset value, both criteria for RBER met As per the tender clause 5.3.14.1 we recommending the equipment for condemnation .

*[Signature]* 08/10/2024  
 GOKUL KT  
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:  
*Verified with the damage of battery, transformer and inverter PCB. Need to replace them parts. Beyond economic repair.*

Recommended for BER (Yes/ No): YES

Date: 08/10/24


*[Signature]* 08/10/24  
 Jr. Consultant Biomedical Engineer  
 NHM Malappuram

Date: 14/10/2024      Seal      Signature of Superintendent / Medical Officer (i/c)

\*Not mandatory

# Based on the period of life and value as per the BER guidelines

\* Attach Photograph



*[Signature]*  
 Medical Officer  
 Family Health Centre  
 Athanikkal



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



No : 1012156

**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

# CYRIX®

HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

### Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility: <u>PHC ATHANIKKAL</u>	Call Registration Date: <u>23/09/2024</u>
Address: <u>MALAPPURAM</u>	Caller ID: <u>136526</u>
Ph: <u>9497710362</u>	Date of Visit: <u>25/09/2024</u>
	Asset No: <u>1040260</u>
	EQPT Name: <u>UPS</u>
	Manufacture: <u>APC</u> Model: <u>UPS 600</u>
	S. No: <u>B21718009363</u> Dept: <u>Laboratory</u>

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : Not switching on.

Action Taken : checked the machine and found main board transformer, 12V 7Ah battery damaged. need to replace those spares for further checking.

Completed  Date : 27/09/2024 Time : 11AM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1. 12V 7Ah battery	1 NOS		
2. 12-230 inverter PCB	1 NOS		
3. Transformer	1 NOS		

Cyrix Engineer Rohith V

Date	Start Time	End Time
<u>27/09/2024</u>	<u>10AM</u>	<u>11AM</u>

Customer Remark  Completed  Pending

Service Engineer Name : <u>Rohith V</u>	Customer Name : <u>MARIE M.V</u>
Signature: <u>[Signature]</u>	Signature: <u>[Signature]</u>
Date: <u>27/09/2024</u>	Date: <u>27/09/24</u>
Contact Number: <u>9539838931</u>	Contact Number: <u>9497710362</u>
	Designation : <u>lab technician</u>
	Hospital Seal

**AION BACKUPS**

YP/VIII/320 AYLARA, KERALA KOLLAM

Date: 04/10/2024

Contact: 8921950217

Email: [aionbackups12@gmail.com](mailto:aionbackups12@gmail.com)

GSTIN: 32CGUPM9450L1ZN

## APC 600VA UPS SERVICE PROPOSAL

TO,

CYRIX HEALTH CARE PVT .LTD

Barcode: 1040260 Ticket ID: 136526

Hospital : PHC, Athanikkal ,Malappuram

Dear Sir/Madam,

We are pleased to submit here with our quote for the same as per below.

SL	DESCRIPTION	QTY	UNIT RATE IN INR	TOTAL	GST	TOTAL PRICE IN INR
1	12-230 INVERTER PCB	1	1100.00	1100.00	18	1298.00
2	7Ah BATTERY	1	1200.00	1200.00	28	1536.00
3	TRANSFORMER	1	550.00	550.00	18	649.00
<b>PAYABLE AMOUNT</b>						<b>3483.00</b>

### Commercial terms and conditions

1. Payment 80% advance along with PO balance before dispatch, 20% after installation.
2. Delivery period; 2weeks.
3. Tax :GST@18% for spares,28%@Battery
4. Installation; Our scope.

Please arrange to release the PO at our above address

Bank name :HDFC

Account no:50200082743176

IFSC Code:HDFC0004083

We hope you will find the above offer in line with our discussions and await your valuable order please free to call or mail us on above given number if any query.

Thank you, Sincerely,  
BHAGATH MOHAN





---

## 1040260 UPS MACHINE MAIN BOARD & BATTEEY SPARE QUOTATION

3 messages

---

**Gokul K T** <gokulcyrix@gmail.com>

Mon, 30 Sep, 2024 at 1:16 pm

To: aionbackups12@gmail.com

Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear Sir, Madam

Can you send the Price quotation for UPS main board & Battery Make :APC, model :Back -UPS 600

Barcorde:1040260

Ticket id:136526

UPS

PHC ATHANIKKAL

GOKUL KT

CYRIX HEALTHCARE PVt Ltd

DISTRICT INCHARGE

MALAPPURAM

9809314280

---

**Gokul K T** <gokulcyrix@gmail.com>

Thu, 3 Oct, 2024 at 9:41 am

To: aionbackups12@gmail.com

Cc: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Reminder

[Quoted text hidden]

---

**Bhagath Mohan** <aionbackups12@gmail.com>

Fri, 4 Oct, 2024 at 9:12 am

To: Gokul K T <gokulcyrix@gmail.com>

**Thanks & Regards**

**Bhagath Mohan**

**PH:-8921950217**

[Quoted text hidden]



*Handwritten signature*


Medical Officer  
Family Health Centre  
Athanikkal



# UPS- MISPA VIVA

7

8.11.2017	Received from store	1	1	558
	Sl. No: B2009363			
8.11.17	Issued to Lab use		1	Ni B
	(Sl. No: B21718009363)			



*[Signature]*  
 Medical Officer  
 Family Health Centre  
 Athanikkal