

**PROFORMA  
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: PHC PADIYUR	Name of District: THRISSUR
Name of Equipment: Sterlizer	Equipment ID & Barcode: 130663 & 0844173
Make: NA	Model: NA
Serial Number: NA	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: Ticket master installation date-NA, As per the Stock book installation date- 31-12-2018	Present status of the equipment: Fully damaged
Date of breakdown: 21-08-2024	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Date of registration of complaint through email/Toll free): Toll free	
Action taken: Machine checked and found machine fully rusted condition and MNX 70, inner chamber, outer chamber, gasket, coil, pressure gauge are defective. Need to replace these spares for further checking working condition of equipment.	


Recommendations for repair (required service details): Not recommending


Cost of spares (specify parts and cost):	Inner chamber Outer chamber Gasket Pressure gauge MNX 70 Coil	- 47000/-
--	--	-----------

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 1220.7%	Asset Value: Ticket master asset value - 3850/-
--	---

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): OEM quotation and Cyrix service report attached


Reasons for recommending the equipment as BER: Machine checked and found machine fully rusted condition and inner chamber, outer chamber, gasket, coil, pressure gauge are defective. Machine installed on 31-12-2018 aged up to 5 years and 8 months. The spare cost is more than 100 % of asset value and Both criteria met for RBER. As per tender clause 5.3.14.1 we are recommending this equipment for RBER.


  
 BIJOI JOY  
 Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:  
*It is autoclave and the machine is fully rusted.*  


Recommended for BER (Yes/ No):	 Signature of JC BM (NHM)
Date:	

Date: 14.10.2024



  
 Signature of Superintendent / Medical Officer (i/c)  
 Padiyoor PHC  
 Thriassur Dt. - 680 666

\*Not mandatory #Based on the period of life and value as per the BER manual. Attach Photograph



**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

No : 1007899

**CYRIX**<sup>®</sup>  
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala  
Ph : 98472 99500 Website :www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <b>RHC</b>	Call Registration Date : <b>21/08/2024</b>
Address..... <b>PADIYUR</b>	Caller ID : <b>130663</b>
<b>THRISSUR</b>	Date of Visit : <b>21/08/2024</b>
Ph : <b>9895903111</b>	Asset No : <b>0844173</b>
	EQPT Name : <b>STERILIZER</b>
	Manufacture : <b>NA</b> Model : <b>NA</b>
	S. No. : <b>NA</b> Dept. : <b>DRESSING ROOM</b>

Service Classification : Breakdown call  PMS  Calibration  Cust. Training

Problem Identified : **MACHINE DAMAGED FULLY**

Action Taken : **MACHINE CHECKED & FOUND MACHINE FULLY RUSTED. CONDITION & MIXX TO, INNER CHAMBER, OUTER CHAMBER, CASSET COIL, PRESSURE GAUGE ARE DEFECTIVE. NEED TO REPLACE THESE SPARES FOR FURTHER CHECKING WORKING CONDITION OF EQUIPMENT**

Completed  Date **21/08/24** Time : **11:00AM** Spare Required


Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<b>AMBADY.VS</b>	<b>21/8/24</b>	<b>10:00AM</b>	<b>11:00AM</b>

Customer Remark Completed  Pending

Service Engineer Name : **AMBADY.VS**  
 Signature: **[Signature]**  
 Date: **21/8/24**  
 Contact Number: **7907996506**



Customer Name :  
 Signature: **[Signature]**  
 Date:  
 Contact Number:  
 Designation : **Medical Officer in Charge**  
 Hospital Seal: **21-8-24**



PALAKKAL ANGADI, THRISSUR - 680 001, KERALA  
 0487 - 2427038, 2427031, Mob : 93870 211122, 8943341569  
 Email : uniage@gmail.com, uniage.sales@gmail.com

**CYRIX HEALTHCARE  
 ERNAKULAM**

**Our Quotation No: UA-Q.1575/2024**

**Dt.09.10.2024**

**Ref: Service Quotation - Vertical Autoclave**

Sir,

We hereby quote our lowest possible rates for the following items for your kind consideration and early favorable orders.

No.	Description	Unit Rate (Rs.)	GST %	Total Rate Inclusive of GST (Rs.)
1	VERTICAL AUTOCLAVE [MNX70 Inner chamber, Outer Chamber Pressure Gauge, Coil, Gasket ]	47,000.00	18	55,460.00

**Terms and Conditions**

Rates are inclusive of GST @ 18%  
 (Note: Civil / Electrical / Water Inlet & Outlet works of any nature will be of your scope)

Rates will remain firm for acceptance till 31.12.2024

GSTIN: 32AAAFU5662Q1ZQ

Account Name: M/s. Universal Agencies  
**Our bank details:**  
 STATE BANK OF INDIA  
 SME BRANCH, PARAMAKKAVU THRISSUR  
 A/c No.00000067213868605  
 IFSCODE SBIN0063880

**CONTACT DETAILS:**

Purchase/Inquiry: 8943112277 / 9387021122  
 Order status/Sales/Service:  
 8943341568/8943341569/0487-2427031  
 Accounts/Payment: 843112288 /0487-  
 2427031  
 Others: Mr. Prasannan ( Manager)  
 9447137031

Thanking You,

Yours Faithfully,  
 For Universal Agencies,

*[Signature]*  
 Manager.





Quotation enquiry- 0844173-130663 - phc padiyur

2 messages

Thu, 17 Oct, 2024 at 6:4

Bijo T Joy <bijocyrix@gmail.com>

To: uniage@gmail.com

Cc: DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Dear sir,

There is a complaint reported from PHC PADIYUR, Thrissur, Kerala, Regarding the complaint with Autoclave machine ,Please share the price and availability of the **MNX70, inner chamber, outer chamber, pressure gauge, coil and gasket** as soon as possible.

Equipment - sterilizer

Make - NA

Model: NA

Serial No: NA

Hospital- phc padiyur

District - Thrissur

Barcode - 0844173

Caller id - 130663

With Regards,

**BIJO T JOY**

DISTRICT INCHARGE THRISSUR,

**CYRIX HEALTHCARE PRIVATE LIMITED**

30/641 B, Petta Junction | Poonithura | Kochi | Kerala-682038 Branches: Kerala | Karnataka || Tamil Nadu || Uttar Pradesh | Kuwait | Saudi Arabia

M:-8921870010 | E:- bijocyrix@gmail.com | W:- www.cyrixhealthcare.com | "

Universal Agencies <uniage@gmail.com>

To: Bijo T Joy <bijocyrix@gmail.com>

Thu, 17 Oct, 2024 at 11:2

[Quoted text hidden]

-

Warm Regards,

M/S. Universal Agencies

(ISO 9001 : 2008 Certified)

H.O - "Palakkal Angadi, Thrissur -680001 Kerala - India"

Ph No: 0487-2440319/2427031

Cel : 8943341568

GST : 32AAAFU5662Q1ZQ

Email : uniage@gmail.com

uniage.sales@gmail.com

Branch Office: " 2nd Floor, Pride of Aristo,

C.S.I Cemetery Road, Mission Quarters



East Fort, Thrissur- 680005"

Tel: 0487-2427038/8943112277

Email: uniage.etender@gmail.com



**VERTICAL AUTOCLAVE**  
**(STEAM JACKETED)**

 **PRESSURE SWITCH**   
Indefea  
MODEL : PSM - 550 -- B2 11 C D  
RANGE : 0.2...3 bar  
DIFF. : 0.12...0.75 bar  
S. NO. : 20075    MFG. : NOV/2016  
Switzer Process Instruments Pvt. Ltd.  
Ghaatolad (www.switzerprocess.com)

**CYRIX** MEDICAL EQUIPMENT MAINTENANCE  
VEEDAKKAL STREET  
TOLL FREE NO.: 1800 - 425-7669  
BAR CODE : (8004-890615225)  
0844173

FAMILY HEALTH CENTRE  
THRISSUR  
Dr. \_\_\_\_\_  
889 0800R - 680Y00R

