

**PROFORMA  
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: CHC KARIVELLUR	Name of District: KANNUR
Name of Equipment: REFRIGERATOR	Equipment ID & Barcode: 134894-1330010
Make: GODREJ	Model: EDGE
Serial Number: N/A	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: 02-11-2011	Present status of the equipment: Fully damaged
Date of breakdown: 13.09.2024 Date of registration of complaint through email/Toll free): Toll free	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Action taken: Checked the Equipment Found Equipment is not Switching ON. Found That Compressor, Thermostat, Filter, Condenser Defective Enquired For quotation.	

Recommendations for repair (required service details): No Recommendations

Cost of spares (specify parts and cost): Compressor- 8,000/-  
Thermostat- 1,800/-  
Filter-600/-  
GAS Charging-2,800/-  
Condenser-4,000 /-  
**Total-17,800/-**

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 67.18%      Asset Value: 26,494/-

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service Report & Quotation attached.

Reasons for recommending the equipment as BER: Checked the Equipment Found Equipment Is not Switching ON. Found That Compressor, Thermostat, Filter, Condenser Defective Enquired For quotation. Equipment Installed On 02.11.2011, aged up to 12 years 10 months, The Repairing cost is 67.18% with respect to asset value. Both criteria for RBER met. As per Tender Clause 5.3.14.1 we recommend the equipment for condemnation.

Shobin vs  26/09/24  
Name & Signature of CYRIX Authority with date


Remarks of Junior Consultant (Biomedical) NHM: checked & identified that machine not switching on due to compressor, thermostat, filter, condenser are defective. Aged up to 12 years & above. As per quotation the service cost economically not feasible. So recommend for BER

Recommended for BER (Yes/ No):

Date: 26/09/24.

 26/09/24  
Signature of JC BM (NHM)

Nirisha. k. P

  
Nursing officer Gmi


Date: 28/9/24

Seal

Signature of Superintendent / Medical Officer (i/c)

\*Not mandatory

#Based on the period of life and value as per the BER guidelines \* Attach Photograph













---

## One time service required for Refrigerator 1330010#134894

2 messages

---

avinash t <avinashtcyrix@gmail.com>

Thu, 26 Sept, 2024 at 11:03

am

To: ecotonicsolutions@gmail.com

Dear sir,

Please provide service for Refrigerator at THQH PAYYANNUR

**Thanks & Regards Avinash**

**T**

**District charge, kannur**

**bemp kerala**

**mobile: +91 6364064948**

**E mail: [avinashtcyrix@gmail.com](mailto:avinashtcyrix@gmail.com)**

**CYRIX HEALTHCARE (P) LTD**

**30/641B | Petta | Punithura | Kochi | Kerala- 682038**

**[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)**

---

ECOTONIC AUTOMATION SOLUTIONS <ecotonicsolutions@gmail.com>

Thu, 26 Sept, 2024 at 11:14

am To: avinash t <avinashtcyrix@gmail.com>

Hi Avinash

Your quotation for above mentioned machine is attached. PFA

Regards,

Ecotonic Automation

Solutions [www.ecotonics.in](http://www.ecotonics.in)

[Quoted text hidden]

## REFRIGERATOR

04/01/14	(MODEL 02L-205K75 ST) Serial No: 311SRSC 01694 Rs. 12,800	1	1
04/01/14	Issued to Lab	1	nil
29/01/14	Len REF 96 314 PMG4. Rs. 30,335/-	1	1
31/01/14	Issued to plate pharmacy	1	nil
02/11/11	Rd from National Radio. Rs. 8400/-	1	1
02/11/11	Issued to ZP	1	nil
15/11/2022	Received from Bidlinga Dada Private Lab, approx inno: BIDL/22-23/127 dated 09/10/2022	1	1 (used project)
	Model - YC395L Make - Zhongke Meiping		Ant: 1,41,550/- warranty 3 yrs upto NIL 14/11/2022
15/11/2022	Issued to Pharmacy (Shree)	1	

*P. N. S. S.*  
 MEDICAL OFFICER  
 COMMUNITY HEALTH CENTRE  
 KANVELLUR-670 854



*Praveen*



**CYRIX™**  
HEALTHCARE PVT LTD

**BIOMEDICAL EQUIPMENT MAINTENANCE  
PROGRAMME (BEMP)**



**TOLL FREE NO: 1800 - 425-7669**

**BAR CODE - (8004-890615225)  
1330010**



*[Handwritten Signature]*  
**MEDICAL OFFICER  
COMMUNITY HEALTH CENTRE  
KARIVELLUR 670 521**

