

**PROFORMA
RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)**

Name of Hospital: UPHC PANNAKKAD	Name of District: MALAPPURAM
Name of Equipment: COLORIMETER	Equipment ID & Barcode: 139924 # 1042931
Make: LABTRONICS	Model: LT - 111
Serial Number: 2018-08-198	Warranty details: No Warranty
Date of purchase/Year of manufacture/ Installation Date: In ticket master installation date is 11/08/2018 As per stock book installation date is 03/12/2017	Present status of the equipment: Fully damaged

Date of breakdown: 10/10/2024	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CAMC
Date of registration of complaint through email/Toll free): Toll free	

Action taken: Checked the machine & found Main board defective. Need to replace these spares for further checking.

Recommendations for repair (required service details): No Recommendation

Cost of spares (specify parts and cost): MAIN BOARD: 4000/-
TOTAL : 4000/-

#Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value: 81%	Asset Value: 4900/-
--	---------------------

Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not): Cyrix Service report attached & vendor Quotation attached.

Reasons for recommending the equipment as BER: Checked the machine and found Main board defective. In ticket master installation date is 11/08/2018, As per stock book installed on 03/12/2017. Aged up to 6 years and 10 months, Repairing cost is 81% of asset value, both criteria for RBER met. As per the tender clause 5.3.14.1 We recommending the equipment the equipment for condemnation



16/10/24 GOKULK T
Name & Signature of CYRIX Authority with date

Remarks of Junior Consultant (Biomedical) NHM:
Verified with the complaint for Main board. Need to replace spare and is beyond economic repair

Recommended for BER (Yes/ No):
Date:

Date: 17/10/24
Seal
MEDICAL OFFICER
URBAN PRIMARY HEALTH CENTRE
MALAPPURAM
Signature of Superintendent / Medical Officer (i/c)
Dr. Jaseela

*Not mandatory
*Based on the period of life and value as per the BER guidelines
* Attach Photograph



Jr. Consultant Biomedical Engineer
NHM Malappuram
Signature of JC BM (NHM).



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

No : 1004579

CYRIX
HEALTH CARE PVT LTD

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi- 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | Email : bemp.kl@cyrix.in

Health Facility..... <u>UPHC</u>	Call Registration Date : <u>10/10/2024</u>
Address..... <u>PANNAKKAD</u>	Caller ID : <u>139924</u>
<u>MALAPPURAM</u>	Date of Visit : <u>11/10/2024</u>
Ph: <u>9847146678</u>	Asset No : <u>1042931</u>
	EQPT Name : <u>Colorimeter</u>
	Manufacture: <u>Labtronics</u> Model : <u>LT-111</u>
	S. No. <u>2018-08-198</u> Dept. <u>LAB.</u>

Service Classification : Breakdown call PMS Calibration Cust. Training

Problem Identified : Result Variation

Action Taken : checked the machine and found main board defective. need to replace these spares for further checking.

Completed Date : 11/10/24 Time : 12:45pm Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>CROKIL K T</u>	<u>11/10/24</u>	<u>10:45 AM</u>	<u>12:45 PM</u>

Customer Remark Completed Pending

Service Engineer Name : CROKIL K T Customer Name : Mamga-T
 Signature: [Signature] Signature: [Signature]
 Date: 11/10/24 Date: 11/10/2024
 Contact Number: 9809314280 Contact Number: 9847146678
 Designation : Lab tech
 Hospital Seal





We Care

Biomedical Solutions and Surgicals

wecarebss@gmail.com
Perinthalmanna, Malappuram
+91 9037614233

No. 102

Quotation

Date 15/10/2024.....

To Cyrix Health Care Pvt Ltd Ernamkulam

Sl.No	Particulars (Description & Specifications)	Qty	Rate	Amount
1.	MAIN BOARD	1	4,000	4,000
			Total	4,000
Rupees				Grand Total 4,000
We Care Biomedical Solution And Surgicals				

Prepared By



Received By



042931 Need price quotation for Colorimetre main board

messages

okul K T <gokulcyrix@gmail.com>

o: wecarebss876@gmail.com <wecarebss876@gmail.com>

c: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Tue, 15 Oct, 2024 at 11:3

Dear Sir, Madam

Can you send the Price quotation for Colorimeter main board , Make:Labtronics , model :LT -111

Barcode:1042931

Ticket id:139924

Colorimeter

UPHC PANNAKAD

GOKUL KT

CYRIX HEALTHCARE PVt Ltd

DISTRICT INCHARGE

9809314280

wecarebss <wecarebss876@gmail.com>

o: Gokul K T <gokulcyrix@gmail.com>

c: Blesson Jose <Zm2.klbemp@cyrix.in>, DM Sarang-KLBEMP <dm1.klbemp@cyrix.in>

Tue, 15 Oct, 2024 at 11:4


Hi gokul,

Your quotaion for the above mentioned spare attached below

[Quoted text hidden]



for Dr. Nushina


 Dr. Nushina



MEDICAL OFFICER
 URBAN PRIMARY HEALTH CENTRE
 MALAPPURAM