



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. :
229461

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>P.H.C Parathodu</u>	Call Registration Date : <u>01-08-2024</u>
Address <u>Kottayam</u>	Caller ID : <u>126790</u>
Ph : <u>9745265086</u>	Date of Visit : <u>03-08-2024</u>
	Asset No. : <u>0542979</u>
	EQPT Name : <u>Fetal doppler</u>
	Manufacture <u>Nidek</u> Model : <u>230</u>
	S. No. <u>F220C01800</u> Dept. <u>Sub center</u> <u>3131</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified :
Not working

Action Taken :
checked and found that problem with internal PCB and probe. Enquire the spare with O.E.M

Completed Date : 3/8/24 Time : 10:00 Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			



Cyrix Engineer	Date	Start Time	End Time
<u>Ajmal Shajahan</u>	<u>3/8/24</u>	<u>10:00</u>	<u>12:00</u>

Customer Remark Completed Pending

Date.....

Service Engineer Name : Ajmal Shajahan
Signature : [Signature]
Date : 03/8/24
Contact Number : 9491275618

Customer Name : Ajmal
Signature : [Signature]
Date : 03/8/24
Contact Number : 9745265086
Designation : JPHN
Hospital Seal :





**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	PHC PARATHODU
3	Name of Equipment with Make, Model and Serial Number	Equipment : Fetal doppler Make : NIDEK Model :230 SN:F220C018003131
4	Equipment ID & Barcode	126790 & 0542979 ✓
5	Date of purchase / Year of manufacture /Installation Date	12/03/2019 18/3/2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	1/08/2024 (TOLL FREE)
9	Action taken	Checked and found out the problem with main pcb and probe . Enquired the spare with OEM
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	Probe -3757/- Main pcb- 1352/- Total-5109/-


13	Asset Value	Rs . 7500/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	68%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report and Quotation are attached
16	Reasons for recommending the equipment as BER	Checked and found out the problem with main pcb and probe . Machine installed in 12/03/2019 and covered up to 5 +.The repairing cost 68% both criteria met. So recommending the machine for condemnation as per the tender clause 5.3.14.1 12/3/2019 is invoice date
17	Name & Signature of CYRIX Authority	Ajmal shajahan  

*Not mandatory #Based on the period of life and value as per the BER guidelines


* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value Rs = 7500
 cost of spare - 68% (5104)
 Life period - 5 year 5 months
 As per the RBER guide line recommended
 as RBER


 2/9/20
 Signature of JC BM (NHM)
 CONSULTANT BIOMEDICAL ENGINEER
 KOTTAYAM

Date _____ Seal _____

FOR DR. VANMUKRISHNAM
 DR. TAHA 
 Signature of
 Superintendent / Medical Officer (i/c)
 The Medical Officer in charge
 F. H. C. Parathodu
 Kottayam District



STOCK REGISTER OF

Name of Article

Duplex - ⁹² fatal

Date	No. and date of voucher or invoice	From whom received or to whom issued	Receipt	Issued	Balance after cash transaction	Initials of Receiver	Remarks
------	------------------------------------	--------------------------------------	---------	--------	--------------------------------	----------------------	---------

18/3/19

Received from LHI Parakkal
 No: A 489 dtd: 12.03.2019 HSN-90181190 RS-3500/-
 Received from general health care (Association)

Carried over

Chitra

10,000x100/25/539/2007/SMT/G Govt. of Kerala



For DR. JAYAN PESHAWAN
 The Medical Officer
 F. H. C. Parakkal
 Kottayam District

Jingey



9744473043, 9744473243

cloud99biotech@gmail.com

1 KMC/XIII-1333
Room No: 22,
CSI Golden Jubilee Complex
Baker Hill
Shastri Road
Kottayam

CLOUD99 BIOTECH

GSTIN: 32AMWPT5633Q1Z0
State: 32 - Kerala

MSME NO.: UDYAM-KL-07-0030714
DL NO.: KL/KTM/MD42/2023/000015

Estimate

Estimate For:

ARUN G C

1ST FLOOR 30/641B CYRIX HEALTHCARE PVT LTD
KOCHI DHANUSHKODI ROAD POONITHURA

Contact No.: 8921992272

GSTIN Number: 32AAFCC2499H2ZM

State: 32-Kerala

Estimate No.: C998-QTN-48

Date: 19/08/2024

Place of Supply: 32-Kerala

#	Item name	HSN/ SAC	Quantity	Unit	Price/ Unit	GST	Amount
1	PROBE FOR FETAL DOPPLER (PROBE FOR FETAL DOPPLER)	9018	1	Pcs	₹ 3,183.90	₹ 573.10 (18%)	₹ 3,757.00
2	BOARD FOR FETAL DOPPLER (BOARD FOR FETAL DOPPLER)	9018	1	Pcs	₹ 1,207.14	₹ 144.86 (12%)	₹ 1,352.00
Total						₹ 717.96	₹ 5,109.00

Pay To:

Bank Name : UNION BANK OF INDIA, KOTTAYAM

Bank Account No. : 351201010035645

Bank IFSC code : UBIN0536121

Account holder's name : CLOUD99 BIOTECH

Sub Total	₹ 4,391.04
SGST@6%	₹ 72.43
CGST@6%	₹ 72.43
SGST@9%	₹ 286.55
CGST@9%	₹ 286.55
Total	₹ 5,109.00

Description

PHC PARATHODU
EQUIPMENT: FETAL DOPPLER
MAKE: NIDEK MEDICAL INDIA
MODEL: 230
SN: F220C018003131

Estimate Amount In Words

Five Thousand One Hundred Nine Rupees only


Terms And Conditions

Thanks for doing business with us!

For : CLOUD99 BIOTECH



Authorized Signatory



FETAL DOPPLER

Model: Model-230
SN: F220C018003131
Power: 2018-08
Power: 2x1.2V

CE 0123

NIDEK MEDICAL INDIA
238B, AJC Bose road Kolkata - 700020
www.nidekindia.com

BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

FREE NO: 1800 - 425-7669

CODE - (8004-890615225) 0542979

For
DR. UMMA KRISHNAN

DR. STANA

Medical Officer in charge
F. H. C. Parathodu
Kottayam

Signature

FAMILY HEALTH CENTER
Date.....
PARATHODU KOTTAYAM

(72469)