



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. : 232226

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>General Hospital</u></p> <p>Address <u>Neyyattinkara</u> <u>Thiruvananthapuram.</u></p> <p>Ph : <u>9495284102</u></p>	<p>Call Registration Date : <u>5-8-2024</u></p> <p>Caller ID : <u>127342</u></p> <p>Date of Visit : <u>7-8-2024</u></p> <p>Asset No. : <u>0113184</u></p> <p>EQPT Name : <u>Nebulizer</u></p> <p>Manufacture <u>Life Care</u> Model <u>CNB 6909</u></p> <p>S. No. <u>209191160</u> Dept. <u>OPD</u> <u>557</u></p>
---	--

Service Classification : Breakdown Call PMS Calibration Cust. Training

Problem Identified : Not working

Action Taken : Checked the unit and identified nebulizer motor defective. Need to replace these spares for further checking and working condition of the equipment.


Completed Date : 7-8-2024 Time : 11:30 a.m Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Ajith George</u>	<u>7-8-2024</u>	<u>11:00 a.m</u>	<u>11:30 a.m</u>

Customer Remark Completed Pending



Service Engineer Name : Ajith George
Signature : [Signature]
Date : 7-8-2024
Contact Number : 7593847111


Customer Name GENERAL HOSPITAL NEYYATTINKARA-68512
Signature : [Signature]
Date : 7/8/24
Contact Number : 9400911819
Designation : NO CH
Hospital Seal :

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GENERAL HOSPITAL NEYYATINKARA
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : NEBULISER Manufacturer : LIFECARE Model : CNB69009 Sl: NA
4	Equipment ID/ Barcode	127342/ 0113184
5	Date of purchase/ Year of manufacture/Installation Date	26/12/2018
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	05/08/2024 (Toll free)
9	Action taken	Checked the unit and identified nebulizer motor defective . Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not recommending for Repair
12	Cost of spares (specify parts and cost)	Not available

13	Asset Value	Rs . 1353/-
14	Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Na
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment RBER	Checked and found nebulizer motor defective. The unit was installed on 26/12/2018 and covered up to 5+ years. Quotation not submitted since spares are not available in the market. So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AJITH GEORGE 

*Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Motor defective . so can be RBER as per lender conditions .

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


23-08-2024
Signature of JC BM (NHM)

Date:





Signature of
Superintendent/Medical Officer (i/c)


SUPERINTENDENT
GENERAL HOSPITAL
NEYYATTINKARA-695121

Nebulizers


18401

Date	Received from Stock	Received	Issued	Balance
19/3/12	1			1
23/3/2015	Received from stock (Dated page no 27)	1		2
17/6/2015	1 from stock			3
5/6/2016	Received from stock	1		4
2/10/16	Received from stock	2		6
8/2/2017	Received from stock MSR page no. 28	2		8
16/10/17 (L.P)	Recd from stock 2 Nos. Consumed as per order no 4920/17 G.M. dated 30/10/2017	3		11
10/4/18	Recd from stock (L.P)	2	2	9
3/10/18	Recd from stock	2		11
26/12/18	Portable nebulizers Received from stock 1 No. OP Su / 5434/10 dated 26/12/18 District Health Family Welfare Society Mysore Order no - DPM Su - 744 / 3435 / Jc dated 2/11/18 (Stock Report Vol-5 pag 357)	1		12
29/5/19	Received from stock	2		14
	MS per order no 42811 dated 1/2/2019 Nebulizer 2 Nos. Consumed and received from General Hospital Neyyattinkara		2	12

SUPERINTENDENT
GENERAL HOSPITAL
NEYYATTINKARA-695124



12 No
max 16




SUPERINTENDENT
GENERAL HOSPITAL
NEYYATTINKARA-695121

