



**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



SERVICE PROVIDER  
Tender No. WO-37/2021-2022/698

**CYRIX**  
HEALTHCARE PVT LTD

No. :

234556

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>PHC Manimala</u>	Call Registration Date : <u>7/8/2024</u>
Address <u>Manimala</u>	Caller ID : <u>128082</u>
<u>Ko.Hayam</u>	Date of Visit : <u>8/8/2024</u>
Ph : <u>8075621156</u>	Asset No. : <u>0540869</u>
	EQPT Name : <u>Nebulizer</u>
	Manufacture <u>Life Line</u> Model : <u>Alpha</u>
	S. No. <u>N/A</u> Dept. <u>OP</u>

Service Classification : Breakdown Call  PMS  Calibration  Cust.Training

Problem Identified : Fumes not coming


Action Taken : Found the motor of compressor is suited and not working. Need to replace the motor of compressor in order to check the working condition of the equipment.

Completed  Date : 8/8/2024 Time : 2:05pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			
Cyrix Engineer		Date	Start Time
<u>Sreeja Suresh</u>		<u>8/8/2024</u>	<u>1:20pm</u>
			<u>2:05pm</u>

Customer Remark



Deepa P.N  
Assistant Surgen  
Primary Health Centre  
Manimala

Service Engineer Name : Sreeja Suresh  
Signature : [Signature]  
Date : 8/8/2024  
Contact Number : 7902854249


Customer Name : Deepa P.N  
Signature : [Signature]  
Date : 8/8/24  
Contact Number : 9744104762  
Designation : Senior Np Officer  
Hospital Seal :

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

PROFORMA

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	KOTTAYAM
2	Name of Hospital	PHC MANIMALA
3	Name of Equipment with Make, Model and Serial Number	Equipment :Nebulizer Make : Life line Model : Alpha SN: Nill
4	Equipment ID & Barcode	128082 & 0540869
5	Date of purchase / Year of manufacture /Installation Date	21/09/2017
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	07-08-2024 (Toll free)
9	Action taken	Checked and found out the problem with compressor and motor
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	N/A

13	Asset Value	Rs .2162 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found out the problem with compressor and motor. Machine installed in 21/09/2017 and covered up to 7+ years. Quotation not submitted since spare are not available in the market. So recommended the machine for condemnation
17	Name & Signature of CYRIX Authority	Sreeju Suresh 

\*Not mandatory **#Based on the period of life and value as per the BER guidelines**


\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Asset Value - 2162  
 Cost of spare - not submitted  
 mentioned as spare not available.  
 Life period 6 years 11 months  
 as per the RBER guideline  
 recommended as RBER

Signature of JC BM/MI/MS  
 NATIONAL HEALTH MISSION  
 KOTTAYAM

Date

  
 Seal

Assistant Surgeon  
 Primary Health Centre  
 Mammalu

  
 Signature of  
 Superintendent / Medical Officer (i/c)

Nebulizer (electric)

141

Date		Rd	Issued	Bal
21/9/17	To DR from S.S. Register vol II Page 43	1	-	1
22	Rs-2162.00/- (LifeLine Model N/A plus)			
22/10/2018	To DR from S.S. Register vol II Page-43 (element, portable) (CNRHM supply) A	1	-	2
16/11/2019	To DR from S.S. Register vol II Page-43 (LifeCare Rs-1232/-)	1	-	3
1/12/2022	To DR from S.S. Register vol II Page-43 Bio + Plus Rs 1776.7/- (DMO supply)	1	-	4

Up.

Up.



Asst. Surgeon  
Primary Health Centre  
Mindhala



**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)  
**TOLL FREE NO: 1800 - 425-7669**  
 BAR CODE - (8004-890615225) 0540869

(14)68(254)05(21)3853

Medical Equipment Maintenance Programme (BEMP)  
 1800-425-7669

PRIMARY HEALTH CENTRE  
 Date  
 MANIMALA

*Dgw*  
 Assistant Surgen  
 Primary Health Centre  
 Manumala



MES CALIBRATION CENTRE  
Medical Engineering & Services  
Thiruvananthapuram - 21, PO 048/240007  
MES  
Cal Date: 08-12-21  
Exp Date: 03-12-22  
ID No. 67053853

14468254105213853

CYRIX MEDICAL EQUIPMENT MAINTENANCE  
PROGRAMME (EMMP)  
TOLL FREE NO: 1-800-425-7600  
SERVING SINCE - 1988

MANIPAL HEALTH SERVICES  
Date \_\_\_\_\_  
MANIPAL

*Dps*  
ASSTANT DIRECTOR  
Primary Health Centre  
Manimala



**Life-Line®**

**PISTON COMPRESSOR NEBULIZER**

Model No. : Alpha

Input : 230V, 50Hz

Current : 0.8A

Mode of operation: 30 Min. ON - 30 Min. OFF

Manufactured by:

LifeLine Medical Devices, Gurgaon. INDIA



CE



*dp*  
Assistant Surgen  
Primary Health Center  
Manimala