



**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



234560

SERVICE PROVIDER  
Tender No. WO-37/2021-2022/698

**CYRIX**  
HEALTHCARE PVT LTD

No. :

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report** 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <u>PHC Manimala</u>	Call Registration Date : <u>7/8/2024</u>
Address <u>Manimala</u>	Caller ID : <u>128083</u>
<u>Kottayam</u>	Date of Visit : <u>8/8/2024</u>
Ph : <u>8075621156</u>	Asset No. : <u>0540870</u>
	EQPT Name : <u>Nebuliser</u>
	Manufacture <u>Life Care</u> Model : <u>CNB 69022</u>
	S. No. <u>22190401799</u> Dept. <u>OP</u>

Service Classification : Breakdown Call  PMS  Calibration  Cust.Training

Problem Identified : Fumes not coming.

Action Taken : Found the motor of compressor is damaged. Need to replace the motor of compressor & check the working condition of the equipment.


Completed  Date : 8/8/2024 Time : 2:30pm Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Sreeju Suresh</u>	<u>8/8/2024</u>	<u>2:05pm</u>	<u>2:30pm</u>

Customer Remark Completed  As Pending   
Primary Health Centre Manimala



Service Engineer Name : <u>Sreeju Suresh</u>	Customer Name : <u>Deppa P.N</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>8/8/2024</u>	Date : <u>8/8/2024</u>
Contact Number : <u>7902854749</u>	Contact Number : <u>9744104762</u>
	Designation : <u>Senior Nursing officer</u>
	Hospital Seal :


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

<b>Sl. No</b>	<b>Particulars</b>	<b>Details</b>
1	Name of District	KOTTAYAM
2	Name of Hospital	PHC MANIMALA
3	Name of Equipment with Make, Model and Serial Number	Equipment :Nebulizer Make : Life care Model : CNB69022 SN: Nill
4	Equipment ID & Barcode	128083 & 0540870
5	Date of purchase / Year of manufacture /Installation Date	06/11/2019
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	07-08-2024 (Toll free)
9	Action taken	Checked and found out the problem with compressor and motor
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	N/A




13	Asset Value	Rs .1232 /-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found out the problem with compressor and motor. Machine installed in 06/11/2019 and covered up to 4+ years. Quotation not submitted since spare are not available in the market. So recommended the machine for condemnation
17	Name & Signature of CYRIX Authority	Sreeju Suresh 

\*Not mandatory #Based on the period of life and value as per the BER guidelines


\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Asset value - 1232  
 Cost of spare - not submitted  
 Life period - 4 years 9 months  
 Cyrix reported that spare not available in the market.

  
 30/8/24  
 Signature of JC BM (NHM)

Date \_\_\_\_\_ Seal \_\_\_\_\_

  
 Signature of Superintendent / Medical Officer (i/c)

Assistant Surgeon  
 Primary Health Centre  
 Manimala



# Nebulizer (Electric)

141

Date		Rd	Issued	Bal
21/9/17	To. In from S.S. Register vol II Page 43 Rs-2162.00/- (Life line Model 11/12/17)	1	-	1
22/10/2018	To. In from S.S. Register vol II Page-43 (Calculation, portable) (CNRHM Supply) A	1	-	2
06/11/2019	To. In from S.S. Register vol II Page-43 (Life line Rs-1232/-)	1	-	3
1/12/2022	To. In from S.S. Register vol II Page-43 Bio T Plus Rs 1776.7/- (DMO Supply)	1	-	4



Assistant Surgeon  
 Primary Health Centre  
 Maunala





*Pgo*  
Assistant Supt  
Primary Health Centre  
Mummalu



ISO 9001:2008 Certified Co. Ph: 9744555073  
CALON 10/11/23

**Lifecare**  
Compressor Nebulizer  
Model No.: CNB69022  
Input : AC-220V/50HZ  
Power : 180VA  
Serial No.: LC219040171A  
Mode of operation-continuous  
No AP/APG IPXO



**TOLL FREE NO: 1800 - 425-7669**  
**BAR CODE - (8004-890615225)**  
**0540870**




*Bo*  
Assistant Surgeon  
Primary Health Centre  
Manimala

**CYRIX** BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)  
TOLL FREE NO: 1800 - 425-7660  
BAR CODE - (8004-890615225) 0540870

Compressor Nebuliser  
Model No. CM89022  
Input AC-220V/50HZ  
Power 180VA  
Serial No. 0540870  
Mode of operation-continuous  
No. of AFM 1FX0

1800-425-7660  
ID: 0540870  
PR: NEDWJZEL  
NI: FKX036 JNINA  
HEALTH REPORT  
**CYRIX**

PRIMARY HEALTH CENTRE  
Date \_\_\_\_\_  
MANIMALA

  
Assistant Surgeon  
Primary Health Centre  
Manimala