



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 232474

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report 30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility ... <u>General Hospital</u>	Call Registration Date : <u>9/8/2024</u>
Address <u>Thiruvananthapuram</u>	Caller ID : <u>128458</u>
Ph : <u>9567271899</u>	Date of Visit : <u>14/8/2024</u>
	Asset No. : <u>0110825</u>
	EQPT Name : <u>Suction Apparatus</u>
	Manufacture <u>Industrial</u> Model : <u>MSI-R902</u>
	S. No. <u>202032</u> Dept. <u>OT</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Not working

Action Taken : Checked and found that Suction motor, Suction Jar and Suction Tubings are damaged. Need to replace these spares for further checking and working condition of the equipment

Completed Date : 14/8/24 Time : 10:00AM Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <u>NA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Jeswin sibi</u>	<u>14/8/24</u>	<u>09:00AM</u>	<u>10:00AM</u>

Customer Remark Completed Pending



Service Engineer Name : Jeswin sibi
Signature : [Signature]
Date : 14/8/2024
Contact Number : 9633315512

Customer Name : Sarulatha
Signature : [Signature]
Date : 14/8/24
Contact Number : 9495895103
Designation : NHIVM
Hospital Seal : [Seal]


General Hospital Thiruvananthapuram

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GENERAL HOSPITAL THIRUVANANTHAPURAM
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : Suction Apparatus Make : Monarch Model : MSI RG02 Serial No : NA
4	Equipment ID/ Barcode	128458/0110825
5	Date of purchase/ Year of manufacture/Installation Date	04/03/2020
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	09/08/2024 (Toll Free)
9	Action taken	Checked and found suction motor and suction jars defective. Need to replace these spares for further checking and working condition of the equipment. Enquired spare from vendor.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	Suction Motor - 4650/- Suction Jar - 3600/- Total = 8250/-

13	Asset Value	9900/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	83.3%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT AND VENDOR QUOTATION ATTACHED
16	Reasons for recommending the equipment as BER	Checked and found suction motor and suction jars defective. The unit was installed on 04/03/2020 and covered up to 4+years. The repair cost is 83.3%. As per the tender clause 5.3.14.1 both criteria met. So recommending the unit for condemnation.
17	Name & Signature of CYRIX Authority	JESWIN SIBI 

*Not mandatory


#Based on the period of life and value as per the BER guidelines

* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Suction motor and suction jars defective.
Can be RBER as per tender conditions.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram


23-07-2023
Signature of JC BM (NHM)





Date:


Signature of
Superintendent/Medical Officer (i/c)
Thiruvananthapuram

Suction apparatus

Date	Name of Issue	ROL	Issued	Balance	Signature	Remarks
1.2.2014	Brought from page no 16.3 Volume No 1 Balance (18)					
19.3.93	16.9.98 10.5.98 7.4.2000	18		18		
5.11.2003	21.3.2005 15.6.2007					
23.5.2011	19.11.2013					
20/5/2014	Suction apparatus (Model Spares plate . 8.8) Stock of Reg. of page no 121 Volume 8A.	2		20		Rs 10,000/- (Ten thousand and two hundred/only)
20/5/2014	Suction apparatus (Part) 15007 G.N. Page No 78	1		21		
	As per order no. 5221/2014 dated 31/5/2015 (Cond. d)					15 Nil
	As per order no. 5608/2016 dated 17/6/2016 (Cond. d)					5 Nil
25/7/18	Received from store	1		6	UM	Stock Regi Page No 84
4/3/2020	Received from store Rs - 11,088 x 2 = 22,176/-	2		8	UM	Issue No. KCC 0087/10/00432/19-20 MSR. VRESB-271
17/4/2021	Received from store No 1400037/RS 100920/21-22	1		9	UM	
14/7/21	Entered in Unsuitable register Page No - 47			2	07	
2.9.21	Transferred to Stock register Vol. 11-A Page 13			07	NIL	
	As per order No. C ₂ -2161/26.7.2022/C.H.T dated 26.7.2022. Condensation done.					



YRIX BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

TOLL FREE NO: 1800-425-7669



BAR CODE - 0110825

KMSCL



☎ 974472041 974472041

✉ cloud99biotech@gmail.com

KMCA/III/1333
Room No. 22
CSI Golden Jubilee Complex
Baker Hill
Shastri Road
Kottayam

CLOUD99 BIOTECH

GSTIN: 32AMWP15633Q1Z0
State: 32 - Kerala

MSME NO.: UDYAM-KI-07-0030714

Estimate

Estimate No.: C99B-QTN-28
Date: 09/07/2024
Place of Supply: 32-Kerala

Estimate For:

KASHYAP

1ST FLOOR 30/641B CYRIX HEALTHCARE PVT LTD
KOCHI DHANUSHKODI ROAD POONITHURA

Contact No.: 7593847134
GSTIN Number: 32AA FCC2499H2ZM
State: 32-Kerala

#	Item name	HSN/SAC	Quantity	Unit	Price/Unit	GST	Amount
1	SUCTION JAR	90189099	1	Nos	₹ 3,214.29	₹ 385.71 (12%)	₹ 3,600.00
2	SUCTION MOTOR (MOTOR FOR HOSPITAL SUCTION)	9018	1	-	₹ 3,940.68	₹ 709.32 (18%)	₹ 4,650.00
Total			2			₹ 1,095.03	₹ 7,154.97



Pay To:
Bank Name : UNION BANK OF INDIA, KOTTAYAM
Bank Account No. : 351201010035646
Bank IFSC code : UBI0536121
Account holder's name : CLOUD99 BIOTECH

Sub Total	₹ 7,154.97
SGST@6%	₹ 192.86
CGST@6%	₹ 192.86
SGST@9%	₹ 354.66
CGST@9%	₹ 354.66
Total	₹ 8,250.00
You Saved	₹ 0.01

Description

DH NEDUMANGAD

SUCTION APPARATUS
Barcode: 0112464
Ticket id: 114956
Make: Supreme Enterprises

Estimate Amount In Words

Eight Thousand Two Hundred Fifty Rupees only

Terms And Conditions

Thanks for doing business with us!

For: CLOUD99 BIOTECH



Authorized Signatory



Quotation suction Machine - DH nedumangad (114956/0112464)

4 messages

Kasyep P V <kasyepcyrix@gmail.com>
To: CLOUD 99 <cloud99biotech@gmail.com>

Tue, 9 Jul 2024 at 5:28 pm

Dear Madam,

This mail is to enquire about the spare quotation for the suction machine;
Make : Supreme Enterprises

Quotation required;
Suction motor
Suction jar

Thanks & Regards,

KASYEP PV
District Incharge ,Trivandrum
CYRIX HEALTHCARE Pvt.Ltd
KLBEMP
Ph : +91 7593847134
kasyepcyrix@gmail.com

Kasyep P V <kasyepcyrix@gmail.com>
To: CLOUD 99 <cloud99biotech@gmail.com>

Sat, 20 Jul 2024 at 9:14 pm

Dear sir,
Consider this mail as gentle Reminder

[Quoted text hidden]

CLOUD 99 <cloud99biotech@gmail.com>
To: Kasyep P V <kasyepcyrix@gmail.com>

Sat, 20 Jul 2024 at 10:08 pm

[Quoted text hidden]