



# BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

## CYRIX

HEALTHCARE PVT LTD

**No. :**  
232647

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

<p>Health Facility <u>General hospital</u></p> <p>Address <u>Thiruvananthapuram</u> <u>Kerala</u></p> <p>Ph : <u>9744773198</u></p>	<p>Call Registration Date : <u>9-8-24</u></p> <p>Caller ID : <u>128524</u></p> <p>Date of Visit : <u>14-8-24</u></p> <p>Asset No. : <u>0112083</u></p> <p>EQPT Name : <u>Electrical Suction Unit</u></p> <p>Manufacture <u>Manach</u> Model : <u>MS1-R602</u></p> <p>S. No. <u>2019027</u> Dept. <u>OT</u></p>
---	--

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Not working

**Action Taken :** Checked and found suction motor and suction jaws defective. Need to replace these spares for further checking and working condition of the equipment


Completed  Date : 14/8/24 Time : 11:45 AM Spare Required

Spare Replaced  Requested

Description	Qty.	Part Number	PR Number
1. <u>WA</u>	<u>NA</u>	<u>NA</u>	<u>NA</u>
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>Sooys Sudheep</u>	<u>14/8/24</u>	<u>11:00 AM</u>	<u>11:45 AM</u>

**Customer Remark** Pending



<p>Service Engineer Name : <u>Sooys Sudheep</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>14/8/24</u></p> <p>Contact Number : <u>7907406244</u></p>	<p>Customer Name : <u>Sneetha</u></p> <p>Signature : <u>[Signature]</u></p> <p>Date : <u>14/8/24</u></p> <p>Contact Number : <u>9495898103</u></p> <p>Designation : <u>Superintendent</u></p> <p>Hospital Seal : <u>GH TUM</u></p>
--	--

General Hospital  
Thiruvananthapuram




**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GENERAL HOSPITAL THIRUVANANTHAPURAM
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : Electrical Suction Unit Make : Monarch Model : MSI RG02 Serial No : NA
4	Equipment ID/ Barcode	128524/0112083
5	Date of purchase/ Year of manufacture/Installation Date	04/03/2020
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	09/08/2024 (Toll Free)
9	Action taken	Checked and found suction motor and suction jars defective. Need to replace these spares for further checking and working condition of the equipment. Enquired spare from vendor.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	Suction Motor - 4650/- Suction Jar - 3600/- Total = 8250/-

13	Asset Value	9900/-
14	* Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	83.3%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT AND REFERENCE QUOTATION ATTACHED
16	Reasons for recommending the equipments as BER	Checked and found suction motor and suction jars defective. The unit was installed on 04/03/2020 and covered up to 4+years. The repair cost is 83.3%. As per the tender clause 5.3.14.1 both criteria met. So recommending the unit for condemnation.
17	Name & Signature of CYRIX Authority	JESWIN SIBI 

\*Not mandatory

#Based on the period of life and value as per the BER guidelines

\* Attach Photograph


Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Suction motor and suction jars defective. So can be recommended beyond economic repair as per tender conditions.

**MANEESHA MOHAN M.R.**  
Junior Consultant (Biomedical)  
National Health Mission  
Thiruvananthapuram

  
23-08-2024  
Signature of JC BM (NHM)

Date:



  
Signature of  
Superintendent/Medical Officer (i/c)



Superintendent  
General Hospital  
Thiruvananthapuram



Suction apparatus

Date	Name of Item	Qty	Issued	Balance	Signature	Remarks
1-2-2014	Brought forward from page no 163 Volume No 1 Balance (R)					
19.3.93 (1)	16.9.98 (1) 10.5.99 (1) 7.4.2000 (1)	18		18		
5.11.2003 (2)	21.3.2005 (3) 15.6.2007 (4)					
23.5.2011 (5)	19.11.2013 (3)					
20/5/2014	Suction apparatus (Model Systeme plate - S.S) Stock of Regi. of Soc page no 121 Volume 8A.	2		20		Rs 10,000/each (Ten thousand and two hundred/each only)
20/5/2014	Suction apparatus (Plate) from G.N. page no 79	1		21		
As per order no. C-5221/2014 GHT dated 3/5/2015 Condensed (6)						
As per order no. C/5608/2016 dated 17/6/2016 Condensed (10)						
21/1/2005 (3) 15/6/2007 (3) 23/5/2011 (5)						
25/7/18	Received from store	1		6	um	Stock Regi Page No 84
4/3/2020	Received from store Rs - 11,088 x 2 = 22,176/-	2		8	um	Issue No. KCR 0087/10/00632/19-20 MSR-VCSB-271
17/4/2021	Received from SHWA NO 1400037/RS/100920/21-22	1		9	um	
14.7.21	Entered in Unservicable register page no - 97			2	07	
2.9.21	Transferred to stock register vol. 11-A Page 63			07	NIL	
As per order no. C-2161/26.7.2022/C.H.T dated 26.7.2022 Condensation done						





BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME (BEMP)

CALL FREE NO: 1800-425-7669

BAR CODE - 0112083







924417303 924417304

cloud99biotech@gmail.com

1 KMG XII TEEB  
Room No: 22,  
CSI Golden Jubilee Complex  
Baker Hill  
Shastri Road  
Kottayam

# CLOUD99 BIOTECH

GSTIN: 32AMWP15633Q120  
State: 32 - Kerala

MSME NO: - UDYAM-KL-07-0030718

## Estimate

Estimate No.: C99B-QTN-28  
Date: 08/07/2024  
Place of Supply: 32-Kerala

Estimate For:

### KASHYAP

1ST FLOOR 30/641B CYRIX HEALTHCARE PVT LTD  
KOCHI DHANUSHKODI ROAD POONITHURA

Contact No.: 7593847134

GSTIN Number: 32AAFCC2499H2ZM

State: 32-Kerala

Sl	Item Name	HSN/SAC	Quantity	Unit	Price/Unit	GST	Amount
1	SUCTION JAR	90189099	1	Nos	₹ 3,214.29	₹ 385.71 (12%)	₹ 3,600.00
2	SUCTION MOTOR (MOTOR FOR HOSPITAL SUCTION)	9018	1	-	₹ 3,940.68	₹ 709.32 (18%)	₹ 4,650.00
Total			2			₹ 1,095.03	₹ 7,154.97



Pay To:

Bank Name : UNION BANK OF INDIA, KOTTAYAM

Bank Account No. : 361201010035646

Bank IFSC code : UBIN0536121

Account holder's name : CLOUD99 BIOTECH



Sub Total	₹ 7,154.97
SGST@6%	₹ 192.86
CGST@6%	₹ 192.86
SGST@9%	₹ 354.66
CGST@9%	₹ 354.66
<b>Total</b>	<b>₹ 8,250.00</b>
You Saved	< 0.01

### Description

DH NEDUMANGAD

SUCTION APPARATUS

Barcode: 0112454

Ticket id: 114956

Make : Supreme Enterprises

### Estimate Amount in Words

Eight Thousand Two Hundred Fifty Rupees only

### Terms And Conditions

Thanks for doing business with us!

For: CLOUD99 BIOTECH



Authorized Signatory



**Quotation suction Machine - DH nedumangad (114956/0112464)**

4 messages

**Kasyep P V <kasyepcyrix@gmail.com>**

To: CLOUD 99 <cloud99biotech@gmail.com>

Tue, 9 Jul 2024 at 5:28 pm

Dear Madam,

This mail is to enquire about the spare quotation for the suction machine;  
Make : Supreme Enterprises

Quotation required;  
Suction motor  
Suction jar

Thanks & Regards,

KASYEP PV  
District Incharge ,Trivandrum  
CYRIX HEALTHCARE Pvt.Ltd  
KLBEMP  
Ph : +91 7593847134  
kasyepcyrix@gmail.com

**Kasyep P V <kasyepcyrix@gmail.com>**

To: CLOUD 99 <cloud99biotech@gmail.com>

Sat, 20 Jul 2024 at 9:14 pm

Dear sir,  
Consider this mail as gentle Reminder

[Quoted text hidden]

**CLOUD 99 <cloud99biotech@gmail.com>**  
Tc: Kasyep P V <kasyepcyrix@gmail.com>

Sat, 20 Jul 2024 at 10:08 pm

[Quoted text hidden]