



BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAMME UNDER NATIONAL HEALTH MISSION



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX

HEALTHCARE PVT LTD

No. :

232475

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility <i>General Hospital</i>	Call Registration Date : <i>9/8/2024</i>
Address <i>Thiruvananthapuram</i>	Caller ID : <i>128527</i>
Ph : <i>9567271899</i>	Date of Visit : <i>14/8/2024</i>
	Asset No. : <i>0110758</i>
	EQPT Name : <i>Suction Apparatus</i>
	Manufacture <i>Suprem Enterprises</i> Model : <i>N.A</i>
	S. No. <i>N.A</i> Dept. <i>OT</i>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : *Not working*

Action Taken : *Checked and found that the suction motor suction jar and suction tubing are defective. Need to replace these spares for further checking and working condition of the equipment.*


Completed Date : *14/8/24* Time : *11:00AM* Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2. <i>NA</i>	<i>NA</i>	<i>NA</i>	<i>NA</i>
3.			

Cyrix Engineer	Date	Start Time	End Time
<i>Jeswin Sibi</i>	<i>14/8/24</i>	<i>10:00AM</i>	<i>11:00AM</i>

Customer Remark *Completed* Pending



Service Engineer Name : *Jeswin Sibi*
Signature : *[Signature]*
Date : *14/8/2024*
Contact Number : *9833315512*

Customer Name : *SreeLakshmi*
Signature : *[Signature]*
Date : *14/8/24*
Contact Number : *9495895703*
Designation : *CHTM*
Hospital Seal : *[Seal]*


Supervisor General Hospital

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GENERAL HOSPITAL THIRUVANANTHAPURAM
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : Suction Apparatus Make : Supreme Plus Model : Supreme Serial No : NA
4	Equipment ID/ Barcode	128527/0110758
5	Date of purchase/ Year of manufacture/Installation Date	25/07/2018
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No CAMC/AMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	09/08/2024 (Toll Free)
9	Action taken	Checked and found suction motor and suction jars defective. Need to replace these spares for further checking and working condition of the equipment. Enquired spare from vendor.
10	Present status of the equipment (Fully damaged / partially damaged)	FULLY DAMAGED
11	Recommendations for repair (required service details)	NOT RECOMMENDING FOR REPAIR
12	Cost of spares (specify parts and cost)	Suction Motor - 4650/- Suction Jar - 3600/- Total = 8250/-

13	Asset Value	10200/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	80.8%
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	CYRIX SERVICE REPORT AND VENDOR QUOTATION ATTACHED
16	Reasons for recommending the equipment as BER	Checked and found suction motor and suction jars defective. The unit was installed on 25/07/2018 and covered up to 6+years. The repair cost is 80.8%. As per the tender clause 5.3.14.1 both criteria met. So recommending the unit for condemnation.
17	Name & Signature of CYRIX Authority	JESWIN SIBI 

*Not mandatory

#Based on the period of life and value as per the BER guidelines

* Attach Photograph



Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Motor and suction jars defective. As per the tender conditions, this can be BER.

MANEESHA MOHAN M.R.
Junior Consultant (Biomedical)
National Health Mission
Thiruvananthapuram
Signature of JC BM (NHM)
23-08-2024




Date:



Signature of 
Superintendent of 
Thiruvananthapuram

Suction apparatus

Date	Name of Item	Qty	Unit	Balance	Signature	Remarks
1.2.2014	Brought from page no 16.3 Volume No 1 Balance (R)					
19.3.93 (1)	16.9.98 (1) 10.5.99 (1) 7.4.2001 (1)	18		18'		
5.11.2003 (2)	21.3.2005 (3) 15.6.2007 (3)					
23.5.2011 (5)	19.11.2013 (3)					
20/5/2014	Suction apparatus (Model Spares plate - 8.5) Stock of Reg. of Gen page no 121 Volume 8A.	2		20		Rs 10,200/each (Gen thousand and two hundred/each only)
20/5/2014	Suction apparatus (Stock of Reg. of Gen page no 79)	1		21		
	As per order no. C-1-5221/2014 dated 3/5/2015 (condemned (6)			15		
	As per order no. C-1-5608/2016 dated 17/6/2016 (condemned (10)			5		
25/7/18	Received from store	1		6	um	Stock Regi Page No 84
4/3/2020	Received from store Rs - 11,088 x 2 = 22,176/-	2		8	um	Issue No. K-1-0087/18/00432/19-20 MSR. Vot 27
17/9/2021	Received from store No 1400037/RS 100920/21-22	1		9	um	
14.7.21	Entered in Unserviceable register page no - 47	2		07		
2.9.21	Transferred to Stock register Vol. 11-A Page 63			07		
	As per order no. C-2-2161/26.7.2022/C.H.T dated 26.7.2022. Condemnation done.					





974473213 974473213

cloud99biotech@gmail.com

1 KMC/2011-1333
Room No: 22
CSI Golden Jubilee Complex
Baker Hill
Shastri Road
Kottayam

CLOUD99 BIOTECH

GSTIN: 32AMWPT5633Q1ZD
State: 32 - Kerala

MSME NO.: UDYAM-KL-07-0030714

Estimate

Estimate No.: C998-QTN-28
Date: 08/07/2024
Place of Supply: 32-Kerala

Estimate For:

KASHYAP

1ST FLOOR 30/621B CYRIX HEALTHCARE PVT LTD
KOCHI DHANUSHKODI ROAD POONITHURA

Contact No.: 7593847134
GSTIN Number: 32AA FCC2499H2ZM
State: 32-Kerala

#	Item name	HSN/SAC	Quantity	Unit	Price/Unit	GST	Amount
1	SUCTION JAR	90189099	1	Nos	₹ 3,214.29	₹ 385.71 (12%)	₹ 3,600.00
2	SUCTION MOTOR (MOTOR FOR HOSPITAL SUCTION)	9018	1	-	₹ 3,940.68	₹ 709.32 (18%)	₹ 4,650.00
Total						₹ 7,154.97	₹ 8,250.00

Sub Total	₹ 7,154.97
SGST@6%	₹ 192.86
CGST@6%	₹ 192.86
SGST@9%	₹ 354.66
CGST@9%	₹ 354.66
Total	₹ 8,250.00
Your Saved	₹ 0.01



Pay To:
Bank Name : UNION BANK OF INDIA, KOTTAYAM
Bank Account No. : 361201010035646
Bank IFSC code : UBIN0536121
Account holder's name : CLOUD99 BIOTECH

Description

DH NEDUMANGAD

SUCTION APPARATUS

Barcode: 0112464
Ticket id: 114956
Make : Supreme Enterprises

Estimate Amount In Words

Eight Thousand Two Hundred Fifty Rupees only

Terms And Conditions

Thanks for doing business with us!

For: CLOUD99 BIOTECH



Authorized Signatory



Quotation suction Machine - DH nedumangad (114956/0112464)

4 messages

Kasyep P V <kasyepcyrix@gmail.com>
To: CLOUD 99 <cloud99biotech@gmail.com>

Tue, 9 Jul 2024 at 5:28 pm

Dear Madam,

This mail is to enquire about the spare quotation for the suction machine;
Make : Supreme Enterprises

Quotation required;
Suction motor
Suction jar

Thanks & Regards,

KASYEP PV
District Incharge ,Trivandrum
CYRIX HEALTHCARE Pvt.Ltd
KLBEMP
Ph : +91 7593847134
kasyepcyrix@gmail.com

Kasyep P V <kasyepcyrix@gmail.com>
To: CLOUD 99 <cloud99biotech@gmail.com>

Sat, 20 Jul 2024 at 9:14 pm

Dear sir,
Consider this mail as gentle Reminder

[Quoted text hidden]

CLOUD 99 <cloud99biotech@gmail.com>
To: Kasyep P V <kasyepcyrix@gmail.com>

Sat, 20 Jul 2024 at 10:08 pm

[Quoted text hidden]