



**BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAMME  
UNDER  
NATIONAL HEALTH MISSION**



**SERVICE PROVIDER**  
Tender No. WO-37/2021-2022/698

**CYRIX**  
HEALTHCARE PVT LTD

**No. :** 232228

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

**Service Report**

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala  
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility .. General Hospital	Call Registration Date : 10 - 8 - 2024
Address .. Neyyattinkava	Caller ID : 128691
..... Thiruvananthapuram	Date of Visit : 12 - 8 - 2024
Ph : 9656784251	Asset No. : 0113222
	EQPT Name : Nebuliser
	Manufacture .. Cool .. Model : NA
	S. No. 0113222 Dept. C ward

**Service Classification :** Breakdown Call  PMS  Calibration  Cust.Training

**Problem Identified :** Not working

**Action Taken :** Checked the unit and identified nebulizer motor defective and piston broken. Need to replace these spares for further checking and working condition of the equipment.

**Completed**  **Date :** 12-8-2024 **Time :** 12:00 p.m **Spare Required**

**Spare Replaced**  **Requested**

Description	Qty.	Part Number	PR Number
1.			
2. NA	NA	NA	NA
3.			

Cyrix Engineer	Date	Start Time	End Time
Ajith George	12-8-2024	11:30 a.m	12:00 p.m

**Customer Remark**  Pending

<b>Service Engineer Name :</b> Ajith George <b>Signature :</b> A <b>Date :</b> 12 - 8 - 2024 <b>Contact Number :</b> 7593847111	<b>Customer Name :</b> GENERAL HOSPITAL <b>Signature :</b> [Signature] <b>Date :</b> 12/8/24 <b>Contact Number :</b> 9656784251 <b>Designation :</b> 10/0 <b>Hospital Seal :</b>
--	---



**SUPERINTENDENT**  
GENERAL HOSPITAL  
NEYYATTINKARA-685124


**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT  
MAINTENANCE PROGRAM (BEMP)**

**Recommendations for Beyond Economic Repair (BER)**

**PROFORMA**

Sl. No	Particulars	Details
1	Name of District	THIRUVANANTHAPURAM
2	Name of Hospital	GENERAL HOSPITAL NEYYATINKARA
3	Name of Equipment with Make, Model and Serial Number	Equipment Name : NEBULISER Manufacturer : COOL Model : NA Sl: NA
4	Equipment ID/ Barcode	128691/ 0113222
5	Date of purchase/ Year of manufacture/Installation Date	08/04/2022
6	Warranty details (Yes/No)	No Warranty
7	*AMC/ CAMC Period agreed at the time of Purchase	No AMC/CMC
8	Date of breakdown (Date of registration of complaint through email/ Toll free)	10/08/2024 (Toll free)
9	Action taken	Checked the unit and identified nebulizer motor defective and piston broken. Need to replace these spares for further checking and working condition of the equipment.
10	Present status of the equipment (Fully damaged / partially damaged)	Fully Damaged
11	Recommendations for repair (required service details)	Not recommending for Repair
12	Cost of spares (specify parts and cost)	Not available



13	Asset Value	Rs. 1353/-
14	Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	Na
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment RBER	Checked and found nebulizer motor defective and piston broken. Quotation not submitted since spares are not available in the market. So recommending the equipment for condemnation.
17	Name & Signature of CYRIX Authority	AJITH GEORGE 

\*Not mandatory


#Based on the period of life and value as per the BER guidelines

\* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:

Nebulizer motor defective and piston broken.  
So as per tender condition, this can be RBER.

MANEESHA MOHAN M.R.  
Junior Consultant (Biomedical)  
National Health Mission  
Thiruvananthapuram

 23-08-2024  
Signature of JC BM (NHM)

Date:






Signature of  
Superintendent/Medical Officer (i/c)

**SUPERINTENDENT**  
GENERAL HOSPITAL  
NEYYATTINKARA-695121

Nebulizers


83

	Rel	ISSUE	Balance	
29/9/2021			440	5
29/9/2021	Rel from Store.	1	542	10
	Donation page No: MSR (069)			
8/4/2022	TR from Casualty	1	642	10
	Sub stock Register page no: 315 (cool BC 0201)			6
29/10/2022	Condemned			
	no. 4.823/2022/			
	20/10/2022 (2)	2	6	
2/9/2023	TR out unscrutable		4	
	page no. 92.2 (1) (2) 19/1/19 (2) 20/1/19			
4/12/2023	Rel from store	1	5	
	Condemned as per order no. A/723/30/2/23 LS/1160/23-24 dated 4/12/23			
10/1/2024	Received from store	1	6	10
	LS/01296/23-24 dt. 10.1.24			
19/2/24	Rel from store	1	7	10
28/2/24	TR out to unscrutable	(2)	5	10
	Register B/L No 012914 page 92.23/2/19 013337			
8/5/24	Trans-fer in from General ward	1	6	10
	ward page NO: 35 (no date)			
20/5/2024	TR out to unscrutable	(2)	4	10
	Register page no 92 BC 0113322 (19.4.18) - 9113328			



SUPERINTENDENT  
GENERAL HOSPITAL  
MEYYATTINKARA-695121





*[Signature]*  
SUPERINTENDENT  
GENERAL HOSPITAL  
NEYYATTINKARA-695121