






PROFORMA

RECOMMENDATION FOR BEYOND ECONOMIC REPAIR (RBER)  
BIOMEDICAL EQUIPMENT MAINTENANCE PROGRAM (BEMP)

Name of Hospital: PHC Athirampuzha	Name of District: Kottayam
Name of Equipment :OT Light	Equipment ID & Barcode:130177 & 0541652
Make:Nil	Model: Nil
Serial Number: Nil	Warranty details: No Warranty
Date of purchase/ Year of manufacture/ Installation Date: Nil	Present status of the equipment: Fully Damaged
Date of breakdown:19/08/2024 (Toll Free) (Date of registration of complaint through email/ Toll free)	*AMC/ CAMC Period agreed at the time of purchase: No AMC/CMC
Action taken: Checked and find out the problem with lamp, lamp holder and power supply. The machine filly rusted need to repaint completely	
Recommendations for repair (required service details): Not recommending for repair.	
Cost of spares (specify parts and cost): NIL	
Percentage value of the cost of spares with respect to Cost of Purchase: Asset Value: NIL	Asset Value: NIL
Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not) : Cyrix service report and photos are attached	
Reasons for recommending the equipment as RBER: This registered item is a spot light not an OT light. It has been completed a life span of more than 10 year. Current this machine is totally in rusted and dilapidated condition. Service of this much old machine is not possible. All parts are damaged including Dom, power supply and body parts. To state the condition we are attending the images as proof.	
<p style="text-align: right;">Sanish Sivan <i>[Signature]</i> Name &amp; Signature of CYRIX Authority with date 21/10/24</p>	
<p>Remarks of Junior Consultant (Biomedical) NHM: <i>Asset value - NIL Cost of Spare - nil installa hos Details and stock entry details NA available - it is very old machine</i></p>	
Recommended for BER (Yes/ No): <i>Yes</i>	<p style="text-align: right;"><i>[Signature]</i> Signature of JC BM (NHM)</p>
Date: <i>21/10/24</i>	
Date: _____ Medical Officer (i/c)	<p style="text-align: right;"><i>[Signature]</i> Signature of Superintendent</p>



MEDICAL OFFICER INCHARGE  
PRIMARY HEALTH CENTRE  
ATHIRAMPUZHA



From

Jethu mol. K K  
Senior Nursing officer  
PHe Athirampuzha.

To

The moic  
PHe Athirampuzha.



MEDICAL OFFICER INCHARGE  
PRIMARY HEALTH CENTRE  
ATHIRAMPUZHA

Sir.

Biomedical equipment maintenance programme

[Barcode - 0541652] ന്റെ ഉപയോഗത്തിന് OG portable light  
work ചെയ്യാനും മറ്റും mainstock Entry ഉണ്ടാക്കുന്ന  
ബാർ. mainstock Entry ഉണ്ടാക്കുന്ന portable light  
Beyond repairable ലെറ്റിഫിക്കേഷൻ നൽകുന്നതിനായി  
പ്രാർത്ഥിക്കുന്നു. അയക്കുന്നതിനുള്ള നടപടികൾ  
കൊണ്ടുവരികെ തീർക്കണമെന്നും അഭ്യർത്ഥിക്കുന്നു.

Athirampuzha  
27/9/24.

മെ  
SNO  
JEGGK MOL





27  
2-80  
52-24  
0.9  
8  
14  
22-23



*[Signature]*  
MEDICAL OFFICER INCHARGE  
PRIMARY HEALTH CENTRE  
ATHIRAMPUZHA



4  
B  
23  
22-23



*[Signature]*  
MEDICAL OFFICER IN CHARGE  
PRIMARY HEALTH CENTRE  
ATHIRAMPUZHA

PRIMARY HEALTH CENTRE  
PIN- 686 562  
ATHIRAMPUZHA