



**BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAMME
UNDER
NATIONAL HEALTH MISSION**



SERVICE PROVIDER
Tender No. WO-37/2021-2022/698

CYRIX
HEALTHCARE PVT LTD

No. : 229499

ISO 13485 : 2012 & ISO 9001-2008 CERTIFIED COMPANY | AERB Approved Service Agency

Service Report

30/64 1 B, Petta Junction, Poonithura, Kochi - 682 038, Kerala
Ph : 98472 99500 Website : www.cyrix.com | E-mail : bemp.kl@cyrix.in

Health Facility : <u>Thygh</u>	Call Registration Date : <u>20/8/2024</u>
Address : <u>Vaikom</u>	Caller ID : <u>130270</u>
<u>Kottayam</u>	Date of Visit : <u>22/8/2024</u>
Ph : <u>9744341417</u>	Asset No. : <u>0520011</u>
	EQPT Name : <u>B.P APPARATUS</u>
	Manufacture <u>DISHA</u> Model : <u>ANTIQUE</u>
	S. No. <u>RM010958</u> Dept. <u>Casualty</u>

Service Classification : Breakdown Call PMS Calibration Cust.Training

Problem Identified : Mercury leak

Action Taken : Checked and found out the glass tube broken mercury tank leaked and spilled out one mercury. Screws of hinges are broken. Need to replace these all items for working condition

Completed Date : 22/8/2024 Time : 10:30am Spare Required

Spare Replaced Requested

Description	Qty.	Part Number	PR Number
1.			
2.			
3.			

Cyrix Engineer	Date	Start Time	End Time
<u>AMAL BABU</u>	<u>22/8/2024</u>	<u>10:00am</u>	<u>10:30am</u>

Customer Remark Equipment is not working Completed Pending


Service Engineer Name : <u>AMAL BABU</u>	Customer Name : <u>Head Quarters Hospital</u>
Signature : <u>[Signature]</u>	Signature : <u>[Signature]</u>
Date : <u>22/8/2024</u>	Date : <u>22/8/2024</u>
Contact Number : <u>9567575983</u>	Contact Number : <u>Malik Pina 688 141</u>
	Designation : <u>[Signature]</u>
	Hospital Seal : <u>[Seal]</u>

**REPAIR OF BIOMEDICAL EQUIPMENTS UNDER BIOMEDICAL EQUIPMENT
MAINTENANCE PROGRAM (BEMP)**

Recommendations for Beyond Economic Repair (BER)

PROFORMA

Sl. No	Particulars	Details
1	Name of District	KOTTAYAM
2	Name of Hospital	THQH VAIKOM
3	Name of Equipment with Make, Model and Serial Number	Equipment :BP APPARATUS Make: DISH Model :ANTIQUE SN: RMO109563
4	Equipment ID & Barcode	130270& 0520011✓
5	Date of purchase / Year of manufacture /Installation Date	08/04/2011
6	Warranty details (Yes/No)	No warranty
7	*AMC/ CAMC Period agreed at the time of purchase	No AMC/CMC
8	Date of breakdown(Date of registration of complaint through email/ Toll free)	20/08/2024 (Toll free)
9	Action taken	Checked and found out the glass tube broken, mercury tank leaked and spilled out the mercury, Screw of hinges are broken
10	Present status of the equipment (Fully damaged / partially damaged)	Fully damaged
11	Recommendations for repair (required service details)	Not recommending for repair.
12	Cost of spares (specify parts and cost)	N/A


13	Asset Value	Rs . 1200/-
14	# Percentage value of the cost of spares with respect to Cost of Purchase/ Asset Value	N/A
15	Abstract of Service Report provided by the OEM/ Authorized Service Provider/ CYRIX (Attached or Not)	Cyrix service report attached
16	Reasons for recommending the equipment as BER	Checked and found out the glass tube broken, mercury tank leaked and spilled out the mercury, Screw of hinges are broken. Machine installed in 8/04/2011 and covered up to 13 + years. Mercury not available in the market. So recommending to condemnation
17	Name & Signature of CYRIX Authority	AMAL BABU 

*Not mandatory **#Based on the period of life and value as per the BER guidelines**


* Attach Photograph

Remarks and Recommendations of Junior Consultant (Biomedical) NHM:


Asset value - 1200
 Cost of spare - Not mentioned
 Life period - 11 year 4 months
 Cyrix reported that glass tube broken
 Mercury spilled and mercury not available
 in the market


 Signature of JC BM (NHM)
 KOTTAYAM

Date



Superintendent
 Taluk Head Quarters Hospital
 * Vaikom, Pin - 686 141


 Signature of
 Superintendent / Medical Officer (i/c)

(Life used 5 years)

B.P. APPARATUS. Sphygmomanometer ¹¹

10/3/2010	B.P. from old 130010.P No. 162			10	22/2012
	26/8/03-01 ✓ 20/2/2012				
	21/11/04.2 ✓ 20/2/2012				
20/8/10	Received from Stone (RRS Box)	1.		11	
20/6/10	Received from Stone	1.		12	
1/2/2011	Received from Stone Rs. 936/- (Mercurial i. value)	1.		13	
5/4/11	Received from Stone (Rs. 2600/-)	3		16	
21/5/12	Rd from Stone (1365) E/KO.	1.		17.	Diabetic clinic.
30/11/12	Rd from Stone " (10)	3		20	
2012	Transferred to UAR Book page NO. 08 [Condemned as per order NO Bg Bs 1180/2011/DMO HK / DT 18/11/11]		3	17	
28/2/13	Rd from Stone "	3 ✓		20	
p.no 344 CIV	maistore Kanya Store) (Price 1590 + Tax)				
7/4/15	Rd from Stone Rs. "1885/- each. (In warranty - 1 yr)	2 ✓		22	SKT
26/11/15	Received de the st. Rs. "2050 (C/KO)	1 ✓		23	SKT
10/6/17	Rd from Stone Rs 1480 Rs. 5440 LP → Troubled to share the book. Condemned as per order NO C2-3735/14011 Vikram chd 13/6/23	3		25	SKT
20/6/17	Rd from the Stone (Life line) Rs 1764 LP book (Rs 182 page)	1		26	SKT
11/2/17	Rd from the Stone. (C/KO Rs. 4233.60) Condemned as per order NO C2-1676/17(1)(B) Supt of UAH V dt 27/3/18 A.S.B.	2		28	SKT
24.11.18	Rd from the Stone with 100/80 Bg " " " " " "	1	+	16	L
25.4.19	Transferred to UAR Book page P-50 (10.7.12-13) 28.2.13-13) 7.4.19 Condemned as per order NO C2-1288/2019 Btl. Vaikom Std 24.9.2019 (10.7.12-13) 28.2.13-13) 7.4.19 (20.10.15-17)			9	17



Superintendent
Taluk Head Quarters Hospital
Vaikom. PIN - 686 141



Superintendent
Tatuk Head Quarters Hospital
Walkom. Pin - 686 141

